





This project is Funded by KDSP

SUPPLY AND DELIVERY OF MEDICAL EQUIPMENTS PAP ONDITI –NYAKACH SUB COUNTY HOSPITAL (KDSP)

TENDER NO: CGK/FIN/KDSP/MEQ/OP/2021/2022/001

CLIENT COUNTY GOVERNMENT OF KISUMU DEPARTMENT OF FINANCE (KDSP) COUNTY GOVERNMENT OF KISUMU P.O BOX 2738-40100 KISUMU

<u>MAY</u>,

TABLE OF CONTENTS

PREFACEiii		
APP	ENDIX TO THE PREFACE	iv
GUII	DELINES FOR PREPARATION OF TENDER DOCUMENTS	iv
1.	GENERAL	iv
2.	PART 1 - TENDERING PROCEDURES	iv
3.	PART 2 – PROCUREMENT ENTITY'S REQUIREMENTS	v
4.	PART 3 – CONDITIONS OF CONTRACT AND CONTRACT FORMS	V
INVI	TATION TO TENDER	vii
Secti	on I - Instructions to Tenderers	1
A 1	General Provisions	1
2	Fraud and Corruption	1
3	Eligible Tenderers	
4	Eligible Goods and Related Services	3
B.	Contents of Request for Tenders Document	3
5.	Sections of Tendering Document	3
6	Clarification of Tender Document	
7	Amendment of Tender Document	4
C.	Preparation of Tenders	
8	Cost of Tendering	
9	Language of Tender Documents Comprising the Tender	
10 11	Form of Tender and Price Schedules	
12	Alternative Tenders	
13	Tender Prices and Discounts	
14	Currencies of Tender and Payment	
15	Documents Establishing the Eligibility and Conformity of the Goods and Related Services	7
16	Documents Establishing the Eligibility and Qualifications of the Tenderer	
17 18	Period of Validity of Tenders	
19	Format and Signing of Tender	
1)		
D . 20	Submission and Opening of Tenders	
21	Sealing and Marking of Tenders Deadline for Submission of Tenders	
22	Late Tenders	
23	Withdrawal, Substitution, and Modification of Tenders	
24	Tender Opening	
E	Evaluation and Comparison of Tenders	11
25	Confidentiality	
26	Clarification of Tenders	
27	Deviations, Reservations, and Omissions	
28	Determination of Responsiveness	
29	Non-conformities, Errors and Omissions	
30	Correction of Arithmetical Errors	13

31	Conversion to Single Currency	
32	Margin of Preference and Reservations	
33	Evaluation of Tenders	
34 35	Comparison of Tenders	
36.	Abnormally Low Tenders	
30.	Abhormany Tiigh Tenders	14
37.	Post Qualification of the Tenderer	15
38.	Lowest Evaluated Tender	
39.	Procuring Entity's Right to Accept Any Tender, and to Reject Any or All	15
F.	Award of Contract	
40.	Award Criteria	
41. 42.	Notice of Intention to enter into a Contract	
42. 43.	Standstill Period Debriefing by the Procuring Entity	
43. 44.	Letter of Award	
45.	Signing of Contract	
46.	Performance Security	
47.	Publication of Procurement Contract	17
48.	Procurement Related Complaint	17
SEC	TION II - TENDER DATA SHEET (TDS)	18
SEC	TION III EVALUATION AND QUALIFICATION CDITEDIA	22
	CTION III - EVALUATION AND QUALIFICATION CRITERIA	
	valuation of Tenders (ITT 33)	
	ARGIN OF PREFERENCE	
	ost Qualification of Tenderers (ITT 37)	
	TION IV - TENDERING FORMS	
	n	of
Forn		0.
Ten	der	0.
Teno Cert	lerficate of Independent Tender Determination	0.
Teno Cert	der	0.
Tend Cert Self-	derificate of Independent Tender Determination	29
Tend Cert Self- Form	der	29
Tend Cert Self- Form	der	29
Tend Cert Self- Form App Corn	der	29
Tend Cert Self- Form App Corn Tend 	der	29
Tend Cert Self- Forr App Corr Tend 	der	29
Tend Cert Self- Form App Corn Tend 	der	29
Tend Cert Self- Form App Corn Tend Tend	der	29
Tend Cert Self- Forr App Corr Tend Tend Price	der	29
Tend Cert Self- Forr App Corr Tend Tend Price FOR	der	29
Tend Cert Self- Forr App Corr Tend Tend Price FOR	der	
Tend Cert Self- Forr App Corr Tend Tend Price FOR FOR	der	
Tend Cert Self- Forr App Corr Tend Tend Price FOR FOR MA	der	
Tend Cert Self- Forr App Corr Tend Tend Price FOR FOR MA	der	
Tend Cert Self- Forr App Corn Tend Tend Price FOR FOR MAX	der	
Tend Cert Self- Forr App Corr Tend Tend Price FOR FOR MA	der	
Tend Cert Self- Forr App Corr Tend Tend Price FOR FOR MA	der	
Tend Cert Self- Forr App Corr Tend Tend Price FOR FOR MAT	declaration	
Tend Cert Self- Forr App Corr Tend Tend Price FOR FOR MAT	der	
Tend Cert Self- Forr App Corr Tend Tend Pricc FOR FOR MAX PAI Sect Note 1.	der	
Tend Cert Self- Forr App Corr Tend Tend Price FOR FOR MA: PAH Sect Note 1. 2. 3.	der	
Tend Cert Self- Forr App Corr Tend Tend Price FOR FOR MA. PAI Sect Note 1. 2. 3. 4. 5.	der	
Tend Cert Self- Forr App Corr Tend Tend Price FOR FOR MAX PAH Sect Note 1. 2. 3. 4. 5.	der	

SECTION VII - SPECIAL CONDITIONS OF CONTRACT	74
SECTION VIII - CONTRACT FORMS	78
FORM No. 1: NOTIFICATION OF INTENTION TO AWARD	79
FORM No. 2: REQUEST FOR REVIEW	82
FORM No. 3: LETTER OF AWARD	82
FORM No. 4 - CONTRACT AGREEMENT	84
FORM No. 5 - PERFORMANCE SECURITY [Option 1- Unconditional Demand Bank Guarantee]	85
FORM No. 6 - PERFORMANCE SECURITY [Option 2– Performance Bond]	86
FORM No. 7 - ADVANCE PAYMENT SECURITY [Demand Bank Guarantee]	88
FORM No. 8 - BENEFICIAL OWNERSHIP DISCLOSURE FORM	8

INVITATION TO TENDER

PROCURING ENTITY: COUNTY GOVERNMENT OF KISUMU, P. O. BOX 2738 KISUMU CONTRACT NAME AND DESCRIPTION: SUPPLY AND DELIVERY OF SUPPLY AND DELIVERY MEDICAL

TENDER NO: CGK/FIN/KDSP/MEQ/OP/2021/2022/001

 The COUNTY GOVERNMENT OF KISUMU invites sealed tenders for the supply of goods I.E._ SUPPLY AND DELIVERY MEDICAL

Tendering will be conducted under open competitive method *[National]* using a standardized tender document. Tendering is open to all qualified and interested Tenderers.

- 2. In case tender is subject to Multiple contracts/lots, insert "Tenderers will be allowed to tender for one or more lots". N/A
- 3. Qualified and interested tenderers may obtain further information and inspect the Tender Documents during office hours 0900 to 1700 hours] at the address given below. Office of the Director Supply Chain Management, P.O. BOX 2738 KISUMU, PROSPERITY BUILDING 2nd Floor, Wing C.
- 4. A complete set of tender documents may be obtained **free of charge**. Tender documents may be obtained electronically from the Kisumu County Website((<u>www.kisumu.go.ke</u>) and the Public Procurement Information Portal(<u>www.tenders.go.ke</u>). Tender documents obtained electronically will be free of charge.
- 5. Tender documents may be viewed and downloaded for free from the website (www.kisumu.go.ke). Tenderers who download the tender document must forward their particulars immediately to procurementcgk@gmail.com and/or

COUNTY GOVERNMENT OF KISUMU P.O. BOX 2738-40100 KISUMU

to facilitate any further clarificationor addendum.

- 6. All Tenders must be accompanied by a *tender Security* of **KES 171,909 valid for 91 days**.
- 7. The Tenderer shall chronologically serialize all pages of the tender documents submitted.
 - 8. Completed tenders must be delivered to the address below on or before 20th MAY, 2021, 1200 noon.
- 9. Electronic Tenders will not be permitted.
- 10. Tenders will be opened immediately after the deadline date and time specified above or anydead line date and time specified later. Tenders will be publicly opened in the presence of the Tenderers' designated representatives who choose to attend at the address below.
- 11. Late tenders will be rejected.
- 12. The addresses referred to above are:
 - a. Address for obtaining further information and for purchasing tender documents

COUNTY GOVERNMENT OF KISUMU OFFICE OF THE DIRECTOR SUPPLY CHAIN PROSPERITY BUILDING 2NDFLOOR WING C P.O. BOX 2738 KISUMU procurementcgk@gmail.com

A. Address for Submission of Tenders.

COUNTY GOVERNMENT OF KISUMU OFFICE OF THE DIRECTOR SUPPLY CHAIN PROSPERITY BUILDING TENDER BOX AT $2^{\rm ND}$ FLOOR WING C P.O. BOX 2738 KISUMU

B. Address for Opening of Tenders.

COUNTY GOVERNMENT OF KISUMU OFFICE OF THE DIRECTOR SUPPLY CHAIN PROSPERITY BUILDING TENDER BOX AT $2^{\rm ND}$ FLOOR WING C P.O. BOX 2738 KISUMU

JOAN. A. KANJEGO DIRECTOR SUPPLY CHAIN MANAGEMENT COUNTY GOVERNMENT OF KISUMU

PART 1 - TENDERING PROCEDURES

SECTION I: INSTRUCTIONS TO TENDERERS

A General Provisions

1. Scope of Tender

- 1.1 The Procuring Entity as defined in the **TDS** invites tenders for supply of goods and, if applicable, any Related Services incidental thereto, as specified in Section V, Supply Requirements. The name, identification, and number of lots (contracts) of this Tender Document are specified in the **TDS**.
- 12 Throughout this tendering document:
 - a) the term "in writing" means communicated in written form (e.g. by mail, e-mail, fax, including if specified in the **TDS**, distributed or received through the electronic-procurement system used by the Procuring Entity) with proof of receipt;
 - b) if the context so requires, "singular" means "plural" and vice versa;
 - c) "Day" means calendar day, unless otherwise specified as "Business Day". A Business Day is any day that is an official working day of the Procuring Entity. It excludes official public holidays.

2. Fraud and Corruption

- The Procuring Entity requires compliance with the provisions of the Public Procurement and Asset Disposal Act, 2015, Section 62 "Declaration not to engage in corruption". The tender submitted by a person shall include a declaration that the person shall not engage in any corrupt or fraudulent practice and a declaration that the person or his or her sub-contractors are not debarred from participating in public procurement proceedings.
- The Procuring Entity requires compliance with the provisions of the Competition Act 2010, regarding collusive practices in contracting. Any tenderer found to have engaged in collusive conduct shall be disqualified and criminal and/or civil sanctions may be imposed. To this effect, Tenders shall be required to complete and sign the "Certificate of Independent Tender Determination" annexed to the Form of Tender.
- Unfair Competitive Advantage Fairness and transparency in the tender process require that thefirms or their Affiliates competing for a specific assignment do not derive a competitive advantage from having provided consulting services related to this tender. To that end, the Procuring Entity shall indicate in the **Data Sheet** and make available to all the firms together with this tender document all information that would in that respect give such firm any unfair competitive advantageover competing firms.

3. Eligible Tenderers

- 3.1 A Tenderer may be a firm that is a private entity, an individual, a state-owned enterprise or institution subject to ITT3.7, or any combination of such entities in the form of a joint venture (JV) under an existing agreement or with the intent to enter into such an agreement supported by aletter of intent. Public employees and their close relatives (spouses, children, brothers, sisters and uncles and aunts) are not eligible to participate in the tender.
 - In the case of a joint venture, all members shall be jointly and severally liable for the execution of the entire Contract in accordance with the Contract terms. The JV shall nominate a Representative who shall have the authority to conduct all business for and on behalf of any and all the members of the JV during the Tendering process and, in the event the JV is awarded the Contract, during contract execution. The maximum number of JV members shall be specified in the **TDS**.
- Public Officers of the Procuring Entity, their Spouses, Child, Parent, Brothers or Sister. Child, Parent, Brother or Sister of a Spouse their business associates or agents and firms/organizations in which they have a substantial or controlling interest shall not be eligible to tender or be awarded a contract. Public Officers are also not allowed to participate in any procurement proceedings.
- A Tenderer shall not have a conflict of interest. Any Tenderer found to have a conflict of interest shall be disqualified. A Tenderer may be considered to have a conflict of interest for the purpose of this Tendering process, if the Tenderer:

- a) directly or indirectly controls, is controlled by or is under common control with another Tenderer; or
- b) receives or has received any direct or indirect subsidy from another Tenderer; or
- c) has the same representative or ownership as another Tenderer; or
- d) has a relationship with another Tenderer, directly or through common third parties, that puts it in a position to influence the Tender of another Tenderer, or influence the decisions of the Procuring Entity regarding this Tendering process; or
- e) or any of its affiliates participated as a consultant in the preparation of the design or technical specifications of the goods that are the subject of the Tender; or
- f) or any of its affiliates has been hired (or is proposed to be hired) by the Procuring Entity or Procuring Entity for the Contract implementation; or
- g) would be providing goods, works, or non-consulting services resulting from or directly related to consulting services for the preparation or implementation of the project specified in the **TDS** ITT 1.1 that it provided or were provided by any affiliate that directly or indirectly controls, is controlled by, or is under common control with that firm; or has a close business or family relationship with a professional staff of the Procuring Entity (or of the project implementing agency, who: (i) are directly or indirectly involved in the preparation of the tendering documentor specifications of the Contract, and/or the Tender evaluation process of such Contract; or (ii) would be involved in the implementation or supervision of such Contract unless the conflict stemming from such relationship has been resolved in a manner acceptable to the Procuring Entity throughout the Tendering process and execution of the Contract.
- 34 A tenderer shall not be involved in corrupt, coercive, obstructive, collusive or fraudulent practice. A tenderer that is proven to have been involved in any of these practices shall be automatically disqualified.
- A firm that is a Tenderer (either individually or as a JV member) shall not submit more than one Tender, except for permitted alternative Tenders. This includes participation as a subcontractor. Such participation shall result in the disqualification of all Tenders in which the firm is involved. A firm that is not a Tenderer or a JV member, may participate as a subcontractor in more than one Tender. Members of a joint venture may not also make an individual tender, be a subcontractor ina separate tender or be part of another joint venture for the purposes of the same Tender.
- A Tenderer may have the nationality of any country, subject to the restrictions pursuant to ITT3.9. A Tenderer shall be deemed to have the nationality of a country if the Tenderer is constituted, incorporated or registered in and operates in conformity with the provisions of the laws of that country, as evidenced by its articles of incorporation (or equivalent documents of constitution or association) and its registration documents, as the case may be. This criterion also shall apply tothe determination of the nationality of proposed subcontractors or sub consultants for any part of the Contract including related Services.
- 37 A Tenderer that has been debarred by the PPRA from participating in public procurement shall be ineligible to tender or be awarded a contract. The list of debarred firms and individuals is available from the PPRA's website www.ppra.go.ke
- Tenderers that are state-owned enterprises or institutions may be eligible to compete and be awarded a Contract(s) only if they are (i) a legal public entity of the state Government and/or public administration, (ii) financially autonomous and not receiving any significant subsidies or budget support from any public entity or Government, and (iii) operating under commercial law and vested with legal rights and liabilities similar to any commercial enterprise to enable it compete with firms in the private sector on an equal basis. Public employees and their close relatives are not eligible to participate in the tender.
- Tenderers may be ineligible if their countries of origin (a) as a matter of law or official regulations, Kenya prohibits commercial relations with that country, or(b) by an act of compliance with a decision of the United Nations Security Council taken under Chapter VII of the Charter of the United Nations, Kenya prohibits any import of goods or contracting for supply of goods or services from that country, or any payments to any country, person, or entity in that country. A tenderer shall provide such documentary evidence of eligibility satisfactory to the Procuring Entity, as the Procuring Entity shall reasonably request.
- 3.10 Tenderers shall provide the qualification information statement that the tenderer (including all members of a joint venture and subcontractors) is not associated, or have been associated in the past, directly or indirectly, with a firm or any of its affiliates which have been engaged by the Procuring entity to provide consulting services for the preparation of the design, specifications, and other documents to be used for the procurement of the goods under this Invitation for tenders.

- 311 Where the law requires tenderers to be registered with certain authorities in Kenya, such registration requirements shall be defined in the **TDS**
- 3.12 The Competition Act of Kenya requires that firms wishing to tender as Joint Venture undertakings which may prevent, distort or lessen competition in provision of services are prohibited unless they are exempt in accordance with the provisions of Section 25 of the Competition Act, 2010. JVs willbe required to seek for exemption from the Competition Authority. Exemption shall not be a condition for tender, but it shall be a condition of contract award and signature. A JV tenderershall be given opportunity to seek such exemption as a condition of award and signature of contract. Application for exemption from the Competition Authority of Kenya may be accessed from the website www.cak.go.ke.
- 3.13 A Kenyan tenderer shall provide evidence of having fulfilled his/her tax obligations by producing a current tax clearance certificate or tax exemption certificate issued by the Kenya Revenue Authority.

4. Eligible Goods and Related Services

- 4.1 All the Goods and Related Services to be supplied under the Contract shall have their origin inany country that is eligible in accordance with ITT 3.9.
- For purposes of this ITT, the term "goods" includes commodities, raw material, machinery, equipment, and industrial plants; and "related services" include services such as insurance, installation, training, and initial maintenance.
- The term "origin" means the country where the goods have been mined, grown, cultivated, produced, manufactured or processed; or, through manufacture, processing, or assembly, another commercially recognized article results that differs substantially in its basic characteristics from its components.
- A procuring entity shall ensure that the items listed below shall be sourced from Kenya and there shall be no substitutions from foreign sources. The affected items are:
 - a) motor vehicles, plant and equipment which are assembled in Kenya;
 - b) furniture, textile, foodstuffs, oil and gas, information communication technology, steel, cement, leather, agro-processed products, sanitary products, and other goods made in Kenya; or
 - c) goods manufactured, mined, extracted or grown in Kenya.
- Any goods, works and production processes with characteristics that have been declared by the relevant national environmental protection agency or by other competent authority as harmful to human beings and to the environment shall not be eligible for procurement.

5. Sections of Tendering Document

5.1 The tendering document consist of Parts 1, 2, and 3, which include all the sections indicated below, and should be read in conjunction with any Addenda issued in accordance with ITT8.

PART 1: Tendering Procedures

- i) Section I Instructions to Tenderers (ITT)
- ii) Section II Tendering Data Sheet (TDS)
- iii) Section III Evaluation and Qualification Criteria
- iv) Section IV Tendering Forms

PART 2: Supply Requirements

v) Section V - Schedule of Requirements

PART 3: Contract

- vi) Section VI General Conditions of Contract (GCC)
- vii) Section VII Special Conditions of Contract (SCC)
- viii) Section VIII- Contract Forms
- The notice of Invitation to Tender or the notice to the prequalified Tenderers issued by the Procuring Entity is not part of the tendering document.

- Unless obtained directly from the Procuring Entity, the Procuring Entity is not responsible for the completeness of the document, responses to requests for clarification, the minutes of the pre-tender meeting (if any), or addenda to the tendering document in accordance with ITT7.
- 54 The Tenderer is expected to examine all instructions, forms, terms, and specifications in the tendering document and to furnish with its Tender all information or documentation as is required by the tendering document.

6. Clarification of Tendering Document

- A Tenderer requiring any clarification of the Tender Document shall contact the Procuring Entity inwriting at the Procuring Entity's address specified in the **TDS** or raise its enquiries during the pre-Tender meeting if provided for in accordance with ITT 6.4. The Procuring Entity will respond in writing to any request for clarification, provided that such request is received no later than the period specified in the **TDS** prior to the deadline for submission of tenders. The Procuring Entity shall forward copies of its response to all tenderers who have acquired the Tender documents in accordance with ITT 5.3, including a description of the inquiry but without identifying its source. If so specified in the **TDS**, the Procuring Entity shall also promptly publish its response at the web page identified in the **TDS**. Should the clarification result in changes to the essential elements of the Tender Documents, the Procuring Entity shall amend the Tender Documents following the procedure under ITT 7.
- The Procuring Entity shall specify in the **TDS** if a pre-tender conference will be held, when and where. The Tenderer's designated representative is invited to attend a pre-Tender meeting. The purpose of the meeting will be to clarify issues and to answer questions on any matter that maybe raised at that stage.
- The Tenderer is requested to submit any questions in writing, to reach the Procuring Entity not later than the period specified in the **TDS** before the meeting.
- Minutes of the pre-Tender meeting, if applicable, including the text of the questions asked by Tenderers and the responses given, together with any responses prepared after the meeting, will be transmitted promptly to all Tenderers who have acquired the Tender Documents in accordance with ITT 6.3. Minutes shall not identify the source of the questions asked.
- The Procuring Entity shall also promptly publish anonymized (*no names*) Minutes of the pre-Tender meeting at the web page identified **in the TDS**. Any modification to the Tender Documents that may become necessary as a result of the pre-Tender meeting shall be made by the Procuring Entityexclusively through the issue of an Addendum pursuant to ITT 7 and not through the minutes of the pre-Tender meeting. Nonattendance at the pre-Tender meeting will not be a cause for disqualification of a Tenderer.

7. Amendment of Tendering Document

- 7.1 At any time prior to the deadline for submission of Tenders, the Procuring Entity may amend the tendering document by issuing addenda.
- Any addendum issued shall be part of the tendering document and shall be communicated in writing to all who have obtained the tender document from the Procuring Entity in accordance with ITT 6.3. The Procuring Entity shall also promptly publish the addendum on the Procuring Entity's web page in accordance with ITT 7.1.
- To give prospective Tenderers reasonable time in which to take an addendum into account in preparing their Tenders, the Procuring Entity may, at its discretion, extend the deadline for the submission of Tenders, pursuant to ITT 21.2.

C. Preparation of Tenders

8. Cost of Tendering

The Tenderer shall bear all costs associated with the preparation and submission of its Tender, and the Procuring Entity shall not be responsible or liable for those costs, regardless of the conduct or outcome of the Tendering process.

9. Language of Tender

9.1 The Tender, as well as all correspondence and documents relating to the Tender exchanged by the Tenderer and the Procuring Entity, shall be written in English Language. Supporting documents and printed literature that are part of the Tender may be in another language provided they areaccompanied by an accurate translation of the relevant passages into the English Language, in which case, for purposes of interpretation of the Tender, such translation shall govern.

10. Documents Comprising the Tender

- 10.1 The Tender shall comprise the following:
 - a) Form of Tender prepared in accordance with ITT11;
 - b) Price Schedules: completed in accordance with ITT 11 and ITT 13;
 - c) Tender Security or Tender-Securing Declaration, in accordance with ITT 18.1;
 - d) Alternative Tender: if permissible, in accordance with ITT12;
 - e) Authorization: written confirmation authorizing the signatory of the Tender to commit the Tenderer, in accordance with ITT19.3:
 - f) Qualifications: documentary evidence in accordance with ITT 16.2 establishing the Tenderer qualifications to perform the Contract if its Tender is accepted;
 - g) Tenderer Eligibility: documentary evidence in accordance with ITT16.1 establishing the Tenderer eligibility to tender;
 - h) Eligibility of Goods and Related Services: documentary evidence in accordance with ITT 15, establishing the eligibility of the Goods and Related Services to be supplied by the Tenderer;
 - i) Conformity: documentary evidence in accordance with ITT15.2 that the Goods and Related Services conform to the tender document; and
 - j) any other document required in the **TDS**.
- In addition to the requirements under ITT 10.1, Tenders submitted by a JV shall include a copy of the Joint Venture Agreement entered into by all members. Alternatively, a letter of intent to execute a Joint Venture Agreement in the event of a successful Tender shall be signed by all members and submitted with the Tender, together with a copy of the proposed Agreement.
- 103 The Tenderer shall furnish in the Form of Tender information on commissions gratuities, and fees, if any, paid or to be paid to agents or any other party relating to this Tender.

11. Form of Tender and Price Schedules

11.1 The Form of Tender and Price Schedules shall be prepared using the relevant forms furnished in Section IV, Tendering Forms. The forms must be completed without any alterations to the text. All blank spaces shall be filled in with the information requested. The Tenderer shall chronologically serialize pages of all tender documents submitted.

12. Alternative Tenders

121 Unless otherwise specified in the TDS, alternative Tenders shall not be considered.

13. Tender Prices and discounts

- 13.1 The prices quoted by the Tenderer in the Form of Tender and in the Price, Schedules shall conform to the requirements specified below.
- 132 All lots (contracts) and items must be listed and priced separately in the Price Schedules.
- The price to be quoted in the Form of Tender in accordance with ITT10.1 shall be the total price of the Tender, including any discounts offered.
- 134 The Tenderer shall quote any discounts and indicate the methodology for their application in the form of tender. Conditional discounts will be rejected.

- Prices quoted by the Tenderer shall be fixed during the performance of the Contract and notsubject to variation on any account, unless otherwise specified **in the TDS.** A Tender submitted with an adjustable price quotation shall be treated as non-responsive and shall be rejected, pursuant to ITT 28. However, if in accordance with **the TDS**, prices quoted by the Tenderer shall besubject to adjustment during the performance of the Contract, a Tender submitted with a fixed price quotation shall not be rejected, but the price adjustment shall be treated as zero.
- If specified in ITT 1.1, Tenders are being invited for individual lots (contracts) or for any combination of lots (packages). Unless otherwise specified **in the TDS**, prices quoted shall correspond to 100 % of the items specified for each lot and to 100% of the quantities specifiedfor each item of a lot. Tenderers wishing to offer discounts for the award of more than one Contract shall specify in their Tender the price reductions applicable to each package, or alternatively, to individual Contracts within the package. Discounts shall be submitted in accordance with ITT 13.4 provided the Tenders for all lots (contracts) are opened at the same time.
- 13.7 The terms EXW, CIP, CIF, DDP and other similar terms shall be governed by the rules prescribed in the current edition of Incoterms, published by the International Chamber of Commerce.
- Prices shall be quoted as specified in each Price Schedule included in Section IV, Tendering Forms. The disaggregation of price components is required solely for the purpose of facilitating the comparison of Tenders by the Procuring Entity. This shall not in any way limit the Procuring Entity's right to contract on any of the terms offered. In quoting prices, the Tenderer shall be free to use transportation through carriers registered in any eligible country. Similarly, the Tenderer may obtain insurance services from any eligible country in accordance with ITT 3.6, Eligible Tenders. Prices shall be entered in the following manner:
 - a) For Goods manufactured in Kenya:
 - I) the price of the Goods quoted EXW (ex-works, ex-factory, ex warehouse, ex showroom, or off-the- shelf, as applicable) final destination point indicated in the **TDS**, includingall customs duties and sales and other taxes already paid or payable on the components and raw material used in the manufacture or assembly of the Goods;
 - ii) any sales tax and other taxes which will be payable in Kenya on the Goods if the Contract is awarded to the Tenderer; and
 - the price for inland transportation, insurance, and other local services required to convey the Goods to their final destination specified in the TDS.
 - b) For Goods manufactured outside Kenya, to be imported:
 - i) the price of the Goods, quoted CIP named place of destination, in Kenya, as specified in the TDS;
 - the price for inland transportation, insurance, and other local services required to convey the Goods from the named place of destination to their final destination specified in the TDS;
 - c) For Goods manufactured outside Kenya, already imported:
 - i) the price of the Goods, including the original import value of the Goods; plus, any markup (or rebate); plus, any other related local cost, and custom duties and other import taxes already paid or to be paid on the Goods already imported;
 - ii) the custom duties and other import taxes already paid (need to be supported with documentary evidence) or to be paid on the Goods already imported;
 - iii) any sales and other taxes levied in Kenya which will be payable on the Goods if the Contract is awarded to the Tenderer; and
 - **iv**) the price for inland transportation, insurance, and other local services required to convey the Goods from the named place of destination to their final destination (Project Site) specified **in the TDS.**
 - d) for Related Services, other than inland transportation and other services required to convey the Goods to their final destination, whenever such Related Services are specified in the Schedule of Requirements, the price of each item comprising the Related Services (inclusive of any applicable taxes).

14. Currencies of Tender and Payment

- The currency (ies) of the Tender, the currency (ies) of award and the currency (ies) of contract payments shall be the same.
- The Tenderer shall quote in Kenya shillings. If allowed in the **TDS**, the Tenderer may express the Tender price in any currency, provided it shall use no more than two foreign currencies in additionto the Kenya Shilling.
- The rates of exchange to be used by the Tenderer shall be based on the exchange rates provided by the Central Bank of Kenya on the date 30 days prior to the actual date of tender opening.

15. Documents Establishing the Eligibility and Conformity of the Goods and Related Services

- 15.1 To establish the eligibility of the Goods and Related Services in accordance with ITT 15, Tenderers shall complete the country of origin declarations in the Price Schedule Forms, included in Section IV, Tendering Forms.
- To establish the conformity of the Goods and Related Services to the tendering document, the Tenderer shall furnish as part of its Tender the documentary evidence that the Goods conform to the technical specifications and standards specified in Section VII, Schedule of Requirements.
- The documentary evidence may be in the form of literature, drawings or data, and shall consist of a detailed item by item description of the essential technical and performance characteristics of the Goods and Related Services, demonstrating substantial responsiveness of the Goods and Related Services to the technical specification, and if applicable, a statement of deviations and exceptions to the provisions of the Section VII, Schedule of Requirements.
- 15.4 The Tenderer shall also furnish a list giving full particulars, including available sources and current prices of spare parts, special tools, etc., necessary for the proper and continuing functioning of the Goods during the period **specified in the TDS** following commencement of the use of the goodsby the Procuring Entity.
- Standards for workmanship, process, material, and equipment, as well as references to brand names or catalogue numbers specified by the Procuring Entity in the Schedule of Requirements, are intended to be descriptive only and not restrictive. The Tenderer may offer other standards of quality, brand names, and/or catalogue numbers, provided that it demonstrates, to the Procuring Entity's satisfaction, that the substitutions ensure substantial equivalence or are superior to those specified in the Section VII, Schedule of Requirements.

16. Documents Establishing the Eligibility and Qualifications of the Tenderer

- To establish Tenderer eligibility in accordance with ITT 4, Tenderers shall complete the Form of Tender, included in Section IV, Tendering Forms.
- The documentary evidence of the Tenderer qualifications to perform the Contract if its Tender is accepted shall establish to the Procuring Entity's satisfaction:
 - a) that, if required **in the TDS**, a Tenderer that does not manufacture or produce the Goods it offers to supply shall submit the Manufacturer's Authorization using the form included in Section IV, Tendering Forms to demonstrate that it has been duly authorized by the manufacturer or producer of the Goods to supply these Goods in Kenya;
 - b) that, if required **in the TDS**, in case of a Tenderer not doing business within the Kenya, the Tenderer is or will be (if awarded the Contract) represented by an Agent in the country equipped and able to carry out the Supplier's maintenance, repair and spare parts-stocking obligations prescribed in the Conditions of Contract and/or Technical Specifications; and
 - c) that the Tenderer meets each of the qualification criterion specified in Section III, Evaluation and Oualification Criteria.

17. Period of Validity of Tenders

- 17.1 Tenders shall remain valid for the Tender Validity period specified **in the TDS**. The Tender Validity period starts from the date fixed for the Tender submission deadline (as prescribed by the Procuring Entity in accordance with ITT 21.1). A Tender valid for a shorter period shall be rejected by the Procuring Entity as non-responsive.
- In exceptional circumstances, prior to the expiration of the Tender validity period, the Procuring Entity may request Tenderers to extend the period of validity of their Tenders. The request and the responses shall be made in writing. If a Tender Security is requested in accordance with ITT 18, it shall also be extended for a corresponding period. A Tenderer may refuse the request without forfeiting its Tender Security. A Tenderer granting the request shall not be required or permitted to modify its Tender, except as provided in ITT 17.3.
- 173 If the award is delayed by a period exceeding the number of days to be specified in the TDS

days beyond the expiry of the initial tender validity period, the Contract price shall be determined as follows:

- a) in the case of **fixed price** contracts, the Contract price shall be the tender price adjusted by the factor specified **in the TDS**;
- b) in the case of **adjustable price** contracts, no adjustment shall be made; or in any case, tenderevaluation shall be based on the tender price without taking into consideration the applicable correction from those indicated above.

18. Tender Security

- 181 The Tenderer shall furnish as part of its Tender, either a Tender-Securing Declaration or a Tender Security, as specified in the TDS, in original form and, in the case of a Tender Security, in the amount and currency specified in the TDS.
- 182 A Tender Securing Declaration shall use the form included in Section IV, Tendering Forms.
- If a Tender Security is specified pursuant to ITT 18.1, the Tender Security shall be a demand guarantee in any of the following forms at the Tenderer option:
 - i) cash;
 - ii) a bank guarantee;
 - iii) a guarantee by an insurance company registered and licensed by the Insurance Regulatory Authority listed by the Authority; or
 - iv) a letter of credit; or
 - v) guarantee by a deposit taking micro-finance institution, Sacco society, the Youth Enterprise Development Fund or the Women Enterprise Fund.
- If an unconditional guarantee is issued by a non-Bank financial institution located outside Kenya, the issuing non-Bank financial institution shall have a correspondent financial institution located in Kenya to make it enforceable unless the Procuring Entity has agreed in writing, prior to Tender submission, that a correspondent financial institution is not required. In the case of a bank guarantee, the Tender Security shall be submitted either using the Tender Security Form included in Section IV, Tendering Forms, or in another substantially similar format approved by the Procuring Entity prior to Tender submission. The Tender Security shall be valid for thirty
 - (30) days beyond the original validity period of the Tender, or beyond any period of extension ifrequested under ITT 17.2.
- If a Tender Security is specified pursuant to ITT 18.1, any Tender not accompanied by a substantially responsive Tender Security shall be rejected by the Procuring Entity as non-responsive.
- If a Tender Security is specified pursuant to ITT 18.1, the Tender Security of unsuccessful Tenderers shall be returned as promptly as possible upon the successful Tenderer signing the Contract and furnishing the Performance Security pursuant to ITT 46. The Procuring Entity shall also promptly return the tender security to the tenderers where the procurement proceedings are terminated, all tenders were determined non-responsive or a bidder declines to extend tender validity period.
- 187 The Tender Security of the successful Tenderer shall be returned as promptly as possible once the successful Tenderer has signed the Contract and furnished the required Performance Security.
- 188 The Tender Security may be forfeited or the Tender Securing Declaration executed:
 - a) if a Tenderer withdraws its Tender during the period of Tender validity specified by the Tenderer in the Form of Tender, or any extension thereto provided by the Tenderer; or
 - b) if the successful Tenderer fails to:
 - i) sign the Contract in accordance with ITT 45; or
 - ii) furnish a Performance Security in accordance with ITT 46.
- Where tender securing declaration is executed, the Procuring Entity shall recommend to the PPRA that PPRA debars the Tenderer from participating in public procurement as provided in the law.
- 18.10 The Tender Security or Tender-Securing Declaration of a JV must be in the name of the JV that submits the Tender. If the JV has not been legally constituted into a legally enforceable JV at the time of Tendering, the Tender Security or Tender-Securing Declaration shall be in the names of all future members as named in the letter of intent referred to in ITT3.1 and ITT 10.2.
- 18.11 A tenderer shall not issue a tender security to guarantee itself.

19. Format and Signing of Tender

- 19.1 The Tenderer shall prepare one original of the documents comprising the Tender as described inITT 11 and clearly mark it "ORIGINAL." Alternative Tenders, if permitted in accordance with ITT 12, shall be clearly marked "ALTERNATIVE." In addition, the Tenderer shall submit copies of the Tender, in the number **specified in the TDS** and clearly mark them "COPY." In the event of any discrepancy between the original and the copies, the original shall prevail.
- 192 Tenderers shall mark as "CONFIDENTIAL" information in their Tenders which is confidential to their business. This may include proprietary information, trade secrets, or commercial or financially sensitive information.
- The original and all copies of the Tender shall be typed or written in indelible ink and shall be signed by a person duly authorized to sign on behalf of the Tenderer. This authorization shall consist of a written confirmation **as specified in the TDS** and shall be attached to the Tender. The name and position held by each person signing the authorization must be typed or printed below the signature. All pages of the Tender where entries or amendments have been made shall be signed or initialed by the person signing the Tender.
- In case the Tenderer is a JV, the Tender shall be signed by an authorized representative of the JV on behalf of the JV, and so as to be legally binding on all the members as evidenced by a power of attorney signed by each members' legally authorized representatives.
- Any inter-lineation, erasures, or overwriting shall be valid only if they are signed or initialed bythe person signing the Tender.

D. Submission and Opening of Tenders

20 Sealing and Marking of Tenders

- Depending on the sizes or quantities or weight of the tender documents, a tenderer may use an envelope, package or container. The Tenderer shall deliver the Tender in a single sealed envelope, or in a single sealed package, or in a single sealed container bearing the name and Reference number of the Tender, addressed to the Procuring Entity and a warning not to open before the time and date for Tender opening date. Within the single envelope, package or container, the Tenderer shall place the following separate, sealed envelopes:
 - a) in an envelope or package or container marked "ORIGINAL", all documents comprising the Tender, as described in ITT 11; and
 - b) in an envelope or package or container marked "COPIES", all required copies of the Tender; and
- 202 The inner envelopes or packages or containers shall:
 - a) bear the name and address of the Procuring Entity.
 - b) bear the name and address of the Tenderer; and
 - c) bear the name and Reference number of the Tender.
- Where a tender package or container cannot fit in the tender box, the procuring entity shall:
 - a) Specify in the **TDS where** such documents should be received.
 - b) maintain a record of tenders received and issue acknowledgement receipt note to each tenderer specifying time and date of receipt.
 - c) Ensure all tenders received are handed over to the tender opening committee for opening at the specified opening place and time.
- 20.4 If an envelope or package or container is not sealed and marked as required, the *Procuring Entity* will assume no responsibility for the misplacement or premature opening of the Tender. Tenders misplaced or opened prematurely will not be accepted.

21. Deadline for Submission of Tenders

- 21.1 Tenders must be received by the Procuring Entity at the address and no later than the date and time specified in the TDS. When so specified in the TDS, Tenderers shall have the option of submitting their Tenders electronically. Tenderers submitting Tenders electronically shall follow the electronic Tender submission procedures specified in the TDS.
- 212 The Procuring Entity may, at its discretion, extend the deadline for the submission of Tenders by

amending the tendering document in accordance with ITT7, in which case all rights and obligations of the Procuring Entity and Tenderers previously subject to the deadline shall thereafter be subject to the deadline as extended.

22. Late Tenders

The Procuring Entity shall not consider any Tender that arrives after the deadline for submission of Tenders. Any Tender received by the Procuring Entity after the deadline for submission of Tenders shall be declared late, rejected, and returned unopened to the Tenderer.

23. Withdrawal, Substitution, and Modification of Tenders

- A Tenderer may withdraw, substitute, or modify its Tender after it has been submitted by sendinga written notice, duly signed by an authorized representative, and shall include a copy of the authorization (the power of attorney) in accordance with ITT19.3, (except that withdrawal notices not require copies). The corresponding substitution or modification of the Tender must accompany the respective written notice. All notices must be:
 - a) prepared and submitted in accordance with ITT 20 and 21 (except that withdrawal notices do not require copies), and in addition, the respective envelopes shall be clearly marked "WITHDRAWAL," "SUBSTITUTION," or "MODIFICATION;" and
 - b) received by the Procuring Entity prior to the deadline prescribed for submission of Tenders, in accordance with ITT 22.
- Tenders requested to be withdrawn in accordance with ITT 23.1 shall be returned unopened to the Tenderers.
- No Tender may be withdrawn, substituted, or modified in the interval between the deadline for submission of Tenders and the expiration of the period of Tender validity specified by the Tendereron the Form of Tender or any extension thereof.

24. Tender Opening

- Except as in the cases specified in ITT 23, the Procuring Entity shall, at the Tender opening, publicly open and read out all Tenders received by the deadline at the date, time and place specified in the **TDS** in the presence of Tenderers' designated representatives who choose to attend, including to attend any specific electronic tender opening procedures if electronic tendering is permitted in accordance with ITT 21.1, shall be as specified in the **TDS**.
- First, envelopes marked "WITHDRAWAL" shall be opened and read out and the envelope with the corresponding Tender shall not be opened, but returned to the Tenderer. If the withdrawal envelope does not contain a copy of the "power of attorney" confirming the signature as a person duly authorized to sign on behalf of the Tenderer, the corresponding Tender will be opened. No Tender withdrawal shall be permitted unless the corresponding withdrawal notice contains a valid authorization to request the withdrawal and is read out at Tender opening.
- Next, envelopes marked "SUBSTITUTION" shall be opened and read out and exchanged with the corresponding Tender being substituted, and the substituted Tender shall not be opened, but returned to the Tenderer. No Tender substitution shall be permitted unless the corresponding substitution notice contains a valid authorization to request the substitution and is read out at Tender opening.
- Next, envelopes marked "MODIFICATION" shall be opened and read out with the corresponding Tender. No Tender modification shall be permitted unless the corresponding modification notice contains a valid authorization to request the modification and is read out at Tender opening.
- Next, all remaining envelopes shall be opened one at a time, reading out: the name of the Tenderer and whether there is a modification; the total Tender Prices, per lot (contract) if applicable, including any discounts and alternative Tenders; the presence or absence of a Tender Security, if required; and any other details as the Procuring Entity may consider appropriate.
- Only Tenders, alternative Tenders and discounts that are opened and read out at Tender opening shall be considered further for evaluation. The Form of Tender and pages of the Bills of Quantities are to be initialed by the members of the tender opening committee attending the opening. The number of representatives of the Procuring Entity to sign shall be specified in the **TDS**.

- 24.7 The Procuring Entity shall neither discuss the merits of any Tender nor reject any Tender (except for late Tenders, in accordance with ITT 22.1).
- 248 The Procuring Entity shall prepare a record of the Tender opening that shall include, as a minimum:
 - a) the name of the Tenderer and whether there is a withdrawal, substitution, or modification;
 - b) the Tender Price, per lot (contract) if applicable, including any discounts;
 - c) any alternative Tenders;
 - d) the presence or absence of a Tender Security or Tender-Securing Declaration, if one was required;
 - e) number of pages of each tender document submitted.
- 249 The Tenderers' representatives who are present shall be requested to sign the record. The omission of a Tenderer signature on the record shall not invalidate the contents and effect of the record. A copy of the tender opening register shall be issued to a Tenderer upon request.

E. Evaluation and Comparison of Tenders

25. Confidentiality

- 25.1 Information relating to the evaluation of Tenders and recommendation of contract award, shall not be disclosed to Tenderers or any other persons not officially concerned with the tendering process until the information on Intention to Award the Contract is transmitted to all Tenderers in accordance with ITT 41.
- Any effort by a Tenderer to influence the Procuring Entity in the evaluation or contract award decisions may result in the rejection of its Tender.
- Notwithstanding ITT 25.2, from the time of Tender opening to the time of Contract Award, if any Tenderer wishes to contact the Procuring Entity on any matter related to the Tendering process, it should do so in writing.

26. Clarification of Tenders

To assist in the examination, evaluation, comparison of the Tenders, and qualification of the Tenderers, the Procuring Entity may, at its discretion, ask any Tenderer for a clarification of its Tender. Any clarification submitted by a Tenderer in respect to its Tender and that is not in response to a request by the Procuring Entity shall not be considered. The Procuring Entity's request for clarification and the response shall be in writing. No change, including any voluntary increase or decrease, in the prices or substance of the Tender shall be sought, offered, or permitted except to confirm the correction of arithmetic errors discovered by the Procuring Entity in the Evaluation of the Tenders, in accordance with ITT 30.

If a Tenderer does not provide clarifications of its Tender by the date and time set in the Procuring Entity's request for clarification, its Tender may be rejected.

27. Deviations, Reservations, and Omissions

- 27.1 During the evaluation of Tenders, the following definitions apply:
 - a) "Deviation" is a departure from the requirements specified in the Tendering document;
 - b) "Reservation" is the setting of limiting conditions or withholding from complete acceptance of the requirements specified in the tendering document; and
 - c) "Omission" is the failure to submit part or all of the information or documentation required in the tendering document.

28. Determination of Responsiveness

- 28.1 The Procuring Entity's determination of a Tender's responsiveness is to be based on the contents of the Tender itself, as defined in ITT28.2.
- 28 A substantially responsive Tender is one that meets the requirements of the tendering document without material deviation, reservation, or omission. A material deviation, reservation, or omission is one that:

- a) if accepted, would:
 - i) affect in any substantial way the scope, quality, or performance of the Goods and Related Services specified in the Contract; or
 - ii) limit in any substantial way, inconsistent with the tendering document, the Procuring Entity's rights or the Tenderer obligations under the Contract; or
- b) if rectified, would unfairly affect the competitive position of other Tenderers presenting substantially responsive Tenders.
- The Procuring Entity shall examine the technical aspects of the Tender submitted in accordance with ITT 15 and ITT 16, in particular, to confirm that all requirements of Section VII, Schedule of Requirements have been met without any material deviation or reservation, or omission.
- 283 If a Tender is not substantially responsive to the requirements of tendering document, it shall be rejected by the Procuring Entity and may not subsequently be made responsive by correction of the material deviation, reservation, or omission.

29. Non-conformities, Errors and Omissions

- 29.1 Provided that a Tender is substantially responsive, the Procuring Entity may waive any non-conformities in the Tender.
- 292 Provided that a Tender is substantially responsive, the Procuring Entity may request that the Tenderer submit the necessary information or documentation, within a reasonable period of time, to rectify nonmaterial non- conformities or omissions in the Tender related to documentation requirements. Such omission shall not be related to any aspect of the price of the Tender. Failure of the Tenderer to comply with the request may result in the rejection of its Tender.
- Provided that a Tender is substantially responsive, the Procuring Entity shall rectify quantifiable nonmaterial non-conformities related to the Tender Price. To this effect, the Tender Price shall be adjusted, for comparison purposes only, to reflect the price of a missing or non-conforming item or component in the manner specified **in the TDS**. The adjustment shall be based on the *average* price of the item or component as quoted in other substantially responsive Tenders. If the price of the item or component cannot be derived from the price of other substantially responsive Tenders, the Procuring Entity shall use its best estimate.

30. Arithmetical Errors

- 30.1 The tender sum as submitted and read out during the tender opening shall be absolute and final and shall not be the subject of correction, adjustment or amendment in any way by any person or entity.
- Provided that the Tender is substantially responsive, the Procuring Entity shall handle errors on the following basis:
 - a) Any error detected if considered a major deviation that affects the substance of the tender, shall lead to disqualification of the tender as non-responsive .
 - b) Any errors in the submitted tender arising from a miscalculation of unit price, quantity, subtotal and total bid price shall be considered as a major deviation that affects the substance of the tender and shall lead to disqualification of the tender as non-responsive. and
 - c) if there is a discrepancy between words and figures, the amount in words shall prevail.
- 303 Tenderers shall be notified of any error detected in their bid during the notification of a ward.

31. Conversion to Single Currency

31.1 For evaluation and comparison purposes, the currency(ies) of the Tender shall be converted in a single currency as specified **in the TDS.**

32 Margin of Preference and Reservations

A margin of preference may be allowed on locally manufactured goods only when the contract is open to international tendering, where the tender is likely to attract foreign goods and where the

contract exceeds the threshold specified in the Regulations.

- For purposes of granting a margin of preference on locally manufactured goods under international competitive tendering, a procuring entity shall not subject the items listed below to international tender and hence no margin of preference shall be allowed. The affected items are:
 - a) motor vehicles, plant and equipment which are assembled in Kenya;
 - b) furniture, textile, foodstuffs, oil and gas, information communication technology, steel, cement, leather agro-processing, sanitary products, and other goods made in Kenya; or
 - c) goods manufactured, mined, extracted or grown in Kenya.
- 323 A margin of preference shall not be allowed unless it is specified so in the **TDS**.
- Contracts procured on basis of international competitive tendering shall not be subject to reservations to specific groups s as provided in ITT 32.5.
- Where it is intended to reserve a contract to a specific group of businesses (these groups are Smalland Medium Enterprises, Women Enterprises, Youth Enterprises and Enterprises of persons living with disability, as the case may be), and who are appropriately registered as such by the authority to be specified in the **TDS**, a procuring entity shall ensure that the invitation to tender specifically indicates that only businesses or firms belonging to the specified group are eligible to tender as specified in the **TDS**. No tender shall be reserved to more than one group. If not so stated in the Tender documents, the invitation to tender will be open to all interested tenderers.

33 Evaluation of Tenders

- 33.1 The Procuring Entity shall use the criteria and methodologies listed in this ITT and Section III, Evaluation and Qualification criteria. No other evaluation criteria or methodologies shall be permitted. By applying the criteria and methodologies, the Procuring Entity shall determine the Lowest Evaluated Tender. This is the Tender of the Tenderer that meets the qualification criteria andwhose Tender has been determined to be:
 - a) substantially responsive to the tender documents; and
 - b) the lowest evaluated price.
- Price evaluation will be done for Items or Lots (contracts), as specified **in the TDS**; and the Tender Price as quoted in accordance with ITT 14. To evaluate a Tender, the Procuring Entity shall consider the following:
 - a) price adjustment due to unconditional discounts offered in accordance with ITT 13.4;
 - b) converting the amount resulting from applying (a) and (b) above, if relevant, to a singlecurrency in accordance with ITT 31;
 - c) price adjustment due to quantifiable nonmaterial non-conformities in accordance with ITT 29.3; and
 - d) any additional evaluation factors specified **in the TDS** and Section III, Evaluation and Qualification Criteria.
- 333 The estimated effect of the price adjustment provisions of the Conditions of Contract, applied over the period of execution of the Contract, shall not be considered in Tender evaluation.
- Where the tender involves multiple lots or contracts, the tenderer will be allowed to tender for one or more lots (contracts). Each lot or contract will be evaluated in accordance with ITT 33.2. The methodology to determine the lowest evaluated tenderer or tenderers based one lot (contract) or based on a combination of lots (contracts), will be specified in Section III, Evaluation and Qualification Criteria. In the case of multiple lots or contracts, tenderer will be will be required to prepare the Eligibility and Qualification Criteria Form for each Lot.
- 335 The Procuring Entity's evaluation of a Tender will include and consider:
 - a) in the case of Goods manufactured in Kenya, sales and other similar taxes, which will bepayable on the goods if a contract is awarded to the Tenderer;
 - b) in the case of Goods manufactured outside Kenya, already imported or to be imported, customs duties and other import taxes levied on the imported Good, sales and other similar taxes, which will be payable on the Goods if the contract is awarded to the Tenderer;

The Procuring Entity's evaluation of a Tender may require the consideration of other factors, in addition to the Tender Price quoted in accordance with ITT 14. These factors may be related to the characteristics, performance, and terms and conditions of purchase of the Goods and Related Services. The effect of the factors selected, if any, shall be expressed in monetary terms to facilitate comparison of Tenders, unless otherwise specified in the **TDS** from amongst those set out in Section III, Evaluation and Qualification Criteria. The additional criteria and methodologies to be used shall be as specified in ITT 33.2(d).

34 Comparison of Tenders

34.1 The Procuring Entity shall compare the evaluated costs of all substantially responsive Tenders established in accordance with ITT 33.2 to determine the Tender that has the lowest evaluated cost. The comparison shall be on the basis of total cost (place of final destination) prices for all goods and all prices, plus cost of inland transportation and insurance to place of destination, for goods manufactured within the Kenya, together with prices for any required installation, training, commissioning and other services.

35. Abnormally Low Tenders

- 35.1 An Abnormally Low Tender is one where the Tender price, in combination with other constituent elements of the Tender, appears unreasonably low to the extent that the Tender price raises material concerns with the Procuring Entity as to the capability of the Tenderer to perform the Contract for the offered Tender price.
- In the event of identification of a potentially Abnormally Low Tender by the evaluation committee, the Procuring Entity shall seek written clarification from the Tenderer, including a detailed price analyses of its Tender price in relation to the subject matter of the contract, scope, delivery schedule, allocation of risks and responsibilities and any other requirements of the tendering document.
- After evaluation of the price analysis, in the event that the Procuring Entity determines that the Tenderer has failed to demonstrate its capability to perform the contract for the offered Tenderprice, the Procuring Entity shall reject the Tender.

36. Abnormally High Tenders

- An abnormally high price is one where the tender price, in combination with other constituent elements of the Tender, appears unreasonably too high to the extent that the Procuring Entity is concerned that it (the Procuring Entity) may not be getting value for money or it may be paying too high a price for the contract compared with market prices or that genuine competition between Tenderers is compromised.
- 36.5 In case of an abnormally high tender price, the Procuring Entity shall make a survey of the marketprices, check if the estimated cost of the contract is correct and review the Tender Documents to check if the specifications, scope of work and conditions of contract are contributory to the abnormally high tenders. The Procuring Entity may also seek written clarification from the tenderer on the reason for the high tender price. The Procuring Entity shall proceed as follows:
 - i) If the tender price is abnormally high based on wrong estimated cost of the contract, the Procuring Entity may accept or not accept the tender depending on the Procuring Entity's budget considerations.
 - ii) If specifications, scope of work and/or conditions of contract are contributory to the abnormally high tender prices, the Procuring Entity shall reject all tenders and may retender for the contract based on revised estimates, specifications, scope of work and conditions of contract, as the case may be.
- 36.6 If the Procuring Entity determines that the Tender Price is abnormally too high because genuine competition between tenderers is compromised (*often due to collusion, corruption or other manipulations*), the Procuring Entity shall reject all Tenders and shall institute or cause relevant Government Agencies to institute an investigation on the cause of the compromise, before retendering.

37. Post-Qualification of the Tenderer

- 37.1 The Procuring Entity shall determine, to its satisfaction, whether the eligible Tenderer that is selected as having submitted the lowest evaluated cost and substantially responsive Tender, meetsthe qualifying criteria specified in Section III, Evaluation and Qualification Criteria.
- The determination shall be based upon an examination of the documentary evidence of the Tenderer qualifications submitted by the Tenderer, pursuant to ITT 15 and 16. The determination shall not take into consideration the qualifications of other firms such as the Tenderer subsidiaries, parent entities, affiliates, subcontractors (other than specialized subcontractors if permitted in the tendering document), or any other firm(s) different from the Tenderer.
- An affirmative determination shall be a prerequisite for award of the Contract to the Tenderer. A negative determination shall result in disqualification of the Tender, in which event the Procuring Entity shall proceed to the Tenderer who offers a substantially responsive Tender with the next lowest evaluated cost to make a similar determination of that Tenderer qualifications to perform satisfactorily.

38 Lowest Evaluated Tender

- 38.1 Having compared the evaluated prices of Tenders, the Procuring Entity shall determine the Lowest Evaluated Tender. The Lowest Evaluated Tender is the Tender of the Tenderer that meets the Qualification Criteria and whose Tender has been determined to be:
 - a) most responsive to the Tender document; and
 - b) the lowest evaluated price.

39. Procuring Entity's Right to Accept Any Tender, and to Reject Any or All Tenders.

39.1 The Procuring Entity reserves the right to accept or reject any Tender, and to annul the Tendering process and reject all Tenders at any time prior to notification Award, without thereby incurring any liability to Tenderers. In case of annulment, all Tenderers shall be notified with reasons and all Tenders submitted and specifically, tender securities, shall be promptly returned to the Tenderers.

F. Award of Contract

40. Award Criteria

40.1 The Procuring Entity shall award the Contract to the successful tenderer whose tender has been determined to be the Lowest Evaluated Tender in accordance with procedures in Section 3: Evaluation and Oualification Criteria.

41. Procuring Entity's Right to Vary Quantities at Time of Award

41.1 The Procuring Entity reserves the right at the time of Contract award to increase or decrease, by the percentage (s) for items as indicated **in the TDS.**

42 Notice of Intention to enter into a Contract

Upon award of the contract and Prior to the expiry of the Tender Validity Period the Procuring Entity shall issue a Notification of Intention to Enter into a Contract / Notification of award to all tenderers which shall contain, at a minimum, the following information:

- a) the name and address of the Tenderer submitting the successful tender;
- b) the Contract price of the successful tender;
- c) a statement of the reason(s) the tender of the unsuccessful tenderer to whom the letter is addressed was unsuccessful, unless the price information in (c) above already reveals the reason;
- d) the expiry date of the Standstill Period; and
- e) instructions on how to request a debriefing and/or submit a complaint during the standstill period;

43. Standstill Period

43.1 The Contract shall not be awarded earlier than the expiry of a Standstill Period of 14 days to

allow any dissatisfied candidate to launch a complaint. Where only one Tender is submitted, the Standstill Period shall not apply.

Where standstill period applies, it shall commence when the Procuring Entity has transmitted to each Tenderer the Notification of Intention to Enter into a Contract to the successful Tenderer.

4. Debriefing by the Procuring Entity

- On receipt of the Procuring Entity's Notification of Intention to Enter into a Contract referred to in ITT 41, an unsuccessful tenderer may make a written request to the Procuring Entity for adebriefing on specific issues or concerns regarding their tender. The Procuring Entity shall provide the debriefing within five days of receipt of the request.
- Debriefings of unsuccessful Tenderers may be done in writing or verbally. The Tenderer shall bearits own costs of attending such a debriefing meeting.

45. Letter of Award

Prior to the expiry of the Tender Validity Period and upon expiry of the Standstill Period specified in ITT 42, upon addressing a complaint that has been filed within the Standstill Period, the Procuring Entity shall transmit the Letter of Award to the successful Tenderer. The letter of award shall request the successful tenderer to furnish the Performance Security within 21days of the date of the letter.

46. Signing of Contract

- Upon the expiry of the fourteen days of the Notification of Intention to enter into contract and upon the parties meeting their respective statutory requirements, the Procuring Entity shall send the successful Tenderer the Contract Agreement.
- Within fourteen (14) days of receipt of the Contract Agreement, the successful Tenderer shall sign,date, and return it to the Procuring Entity.
- The written contract shall be entered into within the period specified in the notification of awardand before expiry of the tender validity period.

47. Performance Security

- Within twenty-one (21) days of the receipt of Letter of Acceptance from the Procuring Entity, the successful Tenderer, if required, shall furnish the Performance Security in accordance with the GCC 18, using for that purpose the Performance Security Form included in Section X, Contract Forms. If the Performance Security furnished by the successful Tenderer is in the form of a bond, it shall be issued by a bonding or insurance company that has been determined by the successful Tenderer to be acceptable to the Procuring Entity. A foreign institution providing a bond shall have a correspondent financial institution located in Kenya, unless the Procuring Entity has agreed in writing that a correspondent financial institution is not required.
- Failure of the successful Tenderer to submit the above-mentioned Performance Security or sign the Contract shall constitute sufficient grounds for the annulment of the award and forfeiture of the Tender Security. In that event the Procuring Entity may award the Contract to the Tendereroffering the next lowest Evaluated Tender.
- Performance security shall not be required for a contract, if so specified in the **TDS**.

48. Publication of Procurement Contract

- 48.1 Within fourteen days after signing the contract, the Procuring Entity shall publish and publicize the awarded contract at its notice boards, entity website; and on the Website of the Authority in manner and format prescribed by the Authority. At the minimum, the notice shall contain the following information:
 - a) name and address of the Procuring Entity;
 - b) name and reference number of the contract being awarded, a summary of its scope and the selection method used:
 - c) the name of the successful Tenderer, the final total contract price, the contract duration.
 - d) dates of signature, commencement and completion of contract;
 - e) names of all Tenderers that submitted Tenders, and their Tender prices as read out at Tender opening;

49. Procurement Related Complaints and Administrative Review

49.1 The procedures for making a Procurement-related Complaint are as specified in the TDS.

492	A request for administrative review shall be made in the form provided under contract forms.

SECTION II – TENDER DATA SHEET (TDS)

The following specific data shall complement, supplement, or amend the provisions in the Instructions to Tenderers (ITT). Whenever there is a conflict, the provisions herein shall prevail over those in ITT.

ITT Reference	Particulars Of Appendix To Instructions To Tenders
A. General	1 at includes Of Appendix 10 mstructions 10 Tenuers
	The reference much as of the Invited as for Ton Jose !-
ITT 1.1	The reference number of the Invitation for Tenders is CGK/KDSP/FIN/MEQ/OP/2021/2022/001 The Procuring Entity is: COUNTY
	GOVERNMENT OF KISUMU
	The name of the Contract is: SUPPLY AND DELIVERY OF MEDICA
	EQUIPMENT PAP ONDITI –NYAKACH SUB COUNTY HOSPITAL (KDSP)
	The number and identification of lots (contracts) comprising this Invitation for Tenders
	is: CGK/KDSP/FIN/MEQ/OP/2021/2022/001
ITT 3.1	MANDATORY REQUIREMENTS
	Tenderers MUST include the following information and documents with their
	Tenders:
	1. Duly filled Form of Tender – Correctly filled, signed, certified and stamped
	2. Certified copy certificate of incorporation
	3. Certified copy of recent CR12 Form-(12Months) from registrar of companies
	and their ID's where applicable
	4. Copy of Valid Tax Compliance Certificate from KRA.
	5. Copy of PIN/VAT Certificate
	6. Tender Security of KES 171,909/- valid for 91 days.
	7. Certified copy of Single Business Permit preferably from Kisumu County.
	8. Bidders shall SEQUENTIALLY SERIALIZE ALL PAGES for each
	tender document submitted failure to which the bidder shall be disqualified.
ITT 3.7	A list of debarred firms and individuals is available on the PPRA's website:
ITT 3.11	www.ppra.go.ke Tenderers shall be required to be to be registered with PUBLIC PROCUREMENT
3.11	REGULATORY AUTHORITY- N/A
	B. Contents of Tendering Document
ITT 6.1	(a) Address where to send enquiries is
	COUNTY GOVERNMENT OF KISUMU
	P.O. BOX 2738-40100,
	KISUMU
	to reach the Procuring Entity not later than 20th MAY, 2022, 1200 noon.
ITT 6.3	The questions to reach the Procuring Entity not later than 7 DAYS FROM
	THE DATE OF ADVERT 11th May, 2022
ITT 6.5	There shall be NO Pre-tender meeting.
	Separation of Tenders
ITT 10 (j)	
• • •	
ITT 12.1	Alternative Tenders "shall not be" considered.
ITT 13.5	The prices quoted by the Tenderer "shall not" be subject to adjustment
	during the performance of the Contract.
ITT 13.6	Prices quoted for each lot (contract) shall correspond at least to [insert figure] percent
	of the items specified for each lot (contract). N/A

ITT Reference	Particulars Of Appendix To Instructions To Tenders
	Prices quoted for each item of a lot shall correspond at least to <i>[insert figure]</i> percent of the
	quantities specified for this item of a lot.N/A
ITT 13.8 (a) (i) and (iii)	Place of final destination: COUNTY GOVERNMENT OF KISUMU (Health Department Pap Onditi)
ITT 13.8 (a) (iii)	Final Destination (Project Site): [insert final destination/project site, if different from named place of destination] - N/A
ITT 13.8 (b) (i)	Named place of destination, in Kenya is KISUMU
ITT 13.8 (b) (ii)	price for inland transportation, insurance, and other local services required to convey the Goods from the named place of destination to their final destination which is N/A
13.8 (c) (iv)	Place of final destination (Project Site) Kisumu.
ITT 14.2	Foreign currency requirements not allowed.
ITT 15.4	Period of time the Goods are expected to be functioning (for the purpose of spare parts): N/A
ITT 16.2 (a)	Manufacturer's authorization is: N/A
ITT 16.2 (b)	After sales service is: N/A The Tenden well-divergeried shall be 84 down
ITT 17.1	The Tender validity period shall be 84 days.
ITT 17.3	(a) The Number of days beyond the expiry of the initial tender validity period will be 30 days.
ITT 18.1	A Tender Security "shall be" required. A Tender-Securing Declaration "shall not be" required. Tender Security of KES 171,909/- FROM A REPUTABLE BANK RECOGNISED BY THE CENTRAL BANK OF KENYA OR ISSUARANCE COMPANY LISTED BY PPRA WITH A 60 DAY VALIDITY.
ITT 19.1	In addition to the original of the Tender, the number of copies is: [1 original and 1 copy]
ITT 19.3	The written confirmation of authorization to sign on behalf of the Tenderer shall consist of: Power of Attorney
	D. Submission and Opening of Tenders
ITT 20.3	A tender package or container that cannot fit in the tender box shall be received as follows: COUNTY GOVERNMENT OF KISUMU OFFICE OF THE DIRECTOR SUPPLY CHAIN PROSPERITY BUILDING TENDER BOX AT 2 ND FLOOR WING C P.O. BOX 2738 KISUMU
	RECEIVED AT THE OFFICE OF THE DIRECTOR SUPPLY CHAIN, 2^{ND} FLOOR WING C,
	MAKE SURE TO RECORD; 1. THE NAME OF THE COMPANY, 2. ITEM DESCRIPTION 3. NAME OF THE PERSON SUBMITTING THE DOCUMENT 4. DATE RECEIVED 5. TIME RECEIVED

ITT Reference	Particulars Of Appendix To Instructions To Tenders
	The deadline for Tender submission is: Date: 20 th MAY, 2022 Time: 12:00 NOON Tenderers "shall not" have the option of submitting their Tenders electronically.
ITT 24.1	The Tender opening shall take place at: COUNTY GOVERNMENT OF KISUMU OFFICE OF THE DIRECTOR SUPPLY CHAIN PROSPERITY BUILDING TENDER BOX AT 2 ND FLOOR WING C P.O. BOX 2738 KISUMU Date: 20 th MAY, 202 Time: 1200 noon
ITT 24.6	The number of representatives of the Procuring Entity to sign is Four (4No.)
ITT 29.3	Evaluation and Comparison of Tenders The evaluation and comparison of tenders shall be conducted as per mandatory requirements as per ITT. 3.1
ITT 31.1	The currency that shall be used for Tender evaluation and comparison purposes to convert at the selling exchange rate all Tender prices expressed in various currencies into a single currency is: <i>KENYA SHILLINGS</i>
ITT 32.3	A margin of preference and/or reservation "shall not" apply and specify the details. If a margin of preference applies, the application methodology shall be defined in Section III – Evaluation and Qualification Criteria.
ITT 32.5	This tender shall be an Open Tender.
ITT 33.2	Price evaluation will be done for bids successful past the mandatory stage.
ITT 33.6	 The adjustments shall be determined using the following criteria, from amongst those set out in Section III, Evaluation and Qualification Criteria: N/A (a) Deviation in Delivery schedule: [insert Yes or No. If yes insert the adjustment factor in Section III, Evaluation and Qualification Criteria] (b) Deviation in payment schedule: [insert Yes or No. If yes insert the adjustment factor in Section III, Evaluation and Qualification Criteria] N/A (c) the cost of major replacement component, mandatory spare parts, and service: [insert Yes or No. If yes, insert the Methodology and criteria in Section III, Evaluation and Qualification Criteria] N/A

ITT Reference	Particulars Of Appendix To Instructions To Tenders
	 (d) the availability in Kenya of spare parts and after-sales services for the equipment offered in the Tender [insert Yes or No. If yes, insert the Methodology and criteria in Section III, Evaluation and Qualification Criteria] N/A (e) Life cycle costs: the costs during the life of the goods or equipment [insert Yes or No. If yes, insert the Methodology and criteria in Section III, Evaluation and Qualification Criteria] N/A (f) the performance and productivity of the equipment offered; [Insert Yes or No. If yes, insert the Methodology and criteria] N/A
	(g) F. Award of Contract
ITT 41.1	The maximum percentage by which quantities may be increased is: N/A
	The maximum percentage by which quantities may be decreased is: N/A
ITT 41.1	The Procuring Entity shall increase or decrease the quantity of Goods and Related Services by an amount not exceed 15 % and without any change in theunit prices or other terms and conditions of the Tender and the tendering document.
ITT 47.3	Performance security if so required shall be in the sum of N/A .
ITT 49.1	The procedures for making a Procurement-related Complaint are detailed in the "Notice of Intention to Award the Contract" herein and are also available from the PPRA Website www.ppra.go.ke.
	If a Tenderer wishes to make a Procurement-related Complaint, the Tenderer should submit its complaint following these procedures, in writing (by the quickest means available, that is either by email or fax), to:
	For the attention: [insert full name of person receiving complaints] Title/position: [insert title/position]
	Procuring Entity: [insert name of Procuring Entity]
	Email address: [insert email address] In summary, a Procurement-related Complaint may challenge any of the following:
	1. the terms of the Tendering Documents; and the Procuring Entity's decision to award the contract.

SECTION III - EVALUATION AND QUALIFICATION CRITERIA

6. General Provisions

- Wherever a Tenderer is required to state a monetary amount, Tenderers should indicate the Kenya Shilling equivalent using the rate of exchange determined as follows:
 - a) For business turnover or financial data required for each year Exchange rate prevailing on the last day of the respective calendar year (in which the amounts for that year is to be converted) was originally established.

- b) Value of single contract Exchange rate prevailing on the date of the contract signature.
- c) Exchange rates shall be taken from the publicly available source identified in **the ITT 14.3.** Any error in determining the exchange rates in the Tender may be corrected by the Procuring Entity.
- This section contains the criteria that the Procuring Entity Procuring Entity shall use to evaluate tender and qualify tenderers. No other factors, methods or criteria shall be used other than those specified in this tender document. The Tenderer shall provide all the information requested in the forms included in Section IV, Tendering Forms. The Procuring Entity should use the Standard Tender Evaluation Report for Goods and Works for evaluating Tenders.

7. Evaluation of Tenders (ITT 33)

7.1 Successful Tender or Tenders

The Procuring Entity shall use the criteria and methodologies listed in this Section to evaluate Tenders. By applying these criteria and methodologies, the Procuring Entity shall determine the successful Tender or Tenders which has/have been determined to:

- a) be substantially responsive to the tender documents;
- b) offer the lowest evaluated cost to the Procuring Entity for all items of Goods to be procured based on either a single Contract or all multiple Contracts combined, as the case may be, in accordance with the ITT 13.6 inviting Tender prices and discounts, and provisions made of the Tender Document for evaluation of tenders and award of contract (s); and
- c) be offered by Tenderer or Tenderers that substantially meet the qualification criteria applicable for Contract or combined Contracts for which they are selected.

72 Evaluation of Tenders

Preliminary examination for Determination of Responsiveness

The Procuring Entity will start by examining all tenders to ensure they meet in all respects the eligibility criteria and other mandatory requirements in the ITT, and thatthe tender is complete in all aspects in meeting the requirements provided for in the preliminary evaluation criteria outlined below. The Standard Tender Evaluation Report Document for Goods and Works for evaluating Tenders provides very clear guide on how to deal with review of these requirements. Tenders that do not pass the Preliminary Examination will be considered non-responsive and will not be considered further.

Tenderers **MUST** include the following information and documents with their Tenders:

- 1. Duly filled Form of Tender Correctly filled, signed, certified and stamped
- 2. **Certified** copy certificate of incorporation
- 3. **Certified** copy of recent CR12 Form-(12Months) from registrar of companies and their ID's where applicable
- 4. Copy of **Valid** Tax Compliance Certificate from KRA.
- 5. Copy of PIN/VAT Certificate
- 6. Tender Security of KES 171,909/- valid for 91 days.
- 7. **Certified** copy of Single Business Permit preferably from Kisumu County.
- 8. Bidders shall **SEQUENTIALLY SERIALIZE ALL PAGES** for each tender document submitted failure to which the bidder shall be disqualified.

PRICE EVALUATION

Consistent with and in addition to the criteria listed in ITT 33.3 and ITT 29.3; and ITT 34 and its subparagraphs the following criteria shall apply:

221 Evaluation of Technical aspects of the Tender

The Procuring Entity shall evaluate the Technical aspects of the Tender to determine compliance with the Procuring Entity's requirements under Section V 'Schedule of Requirement' and whether the Tenders are substantially responsive to the Technical Specifications and other Requirements.

[The Procuring Entity will highlight herein any particular details, characteristics, functional guarantees or

other requirements under the specifications, which the Tenderer is required to specifically confirm or provide details as per Section V, Supply Requirements or other parts of the Tender Document. To facilitate, a template may be attached or clearly described all information and list of documentation to be submitted by Tenderers to enable evaluation of Technical parts of the Tender]

222 Evaluation of Commercial Terms and Conditions of the Tender (ITT 33.1(a)):

The Procuring Entity shall determine whether the Tenders are substantially responsive to the Commercial and Contractual Terms and Conditions (e.g. Performance securities, Payment and delivery schedules).

[The Procuring Entity will highlight herein any particular requirements under the Contract which the Tenderer is required to specifically confirm or provide information to enable evaluation of Commercial Terms and Conditions of the Tender]

223 Evaluation Criteria (Other Factors) (ITT 33.6)

The Procuring Entity's evaluation of a Tender may take into account, in addition to the Tender Price quoted in accordance with ITT 13.8, one or more of the following factors as specified in ITT 33.2(d) and in TDS ITT 33.6, using the following criteria and methodologies.

a) Delivery schedule.

The Goods specified in the List of Goods are required to be delivered within the acceptable time range (after the earliest and before the final date, both dates inclusive) specified in Section V, Schedule of Requirements. No credit will be given to deliveries before the earliest date, and Tenders offering delivery after the final date shall be treated as non-responsive. Within this acceptable period, an adjustment of [insert the adjustment factor], will be added, for evaluation purposes only, to the Tender price of Tenders offering deliveries later than the "Earliest Delivery Date" specified in Section V, Schedule of Requirements.

[An adjustment factor of 0.5% per week of delay would be reasonable. However,the adjustment factor should not be more than the rate of Liquidated Damages tobe applied in case of delay in delivery of Goods and Services under the Contract conditions.]

b) Deviation in payment schedule. [insert one of the following]

- i. tenderers shall state their Tender price for the payment schedule outlined in the SCC. Tenders shall be evaluated on the basis of this base price tenderers are, however, permitted to state an alternative payment schedule andindicate the reduction in Tender price they wish to offer for such alternative payment schedule. The Procuring Entity may consider the alternative payment schedule and the reduced Tender price offered by the tenderer selected on the basis of the base price for the payment schedule outlined in the SCC.
- ii. The SCC stipulates the payment schedule specified by the Procuring Entity. If a Tender deviate from the schedule and if such deviation is considered acceptable to the Procuring Entity, the Tender will be evaluated by calculating interest earned for any earlier payments involved in the terms outlined in the Tender as compared with those stipulated in the SCC, at the rate per annum [insert adjustment rate].
- c) Cost of major replacement components, mandatory spare parts, and service. [insert one of the followings]

The list of items and quantities of major assemblies, components, and selected spare parts, likely to be required during the initial period of operation specified in the TDS 15.4, is in the List of Goods. An adjustment equal to the totalcost of these items, at the unit prices quoted in each Tender, shall be added

to the Tender price, for evaluation purposes only.

or

The Procuring Entity will draw up a list of high-usage and high-value items of components and spare parts, along with estimated quantities of usage in the initial period of operation specified in the TDS 15.4. The total cost of these items and quantities will be computed from spare parts unit prices submittedby the tenderer and added to the Tender price, for evaluation purposes only.

or

Tenderer shall provide along with its Tender, the list of recommended spare parts for Goods offered indicating for each item of spare part the recommended quantity and unit, and total CIP final destination prices required during the initial period of operation specified in the TDS 15.4. The prices offered shall not exceed the prevailing prices charged to other parties by the Tenderer. The cost of such spare parts will not be taken into account for tender evaluation. The Procuring Entity may award the contract for spare parts to the Tenderer that is successful for the supply of Goods, by selecting at its option, from the Tender's list of recommended spare parts, such items and quantities against each as the Procuring Entity may deem appropriate at theunit prices indicated by the Tenderer but not exceeding ----% (present) of the cost of Goods [normally not more than 10% or 15%.]

d) **Availability in Kenya** of spare parts and after sales services for equipment offered in the Tender.

An adjustment equal to the cost to the Procuring Entity of establishing the minimum service facilities and parts inventories if quoted separately, shall be added to the Tender price, for evaluation purposes only.

e) Life Cycle Costs

If specified in TDS 33.6, an adjustment to consider the additional life cycle costs for the period specified below, such as the operating and maintenance costs of the Goods, will be added to the Tender price, for evaluation purposes only. The adjustment will be evaluated in accordance with the methodology specified below and the following information:

[Note to Procuring Entity: Life cycle costing should be used when the costs of operation and/or maintenance over the specified life of the goods are estimated to be considerable in comparison with the initial cost and may vary among different Tenders. Life cycle costs shall be evaluated on a net present value basis. If life cycle costs apply, then specify the factors required to determine them for evaluation purposes.

[Either amend the following text as required, or delete if life cycle cost is not applicable]

- i) number of years for life cycle cost determination [insert the number of years of economic life of Goods];
- *ii)* the discount rate to be applied to determine the net present value of the life-cycle-cost is [insert the discount rate];
- iii) the annual operating and maintenance costs (recurrent costs) shall be determined on the basis of the following methodology: [insert methodology E.G. This should include factors that will be used for determination of life-cycle-cost such as costs of operation and maintenance, residual value at the end of economic life of Goods, major elements that will be used for determination of cost of operation and maintenance such as fuel, power, labor, spare parts, etc. unit prices of elements such as fuel, power, etc., quantity of annual usage such as Kms or Hours of operation of Goods, Formula for calculation of LCC, etc];
- iv) and the following information is required from tenderers [insert any information

required from tenderers, including prices e.g. Guaranteed fuel and/or power consumption, cost of labour, spare parts, etc].

f) Performance and productivity of the equipment: [insert one of the followings]

i) Performance and productivity of the equipment. An adjustment representing the capitalized cost of additional operating costs over the life of the goods will be added to the Tender price, for evaluation purposes if specified in the TDS 33.6. The adjustment will be evaluated based on the drop in the guaranteed performance or efficiency offered in the Tender below the norm of 100, using the methodology specified below.

[Insert the methodology and criteria if applicable e.g. The Following aspects could be considered in the formulation of this methodology and criteria: (i) Tender price for the equipment; ii) Price of spare parts required for AAA years of operations, iii) Adjustments to tender price for omissions, deviations and exceptions to technical and commercial conditions in the tender documents; iv) Capitalized cost savings due to the equipment efficiency at the rate of XXX (specify currency and amount) for each YYY % (percent) above the minimum ZZZ % (percent) efficiency; v) Capitalized cost for the auxiliary power consumption at PPP (specify currency and amount) per KW for AAA years; and vi) Applicable discount rate of BBB%.]

 \mathbf{or}

ii) An adjustment to consider the productivity of the goods offered in the Tender will be added to the Tender price, for evaluation purposes only, if specified in ITT 33.6. The adjustment will be evaluated based on the cost per unit of the actual productivity of goods offered in the Tender with respect to minimum required values, using the methodology specified below.

[Insert the methodology and criteria if applicable E.G. The evaluation and comparison of responsive tenders shall be based on the total life cycle cost for XXX years, per unit of output. The life cycle cost shall be the sum of the initial purchase price of the equipment and the cost of operation in electric energy for XXX years of operation at unit cost of AAA (specify currency and amount) per kwh, discounted to net present value at YYY percent.]

g) Specific additional criteria

[Other specific additional criteria to be considered in the evaluation, and the evaluation method shall be detailed in TDS 34.6][If specific **sustainable procurement technical requirements** have been specified in Section VII- Specification, **either** state that (i) those requirements will be evaluated on a pass/fail (compliance basis) **or** otherwise (ii) in addition to evaluating those requirements on a pass/fail (compliance basis), if applicable, specify the monetary adjustments to be applied to Tender Prices for comparison purposeson account of Tenders that exceed the specified minimum sustainable procurement technical requirements.]

224. Multiple Contracts (ITT 33.4)

Multiple contracts will be permitted in accordance with ITT 33.4. Tenderers are evaluated on basis of Lots and the lowest evaluated tenderer identified for each Lot. The Procuring Entity will select one Option of the two Options listed below for awardof Contracts.

OPTION 1

i) If a tenderer wins only one Lot, the tenderer will be awarded a contract forthat Lot, provided the tenderer meets the Eligibility and Qualification Criteria for that Lot.

ii) If a tenderer wins more than one Lot, the tender will be awarded contracts for all won Lots, provided the tenderer meets the aggregate Eligibility and Qualification Criteria for all the Lots. The tenderer will be awarded the combination of Lots for which the tenderer qualifies and the others will be considered for award to second lowest the tenderers.

OPTION 2

The Procuring Entity will consider all possible combinations of won Lots[contract(s)] and determine the combinations with the lowest evaluated price. Tenders will then be awarded to the Tenderer or Tenderers in the combinations provided the tenderer meets the aggregate Eligibility and Qualification Criteria for all the won Lots.

225. Alternative Tenders

(ITT 13.1) An alternative if permitted under

ITT 13.1, will be evaluated as follows: [insert

one of the following]

"A Tenderer may submit an alternative Tender only with a Tender for the base case. The Procuring Entity shall only consider the alternative Tenders offered by the Tenderer whose Tender for the base case was determined to be the Lowest Evaluated Tender." or

"A Tenderer may submit an alternative Tender with or without a Tender for thebase case. The Procuring Entity shall consider Tenders offered for alternatives as specified in the Technical Specifications of Section V, Schedule of Requirements. All Tenders received, for the base case, as well as alternative Tenders meeting the specified requirements, shall be evaluated on their own merits in accordance with the same procedures, as specified in the ITT 33."

8. MARGIN OF PREFERENCE

- 81 If the TDS so specifies, the Procuring Entity will grant a margin of preference of 15% (fifteen percent) to Tenderers offering goods manufactured, mined, extracted, grown, assembled or semi-processed in Kenya. Goods assembled or semi-processed in Kenya shall have a local content of not less than 40%.
- 82 The margin of preference will be applied in accordance with, and subject to, the following provisions:
 - a) Tenderers applying for such preference on goods offered shall provide, as part of the data for qualification, such information, including details of the goods produced in Kenya, so as to determine whether, according to the classification established by the Procuring Entity, a particular category of goods or group of goods qualifies for a margin of preference.
 - b) After Tenders have been received and reviewed by the Procuring Entity, goods offered in the responsive Tenders shall be assessed to ascertain they are manufactured, mined, extracted, grown, assembled or semi-processed in Kenya. Responsive tenders shall be classified into the following groups:
 - i) **Group A:** Tenders offering goods manufactured in Kenya, for which (a) labour, raw materials, and components from within Kenya account for more than forty (40) percent of the Ex-Works price; and (b) the production facility in which they will be manufactured or
 - assembled has been engaged in manufacturing or assembling such goods at least since the
 - date of Tender Submission date:
 - ii) Group B: All other Tenders offering Goods manufactured in Kenya;
 - iii) Group C: Tenders offering Goods manufactured outside Kenya that have

been already imported or that will be imported.

- c) To facilitate this classification by the Procuring Entity, the Tenderer shall complete whichever version of the Price Schedule furnished in the Tender Documents is appropriate. Incorrect classification may render the Tender non- responsive as no reclassification will be permitted after Tender opening. Tenderers shall provide correct information especially with respect to duties, taxes etc. paid on previously imported Goods and percentage of local labour, materials and components for Goods manufactured in Kenya as any false information which cannot be supported by documentation may render the Tender non-responsive besides other sanctions for providing falsifiedinformation.
- d) The Procuring Entity will first review the Tenders to confirm the appropriateness of the Tender group classification to which Tenderers assigned their Tenders in preparing their Tender Forms and Price Schedules.
- e) All evaluated Tenders in each group will then be compared to determine the lowest evaluated Tender of each group. Such lowest evaluated Tenders shall becompared with each other and if as a result of this comparison a Tender from Group A or Group B is the lowest, it shall be selected for the award.
- f) If as a result of the preceding comparison, the lowest evaluated Tender is a Tender from Group C, all Tenders from Group C shall be further compared with the lowest evaluated Tender from Group A after adding to the evaluated price of goods offered in each Tender from Group C, for the purpose of this further comparison only, an amount equal to 15% (fifteen percent) of therespective CIP Tender price for goods to be imported and already imported goods. Both prices shall include unconditional discounts and be corrected for arithmetical errors. If the Tender from Group A is the lowest, it shall be selected for award. If not, the lowest evaluated Tender from Group C shall be selected as per paragraph (e) above."

9. Post-Qualification of Tenderers (ITT 37)

[Note for Procuring Entity to be deleted before issuing the tender documents.

This STD for Procurement of Goods assumes that no Prequalification has taken place before tendering. However, if a Prequalification process is undertaken, the Qualification Criteria stipulated in this Section III, Evaluation and Qualification Criteria must be updated to ensure that the Tenderer and any Sub-Suppliers shall meet or continue to meet the Criteria used at the time of Prequalification.]

9.1 Post-Qualification Criteria (ITT 37.1)

In case the tender <u>was not subject to pre-qualification</u>, the tender that has been determined to be the lowest evaluated tenderer shall be considered for contractaward, subject to meeting each of the following conditions (post qualification Criteriaapplied on a GO/NO GO basis). The Procuring Entity shall carry out the post-qualification of the Tenderer in accordance with ITT 37, using only the requirements specified herein. Requirements not included in the text below shall not be used inthe evaluation of the Tenderer's qualifications. The minimum qualification requirements for multiple contracts will be the sum of the minimum requirements for respective individual contracts, unless otherwise specified.

[Note for Procuring Entity to be deleted before issuing the tender documents.]
Select requirements (criteria) for post qualification from below as relevant and appropriate for the nature, size and type of Goods and Services to be procured. Generally, for procurement of Goods, unless the value of the item is very large, the criteria for assessment of Manufacturer's technical capability should always be considered more important than its financial resources. For very small value items, the criteria for financial capability may even be omitted].

92 If the Tenderer is a manufacturer

a) i)	Financial Capability The Tenderer shall demonstrate that it has access to, or has available, liquid assets, unencumbered real assets, lines of credit, and other financial means (independent of any contractual advance payment) sufficient to meet the supply	
ii)	cash flow of Kenya Shillings	[or
	number of years). In case of multiple contracts, limitation will be placed on the number of item(s) that will be awarded to the Tenderer.	
b)	Experience and Technical Capacity	
expe is a J acco aggr offer mod char- inter demon wor versi the r	Tenderer shall furnish documentary evidence to demonstrate that it meets the following erience requirement(s) using the form provided in Section IV. In case the Tenderer JV, experience and demonstrated technical capacity of only the JV shall be taken into bunt and not of individual members nor their individual experience/capacity will be regated unless all members of the JV have been manufacturing and supplying Goods ared in the Tender to the same technology, processing, design, materials, specifications, lel number, etc. in all respects such that Goods manufactured have the same functional racteristics, performance parameters, outputs and other guarantees and fully rechangeable which shall be documented along with other required documents constrating capacity to the satisfaction of the Procuring Entity in case individual members claim experience. Otherwise, documents evidencing experience and technical acity shall be in the name of the JV that submitted the Tender. Wherever the reds "Similar Goods" have been used it includes upgrades, latest and improved itions or models of similar specifications and technology. Refer to Form Exp-1 to provide required information. The requirement(s), including experience in successfully implementing sustainable curement requirements, if specified in the tender document.] Samples of Experience uirements: The Tenderer shall be manufacturing similar Goods for the last (spec	
ii)	ify the number of years to cover a sufficiently long period ranging from 2 to5 years depending upon the Goods to be procured). The Tenderer shall furnish documentary evidence to demonstrate successful completion of at least (Insert number) of contracts of similar Goods in the last	
	(Insert number) of contracts of similar Goods in the last	
iii)	(Optional) The installed capacity to manufacturenumber of items (<i>specify the relevant item number</i>) shall not be less than units per (<i>specify</i>	
	week or month).	
c)	(Optional) Documentary Evidence of Usage of Goods (When appropriate) The Tenderer shall furnish documentary evidence satisfactory to the Procuring Entity to demonstrate that similar Goods as offered in the Tender have been in successful use or operation for the last	

93 If Tenderer is a Supplier:

If a Tenderer is a Supplier offering the Goods on behalf of or from a Manufacturer

under Manufacturer's Authorization Form (Section IV, Tendering Forms), the Manufacturer shall demonstrate the above qualifications 4.2 (b) (i), (ii), and (iii) and the Tenderer shall demonstrate it meets the following criteria.

i)	unencumbe	er shall dem red real asset advance pay	s, lines of cr	edit, and o	ther financia meet the su	al means (i pply cash	ndepende	
		of		Kenya	S	hillings		
ii)	Minimum	average	annual	supply	turnover	of K	enya Sl	hillings <i>[in</i>
	in progress	t] or equivale and/or compl years] years	eted within		1 .			
iii)	number) co	ctorily and sontract(s) of a or abroad,	a similar na	ture either	within Ken	ya, the Eas	st Africa	n
	a	minimum	valu		in	Kenya		shillings
	equi	ivalent.						
Hist	tory of non-p	erforming c	ontracts:					
	derer (Suppli JV, shall de							
	he default of the in the last_	of the Tende	erer, manuf	acturer or	the member	er of JV a (specify		

95 Pending Litigation

94

Financial position and prospective long-term profitability of the Single Tenderer, and in the case the Tenderer is a JV, of each member of the JV, shall remain sound according to criteria established with respect to Financial Capability under paragraphI (i) above assuming that all pending litigation will be resolved against the Tenderer. Tenderer shall provide information on pending litigations as per Form CON-2.

required information shall be furnished as per form CON-2].

4.6. Litigation History

SECTION IV - TENDERING FORMS

Form of Tender Tenderer Information Form Tenderer JV Members Information Form

Price Schedule: Goods Manufactured Outside Kenya, to be Imported Price Schedule: Goods

Manufactured Outside Kenya, already imported Price Schedule: Goods Manufactured in Kenya

Price and Completion Schedule – Related Services Form of Tender Security – Demand

Guarantee Form of Tender Security (Tender Bond)

Form of Tender-Securing Declaration Manufacturer's Authorization Form

FORM OF TENDER

INSTRUCTIONS TO TENDERERS

- i) The Tenderer must prepare this Form of Tender on stationery with its letterhead clearly showing the Tenderer's complete name and business address.
- ii) All italicized text is to help Tenderer in preparing this form.
- iii) Tenderer must complete and sign CERTIFICATE OF INDEPENDENT TENDER DETERMINATION and the SELF DECLARATION OF THE TENDERER attached to this Form of Tender.

Date of this	1 ender	submission:[insert	aate (as	aay, montn	ana yeai	r) oj Tenaei
submission]	Tender		Name	:		and
Identification	· ·	[insert		identij	fication]	Alternative
No.:		[insert identification	n No if thi	is is a Tend	ler for a	n alternative)
То:		[Insert complete name of	Procuring I	Entity]		

- a) **No reservations:** We have examined and have no reservations to the Tendering document, including Addenda issued in accordance with Instructions to tenderers(ITT 7);
- b) **Eligibility**: We meet the eligibility requirements and have no conflict of interest in accordance with ITT 3;
- c) **Tender/Proposal-Securing Declaration**: We have not been suspended nor declared ineligible by the Procuring Entity based on execution of a Tender-Securing Declaration. or
 - Proposal-Securing Declaration in Kenya in accordance with ITT 3.6;
- d) **Conformity:** We offer to supply in conformity with the Tendering document and in accordance with the Delivery Schedules specified in the Schedule of Requirementsthe following Goods: [insert a brief description of the Goods and Related Services];
- e) **Tender Price**: The total price of our Tender, excluding any discounts offered initem (f) below is:
 - Option 1, in case of one lot: Total price is: [insert the total price of the Tender in words and figures, indicating the various amounts and the respective currencies];

or

Option 2, in case of multiple lots: (a) Total price of each lot [insert the total price of each lot in words and figures, indicating the various amounts and the respective currencies]; and (b) Total price of all lots (sum of all lots) [insert the total price of all lots in words and figures, indicating the various amounts and the respective currencies];

- f) **Discounts**: The discounts offered and the methodology for their application are:
 - i) The discounts offered are: [Specify in detail each discount offered.]
 - ii) The exact method of calculations to determine the net price after application of discounts are shown below: [Specify in detail the method that shall be used to

- g) **Tender Validity Period**: Our Tender shall be valid for the period specified in TDS 17.1 (as amended, if applicable) from the date fixed for the Tender submission deadline specified in TDS 21.1 (as amended, if applicable), and it shall remain binding upon us and may be accepted at any time before the expiration of that period;
- **(h) Performance Security**: If our Tender is accepted, we commit to obtain a performance security in accordance with the Tendering document;
- i) One Tender per tenderer: We are not submitting any other Tender(s) as an individual tenderer, and we are not participating in any other Tender(s) as a Joint Venture member, or as a subcontractor, and meet the requirements of ITT 3.9, other than alternative Tenders submitted in accordance with ITT 12;
- j) Suspension and Debarment: We, along with any of our subcontractors, suppliers, consultants, manufacturers, or service providers for any part of the contract, are not subject to, and not controlled by any entity or individual that is subject to, a temporary suspension or a debarment imposed by the Procuring Entity. Further, we are not ineligible under the Kenya laws or official regulations or pursuant to a decision of the United Nations Security Council;
- k) **State-owned enterprise or institution**: [select the appropriate option and delete the other] [We are not a state- owned enterprise or institution] / [We are a state- owned enterprise or institution but meet the requirements of ITT 3.7];
- l) Commissions, gratuities, fees: We have paid, or will pay the following commissions, gratuities, or fees with respect to the Tendering process or execution of the Contract: [insert complete name of each Recipient, its full address, the reason for which each commission or gratuity was paid and the amount and currency of each such commission or gratuity]

Name of Recipient	Address	Reason	Amount

(If none has been paid or is to be paid, indicate "none.")

- m) **Binding Contract**: We understand that this Tender, together with your written acceptance thereof included in your Letter of Acceptance, shall constitute a binding contract between us, until a formal contract is prepared and executed;
- n) **Procuring Entity Not Bound to Accept**: We understand that you are not bound to accept the lowest evaluated cost Tender, the Best Evaluated Tender or any other Tender that you may receive; and
- o) **Fraud and Corruption**: We hereby certify that we have taken steps to ensure that no person acting for us or on our behalf engages in any type of Fraud and Corruption.
- (p) Code of Ethical Conduct: We undertake to adhere by the Code of Ethics for Persons Participating in Public Procurement and Asset Disposal, copy available from _______(specify website) during the procurement process and the execution of any resulting contract.
- (q) **Collusive practices**: We hereby certify and confirm that the tender is genuine, non-collusive and made with the intention of accepting the contract if awarded. To this effect we have signed the "Certificate of Independent tender Determination" attached

below.

- (r) We, the Tenderer, have completed fully and signed the following Forms as part of our Tender:
 - a) Tenderer's Eligibility; Confidential Business Questionnaire to establish we are not in any conflict to interest.
 - b) Certificate of Independent Tender Determination to declare that we completed the tender without colluding with other tenderers.
 - c) Self-Declaration of the Tenderer to declare that we will, if awarded a contract, not engage in any form of fraud and corruption.
 - d) Declaration and commitment to the Code of Ethics for Persons Participating in Public Procurement and Asset Disposal.

Further, we confirm that we have read and understood the full content and scope of fraud and corruption as informed in "Appendix 1- Fraud and Corruption" attached to the Formof Tender

Name of the tenderer: *[insert complete name of the tenderer]

Name of the person duly authorized to sign the Tender on behalf of the tenderer: **[insert complete name of person duly authorized to sign the Tender]

Title of the person signing the Tender: [insert complete title of the person signing the Tender]

Signature of the person named above: [insert signature of person whose name and capacity are shown above] Date signed [insert date of signing] day of [insert month], [insert year]

- *: In the case of the Tender submitted by a Joint Venture specify the name of the Joint Venture as tenderer.
- **: Person signing the Tender shall have the power of attorney given by the tenderer. The power of attorney shall be attached with the Tender Schedules.

CERTIFICATE OF INDEPENDENT TENDER DETERMINATION

I,	the	undersigned,	1n	submitting	the	accompanying	Letter	of	Tende	er to	the
									[.	Name	of
Pr	ocurir	ng Entity] for:							[Nan	ıe	and
nu	mber	of tender] in	respo	onse to the re	equest	for tenders mad	le by:		[.	Name	of
		r] do hereby m spect:	nake	the following	g stat	ements that I ce	ertify to	be	true and	d com	pletein
	•	on behalf of rer] that:								[N	lame

- 1. I have read and I understand the contents of this Certificate;
- 2. I understand that the Tender will be disqualified if this Certificate is found not to be true and complete in every respect;
- 3. I am the authorized representative of the Tenderer with authority to sign this Certificate, and to submit the Tender on behalf of the Tenderer;
- 4. For the purposes of this Certificate and the Tender, I understand that the word"competitor" shall include any individual or organization, other than the Tenderer, whether or not affiliated with the Tenderer, who:
 - a) has been requested to submit a Tender in response to this request for tenders;
 - b) could potentially submit a tender in response to this request for tenders, based on their qualifications, abilities or experience;
- 5. The Tenderer discloses that [check one of the following, as applicable]:
 - a) The Tenderer has arrived at the Tender independently from, and without consultation, communication, agreement or arrangement with, any competitor;
 - b) the Tenderer has entered into consultations, communications, agreements or arrangements with one or more competitors regarding this request for tenders, and the Tenderer discloses, in the attached document(s), complete details thereof,including the names of the competitors and the nature of, and reasons for, such consultations, communications, agreements or arrangements;
- 6. In particular, without limiting the generality of paragraphs (5)(a) or (5)(b) above, therehas been no consultation, communication, agreement or arrangement with any competitor regarding:
 - a) prices;
 - b) methods, factors or formulas used to calculate prices;
 - c) the intention or decision to submit, or not to submit, a tender; or
 - d) the submission of a tender which does not meet the specifications of the request for Tenders; except as specifically disclosed pursuant to paragraph (5)(b) above;
- 7. In addition, there has been no consultation, communication, agreement or arrangement with any competitor regarding the quality, quantity, specifications or delivery particulars of the works or services to which this request for tenders relates, except as specifically authorized by the procuring authority or as specifically disclosed pursuant to paragraph (5)(b) above;
- 8. the terms of the Tender have not been, and will not be, knowingly disclosed by the Tenderer, directly or indirectly, to any competitor, prior to the date and time of the official tender opening, or of the awarding of the Contract, whichever comes first, unless otherwise required by law or as specifically disclosed pursuant to paragraph(5)(b) above.

Name_			

SELF-DECLARATION FORMS

FORM SD1

SELF DECLARATION THAT THE PERSON/TENDERER IS NOT DEBARRED IN THE MATTER OF THE PUBLIC PROCUREMENT AND ASSET DISPOSAL ACT 2015.

I		of Post Office Box	being
a resi		in the Republic of	
1.	Officer/Director of	(insert name of the Pr	the Company) who is (insert tender
2.		Directors and subcontractors have ceeding under Part IV of the Act.	
3.	THAT what is deponed to herein and belief.	above is true to the best of my	knowledge, information
(Title	······································	(Signature)	(Date)
Bidde	er Official Stamp		

FORM SD2

SELF DECLARATION THAT THE PERSON/TENDERER WILL NOT ENGAGE IN ANY CORRUPT OR FRAUDULENT PRACTICE

of	e a statement as follows:-	of P.O. Box in the Republic	ofbeing a resident do hereby
1.	ofwho is a Bidder in respect of for(In	f Tender No sert tender title/description)	Director/Principal Officer/Director(insert name of the Company) for nd competent to make this statement.
2.	corrupt or fraudulent practi member of the Board, Mana	ice and has not been requestagement, Staff and/or employed	abcontractors will not engage inany sted to pay any inducement toany es and/or agents of which is the procuring entity.
3.		of the Board, Management, S	subcontractors have not offered any Staff and/or employees and/oragents
4.	THAT the aforesaid Bidder w bidders participating in the s		in any corrosive practice with other
5.	THAT what is deponed to h belief.	nerein above is true to the be	st of my knowledge information and
	(Title)	(Signature)	. (Date)

Bidder's Official Stamp

DECLARATION AND COMMITMENT TO THE CODE OF ETHICS

I	(Person) on behalf of (Name
of the Business/ Company/Firm)	declare that I have read and
fully understood the contents of the Public Procurement &	1
the Code of Ethics for persons participating in Public responsibilities under the Code.	Procurement and Asset Disposal and m
I do hereby commit to abide by the provisions of the Code	of Ethics for persons participating in Publi
Procurement and Asset Disposal.	
Name of Authorized signatory	
Sign	
Position	
Office address	Telephone
E-mail	
Name of the Firm/Company	
Date	
(Company Seal/ Rubber Stamp where applicable)	
Witness	
Name	
Sign	
Date	

APPENDIX 1- FRAUD AND CORRUPTION

(Appendix 1 shall not be modified)

1. Purpose

1.1 The Government of Kenya's Anti-Corruption and Economic Crime laws and their sanction's policies and procedures, Public Procurement and Asset Disposal Act (no.33 of 2015) and its Regulation, and any other Kenya's Acts or Regulations related to Fraud and Corruption, and similar offences, shall apply with respect to Public Procurement Processes and Contracts that are governed by the laws of Kenya.

2. Requirements

- The Government of Kenya requires that all parties including Procuring Entities, Tenderers, (applicants/proposers), Consultants, Contractors and Suppliers; any Sub-contractors, Sub-consultants, Service providers or Suppliers; any Agents (whether declared or not); and any of their Personnel, involved and engaged in procurement under Kenya's Laws and Regulation, observe the highest standard of ethics duringthe procurement process, selection and contract execution of all contracts, and refrain from Fraud and Corruption and fully comply with Kenya's laws and Regulations as per paragraphs 1.1 above.
- Kenya's public procurement and asset disposal act (no. 33 of 2015) under Section 66 describes rules to be followed and actions to be taken in dealing with Corrupt, Coercive, Obstructive, Collusive or Fraudulent practices, and Conflicts of Interest in procurement including consequences for offences committed. A few of the provisions noted below highlight Kenya's policy of no tolerance for such practices and behavior:
 - 1) a person to whom this Act applies shall not be involved in any corrupt, coercive, obstructive, collusive or fraudulent practice; or conflicts of interest in any procurement or asset disposal proceeding;
 - 2) A person referred to under subsection (1) who contravenes the provisions of that sub-section commits an offence;
 - 3) Without limiting the generality of the subsection (1) and (2), the person shall be
 - a) disqualified from entering into a contract for a procurement or asset disposal proceeding; or
 - b) if a contract has already been entered into with the person, the contract shall be voidable:
 - 4) The voiding of a contract by the procuring entity under subsection (7) does not limit any legal remedy the procuring entity may have;
 - 5) An employee or agent of the procuring entity or a member of the Board or committee of the procuring entity who has a conflict of interest with respect to a procurement:
 - a) shall not take part in the procurement proceedings;
 - b) shall not, after a procurement contract has been entered into, take part in any decision relating to the procurement or contract; and
 - c) shall not be a subcontractor for the bidder to whom was awarded contract, or a member of the group of bidders to whom the contract wasawarded, but the subcontractor appointed shall meet all the requirements of this Act.
 - 6) An employee, agent or member described in subsection (1) who refrains from doing anything prohibited under that subsection, but for that subsection, would have been within his or her duties shall disclose the conflict of interest to the procuring entity;
 - 7) If a person contravenes subsection (1) with respect to a conflict of interest described in subsection (5)(a) and the contract is awarded to the person or his relative or to another person in whom one of them had a direct or indirect pecuniary interest, the contract shall be terminated and all costs incurred by the public entity shall be made good by the awarding officer. Etc.

- In compliance with Kenya's laws, regulations and policies mentioned above, the Procuring Entity:
 - a) Defines broadly, for the purposes of the above provisions, the terms set forthbelow as follows:
 - i) "corrupt practice" is the offering, giving, receiving, or soliciting, directly or indirectly, of anything of value to influence improperly the actions of another party;
 - ii) "fraudulent practice" is any act or omission, including misrepresentation, that knowingly or recklessly misleads, or attempts to mislead, a party to obtain financial or other benefit or to avoid an obligation;
 - iii) "collusive practice" is an arrangement between two or more parties designed to achieve an improper purpose, including to influence improperly the actions of another party;
 - iv) "coercive practice" is impairing or harming, or threatening to impair or harm, directly or indirectly, any party or the property of the party to influence improperly the actions of a party;
 - v) "obstructive practice" is:
 - deliberately destroying, falsifying, altering, or concealing of evidence material to the investigation or making false statements to investigators in order to materially impede investigation by Public Procurement Regulatory Authority (PPRA) or any other appropriate authority appointed by Government of Kenya into allegations of a corrupt, fraudulent, coercive, or collusive practice; and/or threatening, harassing, or intimidating any party to prevent it from disclosing its knowledge of matters relevant to the investigation or from pursuing the investigation; or
 - acts intended to materially impede the exercise of the PPRA's or the appointed authority's inspection and audit rights provided for under paragraph 2.3 e. below.
 - b) Defines more specifically, in accordance with the above procurement Act provisions set forth for fraudulent and collusive practices as follows:
 - "fraudulent practice" includes a misrepresentation of fact in order to influence a procurement or disposal process or the exercise of a contract to the detriment of the procuring entity or the tenderer or the contractor, and includes collusive practices amongst tenderers prior to or after tender submission designed to establish tender prices at artificial non-competitive levels and to deprive the procuring entity of the benefits of free and open competition.
 - c) Rejects a proposal for award of a contract if PPRA determines that the firm orindividual recommended for award, any of its personnel, or its agents, or its sub-consultants, sub-contractors, service providers, suppliers and/ or their employees, has, directly or indirectly, engaged in corrupt, fraudulent, collusive, coercive, or obstructive practices in competing for the contract in question;
 - d) Pursuant to the Kenya's above stated Acts and Regulations, may sanction or debar or recommend to appropriate authority (ies) for sanctioning anddebarment of a firm or individual, as applicable under the Acts and Regulations;
 - e) Requires that a clause be included in Tender documents and Request for Proposal documents requiring (i) Tenderers (applicants/proposers), Consultants, Contractors, and Suppliers, and their Sub-contractors, Sub-consultants, Service providers, Suppliers, Agents personnel, permit the PPRA or any otherappropriate authority appointed by Government of Kenya to inspect² all accounts, records and other documents relating to the procurement process,

- selection and/or contract execution, and to have them audited by auditors appointed by the PPRA or any other appropriate authority appointed by Government of Kenya; and
- f) Pursuant to Section 62 of the above Act, requires Applicants/Tenderers to submit along with their Applications/Tenders/Proposals a "Self-Declaration Form" as included in the procurement document declaring that they and all parties involved in the procurement process and contract execution have not engaged/will not engage in any corrupt or fraudulent practices.

¹For the avoidance of doubt, a party's ineligibility to be awarded a contract shall include, without limitation, (i) applying for pre-qualification, expressing interest in a consultancy, and tendering, either directly or as a nominated sub-contractor, nominated consultant, nominated manufacturer or supplier, or nominated service provider, in respect of such contract, and (ii) entering into an addendum or amendment introducing a material modification to any existing contract.

Inspections in this context usually are investigative (i.e., forensic) in nature. They involve fact-finding activities undertaken by the Investigating Authority or persons appointed by the Procuring Entity to address specific matters related to investigations/audits, such as evaluating the veracity of an allegation of possible Fraud and Corruption, through the appropriate mechanisms. Such activity includes but is not limited to: accessing and examining a firm's or individual's financial records and information, and making copies thereofas relevant; accessing and examining any other documents, data and information (whether in hard copy or electronic format) deemed relevant for the investigation/audit, and making copies thereof as relevant; interviewing staff and other relevant individuals; performing physical inspections and site visits; and obtaining third party verification of information.

TENDERER INFORMATION FORM

[The tenderer shall fill in this Form in accordance with the instructions indicated below. No alterations to its format shall be permitted and no substitutions shall be accepted.]
Date: [insert date (as day, month and year) of Tender submission]
Tender Name and Identification[Insert identification
Alternative No.: [insert identification No if this is a Tender for an
alternative] Pageofpages
1. Tenderer's Name [insert Tenderer's legal name]
2. In case of JV, legal name of each member: [insert legal name of each member in JV]
3. Tenderer's actual or intended country of registration: [insert actual or intended country of registration]
4. Tenderer's year of registration: [insert Tenderer's year of registration]
5. Tenderer's Address in country of registration: [insert Tenderer's legal address in country of registration]
6. Tenderer's Authorized Representative Information
Name: [insert Authorized Representative's name]
Address: [insert Authorized Representative's Address]
Telephone/Fax numbers: [insert Authorized Representative's telephone/fax numbers]
Email Address: [insert Authorized Representative's email address]
7. Attached are copies of original documents of [check the box(es) of the attached original documents]
☐ For Kenyan Tenderers a current tax clearance certificate or tax exemption certificate issued by the Kenya Revenue Authority in accordance with ITT 3.14.
□Articles of Incorporation (or equivalent documents of constitution or association), and/or documents of registration of the legal entity named above, in accordance with ITT 3.4.
☐ In case of JV, letter of intent to form JV or JV agreement, in accordance with ITT 3.1.
☐ In case of state-owned enterprise or institution, in accordance with ITT 4.6 documents establishing:
(i) Legal and financial autonomy
(ii) Operation under commercial law
(iii) Establishing that the tenderer is not under the supervision of the Procuring Entity
2. Included are the organizational chart, a list of Board of Directors, and the beneficial ownership.

TENDERER'S ELIGIBILITY- CONFIDENTIAL BUSINESS QUESTIONNAIRE FORM

a) Instruction to Tenderer

Tender is instructed to complete the particulars required in this Form, *one form for eachentity if Tender is a JV*. Tenderer is further reminded that it is an offence to give false information on this Form.

A. Tenderer's details

	ITEM	DESCRIPTION
1	Name of the Procuring Entity	
2	Name of the Tenderer	
3	Full Address and Contact Details of the Tenderer.	 Country City Location Building Floor Postal Address Name and email of contact person.
4	Reference Number of the Tender	
5	Date and Time of Tender Opening	
6	Current Trade License No and Expiring date	
7	Maximum value of business which the Tenderer handles.	
8		_

General and Specific Details		
Sole Proprietor, provide the following details.		
Name in full		
AgeNationality		
Country of OriginCitizenship		
c) Partnership, provide the following details.		
Names of Partners Nationality Citizenship % Shares owned		
3		

ii) State the nominal and issued capital of the Company-

Nominal Kenya Shillings (Equivalent)	
Issued Kenya Shillings (Equivalent)	

iii) Give	details	of Direct	ors as follows

	Names of Director	Nationality	Citizenship	%	Shares
				owned	
1					
2					
3					

- (e) DISCLOSURE OF INTEREST- Interest of the Firm in the Procuring Entity.

If yes, provide details as follows.

	Names of Person	Designation in the Procuring Entity	Interest or Relationship with Tenderer
1			
2			
3			

(ii) Conflict of interest disclosure

	Type of Conflict	Disclosure YES OR NO	If YES provide details of the relationship with Tenderer
1	Tenderer is directly or indirectly controlled by or is under common control with another tenderer.		
2	Tenderer receives or has received any direct or indirect subsidy from another tenderer.		
3	Tenderer has the same legal representative as another Tenderer		
4	Tender has a relationship with another tenderer, directly or through common third parties that puts it in a position to influence the tender of another tenderer, or influence the decisions of the Procuring Entity regarding this tendering process.		
5	Any of the Tenderer's affiliates participated as a consultant in the preparation of the design or technical specifications of the works that are the subject of the tender.		
6	Tenderer would be providing goods, works, non-consulting services or consulting services during implementation of the		

	Type of Conflict	Disclosure YES OR NO	If YES provide details of the relationship with Tenderer
	contract specified in this Tender Document.		
7	Tenderer has a close business or family relationship with a professional staff of the Procuring Entity who are directly or indirectly involved in the preparation of the Tender document or specifications of the Contract, and/or the Tender evaluation process of such contract.		
8	Tenderer has a close business or family relationship with a professional staff of the Procuring Entity who would be involved in the implementation or supervision of the Contract.		
9	Has the conflict stemming from such relationship stated in item 7 and 8 above been resolved in a manner acceptable to the Procuring Entity throughout the tendering process and execution of the Contract?		

10	\sim			
(f)	(Cer	•t1 †1	LC at	nn
111		LIII	uai	11/11

On behalf of the Tenderer, I certify that the information	ation given above is correct.
Full Name	
Fitle or Designation	
Signature)	(Date)

TENDERER'S JV MEMBERS INFORMATION FORM

[The tenderer shall fill in this Form in accordance with the instructions indicated below. The following table shall be filled in for the tenderer and for each member of a Joint Venture]]. Tender Name and Identification: [insert identification Alternative No:.....[insert identification No if this is a Tender for an alternative]. Page of pages Tenderer's Name: [insert Tenderer's legal name] 2. Tenderer's JV Member's name: [insert JV's Member legal name] 3. Tenderer's JV Member's country of registration: [insert JV's Member country of registration] 4. Tenderer's JV Member's year of registration: [insert JV's Member year of registration] 5. Tenderer's JV Member's legal address in country of registration: finsert JV's Member legal address in country of registration] Tenderer's JV Member's authorized representative information Name: [insert name of JV's Member authorized representative] Address: [insert address of JV's Member authorized representative] Telephone/Fax numbers: [insert telephone/fax numbers of JV's Member authorized representative] Email Address: [insert email address of JV's Member authorized representative] 7. Attached are copies of original documents of [check the box(es) of the attached original documents] Articles of Incorporation (or equivalent documents of constitution or association), and/or registration documents of the legal entity named above, in accordance with ITT 4.4. In case of a state-owned enterprise or institution, documents establishing legal and financial autonomy, operation in accordance with commercial law, and that they are not under the supervision of the Procuring Entity, in accordance with ITT 4.6. 8. Included are the organizational chart, a list of Board of Directors, and the beneficial ownership.

Price Schedule Forms

[The tenderer shall fill in these Price Schedule Forms in accordance with the instructions indicated. The list of line items in column 1 of the **Price Schedules** shall coincide withthe List of Goods and Related Services specified by the Procuring Entity in the Schedule of Requirements.]

Price Schedule: Goods Manufactured Outside Kenya, to be Imported

					(Group C Tenders, goods to be imported) Currencies in accordance with ITT 15			Date: ITT No:			
								Alternative No: of			
1	2		3	4	5	6	7	8	9		
Line	Description	of	Country	Delivery	Quantity	Unit price	CIP Price per	Price per line item for	Total Price per Line		
Item	Goods		of Origin	Date as	and physical	CIP [insert place of	line item	inland transportation and	item		
N°				defined	Unit	destination]	(Col. 5x6)	other services required in	(Col. 7+8)		
				by		in accordance with ITT		Kenya to convey the			
				Incoterm		14.8(b)(i)		Goods to their final			
				S				destination specified in			
								TDS			
[inser	[insert name	of	[insert	[insert	[insert	[insert unit price CIP	[insert total	[insert the corresponding	[insert total price of the		
t	good]		country	quoted	number of	per unit]	CIP price per	price per line item]	line item]		
numb			of origin	Delivery	units to be		line item]				
er of			of the	Date]	Supplied						
the			Good		and name of						
item]					the physical						
					unit]						
,								Total Price			

Name of tenderer [insert complete name of tenderer] Signature of tenderer [signature of person signing the Tender] Date [Insert Date]

Price Schedule: Goods Manufactured Outside Kenya, already imported*

				enders, Goods al in accordance wi	th ITT 15			Date:ITT No:Alternative No:of			
1	2	3	4	5	6	7	8	9	10	11	12
Line Item N°	Descripti on of Goods	Country of Origin	Delivery Date as defined by Incoterms	Quantity and physical unit	Unit price including Custom Duties and Import Taxes paid, in accordance with ITT 14.8(c)(i)	and Import Taxes paid per unit in accordance with ITT 14.8(c)(ii), [to	duties and import taxes, in accordance with ITT 14.8 (c) (iii) (Col. 6 minus	item net of Custom Duties and Import Taxes paid, in accordance	Price per line item for inland transportation and other services required in Kenya to convey the goods to their final destination, as specified in TDS in accordance with ITT 14.8 (c)(v)	Sales and other taxes paid or payable peritem if Contract is awarded (in accordance with ITT 14.8(c)(iv)	per line item
[insert number of the item]	[insert name of Goods]	[insert country of origin of the Good]	[insert quoted Delivery Date]	[insert number of units to be supplied and name of the physical unit]	price per		price net of custom duties	per line item net of custom	item for inland transportation and other services required	other taxes payable per item if Contract is awarded]	price per line
										Total Tender Price	

Name of tenderer [insert complete name of tenderer] Signature of tenderer [signature of person signing the Tender] Date [insert date]

^{* [}For previously imported Goods, the quoted price shall be distinguishable from the original import value of these Goods declared to customs and shall include any rebate or mark-up of the local agent or representative and all local costs except import duties and taxes, which have been and/or have to be paid by the Procuring Entity. For clarity, the tenderers are asked to quote the price including import duties, and additionally to provide the import duties and the price net of import duties which is the difference of those values.]

Price Schedule: Goods Manufactured in Kenya

Kenya				d B Tenders) n accordance v	Date: ITT No: Alternative No: Page N° of				
Line Item N°	2 Description of Goods	Delivery Date as defined by Incoterms	Quantit y and physica l unit	5 Unit price EXW	Total EXW price per line item (Col. 4×5)	for inland transportation and	Cost of local labor, raw materials and components from with origin in Kenya % of Col. 5	Sales and other taxes payable per line item if Contract is awarded (in accordance with ITT 14.8(a)(ii)	-
[insert numbe r of the item]	[insert name of Good]	[insert quoted Delivery Date]	[insert number of units to be supplie d and name of the physica l unit]		[insert total EXW price per line item]	corresponding price	[Insert cost of local labor, raw material and components from within the Purchase's country as a % of the EXW price per line item]	taxes payable per line	
								Total Price	

Name of tenderer [insert complete name of tenderer] Signature of tenderer [signature of person signing the Tender] Date [insert date]

Price and Completion Schedule - Related Services

	Date:					
					ĪTT	No
					Alternative	No
					Page N°	of
1	2	3	4	5	6	7
Service N°	Description of Services (excludes inland transportation and other services required in Kenya to convey the goods to their final destination)	Country of Origin	Delivery Date at place of Final destination	Quantity and physical unit	Unit price	Total Price per Service (Col. 5*6 or estimate)
[insert number of the Service]	[insert name of Services]	[insert country of origin of The Services]	[insert delivery date at place of final destination per Service]	[insert number of units to be supplied and name of the physical unit]	[insert unit price per item]	[insert total price per item]
T-		•	•	Total Tender Price		

Name of tenderer [insert complete name of tenderer] Signature of tenderer [signature of person signing the Tender] Date [insert date]

Be	neficiary:			
Request for Tenders No:				
Da	te:			
	ENDER GUARANTEE No.:			
Gu	arantor:			
1.	We have been informed that(here inafter called "the Applicant") has submitted of will submit to the Beneficiary its Tender (here inafter called" the Tender") for the execution of under Request for Tenders No("the ITT").			
2.	Furthermore, we understand that, according to the Beneficiary's conditions, Tenders must be supported by a Tender guarantee.			
3.	At the request of the Applicant, we, as Guarantor, hereby irrevocably undertake to pay the Beneficiary any sum or sums not exceeding in total an amount of() upon receipt by us of theBeneficiary's complying demand, supported by the Beneficiary's statement, whether in the demand itself or a separate signed document accompanying or identifying the demand, stating that either the Applicant:			
(a)	has withdrawn its Tender during the period of Tender validity set forth in the Applicant's Letter of Tender ("the Tender Validity Period"), or any extension thereto provided by the Applicant; or			
b)	having been notified of the acceptance of its Tender by the Beneficiary during the Tender Validity Period or any extension there to provided by the Applicant, (i) has failed to execute the contract agreement, or (ii) has failed to furnish the Performance.			
4.	This guarantee will expire: (a) if the Applicant is the successful Tenderer, upon our receipt of copies of the contract agreement signed by the Applicant and the Performance Security and, or (b) if the Applicant is not the successful Tenderer, upon the earlier of (i) our receipt of a copy of the Beneficiary's notification to the Applicant of the results of the Tendering process; or (ii) thirty days after the end of the Tender Validity Period.			
5.	Consequently, any demand for payment under this guarantee must be received by us at the office indicated above on or before that date.			
	[signature(s)]			

FORM OF TENDER SECURITY-[Option 1–Demand Bank Guarantee]

Note: All italicized text is for use in preparing this form and shall be deleted from the final product.

FORMAT OF TENDER SECURITY [Option 2–Insurance Guarantee]

the Procuring Entity's first written of mand, provided that in its demand the currence of any of the above events, so is guarantee will expire: (a) if the Applicant agreement signed by the Applicant of the results of the ander Validity Period.	demand, without the Procuring Entity having to substantiate its ne Procuring Entity shall state that the demand arises from the pecifying which event(s) has occurred. Splicant is the successful Tenderer, upon our receipt of copiesor licant and the Performance Security and, or (b) if the Applicant earlier of (i) our receipt of a copy of the Beneficiary's notification Tendering process; or (ii)twenty-eight days after the end of the inder this guarantee must be received by us at the office indicated and the Islandar of the Guarantee of the Guaran
the Procuring Entity's first written of mand, provided that in its demand the currence of any of the above events, so its guarantee will expire: (a) if the Applicant agreement signed by the Applicant of the results of the hader Validity Period.	demand, without the Procuring Entity having to substantiate its ne Procuring Entity shall state that the demand arises from the pecifying which event(s) has occurred. Tenderer, upon our receipt of copiesof licant and the Performance Security and, or (b) if the Applicant earlier of (i) our receipt of a copy of the Beneficiary's notification Tendering process; or (ii)twenty-eight days after the end of the
the Procuring Entity's first written of mand, provided that in its demand the currence of any of the above events, so its guarantee will expire: (a) if the Applicant agreement signed by the Applicant of the results of the	demand, without the Procuring Entity having to substantiate its ne Procuring Entity shall state that the demand arises from the pecifying which event(s) has occurred. Splicant is the successful Tenderer, upon our receipt of copieson licant and the Performance Security and, or (b) if the Applicant earlier of (i) our receipt of a copy of the Beneficiary's notification
the Procuring Entity's first written of mand, provided that in its demand the currence of any of the above events, so is guarantee will expire: (a) if the Ap	demand, without the Procuring Entity having to substantiate its ne Procuring Entity shall state that the demand arises from the pecifying which event(s) has occurred. Splicant is the successful Tenderer, upon our receipt of copieson
the Procuring Entity's first written command, provided that in its demand the	lemand, without the Procuring Entity having to substantiate its ne Procuring Entity shall state that the demand arises from the
n the guarantee undertakes to immediatel	y pay to the Procuring Entity up to the above amount upon receipt
Period or any extension thereto pro	e of its Tender by the Procuring Entity during the Tender Validity vided by the Principal; (i) failed to execute the Contract agreement formance Security, in accordance with the Instructions to tenderers tendering document.
	he period of Tender validity set forth in the Principal's Letter of d''), or any extension thereto provided by the Principal; or
OW, THEREFORE, THE CONDITION	OF THIS OBLIGATION is such that if the Applicant:
aled with the Common Seal of the said G	buarantor thisday of20
ving our registered office at	t WE
	der] for the
nereas [Name of the tenderer]	(hereinafter called "the tenderer") has submitted its tender
ne er r r r r r r r r r r r r r r r r r	reinafter called "the Tender") for the exeminater called "the Tender of Procuring Entity] (Instruction of Procuring Entity) (Instruction of Entity) (Instruction of Entity), the Guarantor binds itself carried with the Common Seal of the said Gowy of the Sents. When the Tender of State of the Sent of the Sent of the Ender of the Tender of the Ender of

Note: All italicized text is for use in preparing this form and shall be deleted from the final product.

FORM OF TENDER-SECURING DECLARATION

Seal or stamp.

[The Bidder shall complete this Form in accordance with the instructions indicated]		
Date:[insert date (as day, month and year) of Tender Submission]		
Tender No.:		
To:[insert complete name of		
Purchaser] I/We, the undersigned, declare that:		
1. I/We understand that, according to your conditions, bids must be supported by a Tender-Securing Declaration.		
I/We accept that I/we will automatically be suspended from being eligible for tendering in any contract with the Purchaser for the period of time of[insert number of months or years] starting on[insert date], if we are in breach of our obligation(s) under the bid conditions, because we – (a) have withdrawn our tender during the period of tender validity specified by usin the Tendering Data Sheet; or (b) having been notified of the acceptance of our Bid by the Purchaser during the period of bid validity, (i) fail or refuse to execute the Contract, if required, or (ii) fail or refuse to furnish the Performance Security, in accordance with the instructions to tenders.		
3. I/We understand that this Tender Securing Declaration shall expire if we are not the successful Tenderer(s), upon the earlier of:		
a) our receipt of a copy of your notification of the name of the successful Tenderer; or		
b) thirty days after the expiration of our Tender.		
4. I/We understand that if I am/we are/in a Joint Venture, the Tender Securing Declaration must bein the name of the Joint Venture that submits the bid, and the Joint Venture has not been legally constituted at the time of bidding, the Tender Securing Declaration shall be in the names of all future partners as named in the letter of intent.		
Signed:		
Capacity / title (director or partner or sole proprietor, etc.)		
Name:		
Duly authorized to sign the bid for and on behalf of:[insert complete name of		
Tenderer]. Dated on		
signing].		

MANUFACTURER'S AUTHORIZATION FORM

[The tenderer shall require the Manufacturer to fill in this Form in accordance with the instructions indicated. This letter of authorization should be on the letterhead of the Manufacturer and should be signed by a person with the proper authority to sign documents that are binding on the Manufacturer. The tenderer shall include it in its Tender, if so indicated in the **TDS.**]

Date:	[inse	ert date (as day, n	nonth and year)	of Tender su	bmission]	
ITT	No.:	[insert	number	of	ITT	
process]	Alternative No	.:[inse	ert identification l	No if this is a		
Tender j	for an alternat	ive]				
То:	[Inse	rt complete name of	Procuring			
Entity] \	WHEREAS					
type of g authorized following	goods manufac E [insert compl g Goods, manu	tured], having fact ete name of tendere	ories at [insert er] to submit a[inser	full address Tender the	of Manufacturer's purpose of which	factories], do hereby is to provide the of the Goods], and
		all guarantee and wa with respect to th				al
Signed	[1	Insert signature(s) o	f authorized repre	esentative(s) o	of the Manufacture.	r]
Name:	[In	sert complete name	e(s) of authorized	d representai	tive(s) of the Man	ufacturer]
Title:	[Ins	sert title]				
Dated on	1	_day of	•	[insert date	e of signing]	



Section V - Schedule of Requirements

1. List of Goods and Delivery Schedule

[The Procuring Entity shall fill in this table, with the exception of the column "Tenderer's offered Delivery date" to be filled by the tenderer]

	PAP ONDITI MEDICAL EQUIPMENT LIST			
S.NO.	EQUIPMENT (AS PER SPECIFICATIONS)	QUANTITY	UNIT PRICE	TOTAL PRICE
1	OPERATING THEATRE TABLE	1		2
2	OPERATING THEATRE LIGHT	2		
3	ANAESTHETIC MACHINE WITH VENTILLATOR	1		
4	Crash Cart & Defibrillator	1		
5	ELECTROSURGICAL DIATHERMY MACHINE	1		
6	INFANT INCUBATOR	2		
7	INFANT RESCUTCITAIRE	1		
8	Phototherapy Unit	1		
9	Space Heater	3		
10	Oxygen splitter,5way,2ltrs PM	2		
11	PATIENT MONITOR	2		
12	Fetal Doppler Ultrasound Machine	1		
13	Fetal Doppler, Handheld	2		
14	Oxygen Therapy(regulator)apparatus;Gauge,flowmeter ,Humidifier Bottle	4		
15	Wall Oxygen Flowmeter (Dual)	4		
16	Oxygen Concentrator s	1		
17	Adult Weighing Scale with heightometre	1		
18	Infant Weighing Scale	1		
19	Suction Machine, Electrical	2		

20	Suction Machine, Manual	2	
21	Examination Lamp	3	
22	Digital Blood Pressure machine	3	
23	Nebulizer Machine	2	
24	Syringe Infusion Pump	4	
25	Pulse Oximeter	3	
26	Ambu bag adult	2	
27	Ambu bag Paediatrics	2	
28	Delivery Set	3	
29	Ceasarian Set	3	
30	Patient adjustable beds, Single crank, complete with mankintosh matress (for the wards)	15	
31	Bed Side locker	3	
32	Delivery Set, Hydraulic	5	
33	Baby Cot	6	
34	Examination Couch Metal	2	
35	Trolley Instrument	2	
36	Trolley ,Dressing	2	
37	Trolley ,Gas Cylinder	2	
38	Trolley ,General Purpose	2	
39	Trolley,Emergency	2	
40	Trolley ,Medicine /Drug	2	
41	Trolley ,Patient Stretcher	2	
42	Wheel Chair Adult		

2. List of Related Services and Completion Schedule

[This table shall be filled in by the Procuring Entity. The Required Completion Dates should be realistic, and consistent with the required Goods Delivery Dates (as per Incoterms)].

Service	Description of Service	Quantity ¹	Physical Unit	Place where Services shall be performed	Final Completion Date(s) of Services
[insert Service No]	[insert description of Related Services]	[insert quantity of items to be supplied]	[insert physical unit for the items]	[insert name of the Place]	[insert required Completion Date(s)]

¹If applicable

3. Technical Specifications

- 3.1 The purpose of the Technical Specifications (TS), is to define the technical characteristics of the Goods and Related Services required by the Procuring Entity. The Procuring Entity shall prepare the detailed TS consider that:
 - i) The TS constitute the benchmarks against which the Procuring Entity will verify the technical responsiveness of Tenders and subsequently evaluate the Tenders. Therefore, well-defined TS will facilitate preparation of responsive Tenders by tenderers, as well as examination, evaluation, and comparison of the Tenders by the Procuring Entity.
 - ii) The TS shall require that all goods and materials to be incorporated in the goods be new, unused, and of the most recent or current models, and that they incorporate all recent improvements in design and materials, unless provided for otherwise in the contract.
 - iii) The TS shall make use of best practices. Samples of specifications from successful similar procurements in the same country or sector may provide a sound basis for drafting the TS.
 - iv) The PPRA encourages the use of metric units.
 - v) Standardizing technical specifications may be advantageous, depending on the complexity of the goods and the repetitiveness of the type of procurement. Technical Specifications should be broad enough to avoid restrictions on workmanship, materials, and equipment commonly used in manufacturing similar kinds of goods.
 - vi) Standards for equipment, materials, and workmanship specified in the Tendering document shall not be restrictive. Recognized international standards should be specified as much as possible. Reference to brand names, catalogue numbers, or other details that limit any materials or items to a specific manufacturer should be avoided as far as possible. Where unavoidable, such item description should always be followed by the words "or substantially equivalent." When other particular standards or codes of practice are referred to in the TS, whether from the Procuring Entity's or from other eligible countries, a statement should follow other authoritative standards that ensure at least a substantially equal quality, then the standards mentioned in the TS will also be acceptable.
 - vii) Reference to brand names and catalogue numbers should be avoided as far as possible; where unavoidable the words "or at least equivalent" shall always follow such references.
 - viii) Technical Specifications shall be fully descriptive of the requirements in respect of, but not limited to, the following:
 - a) Standards of materials and workmanship required for the production and manufacturing of the Goods.
 - b) Any sustainable procurement technical requirements shall be clearly specified.
- 3.2 To encourage tenderers' innovation in addressing sustainable procurement requirements, as long as the Tender evaluation criteria specify the mechanism for monetary adjustments for the purpose of Tender comparisons, tenderers may be invited to offer Goods that exceeds the specified minimum sustainable procurement requirements.
 - i) Detailed tests required (type and number).
 - ii) Other additional work and/or Related Services required to achieve full delivery/completion.
 - iii) Detailed activities to be performed by the Supplier, and participation of the Procuring Entitythereon.
 - iv) List of detailed functional guarantees covered by the Warranty and the specification of the liquidated damages to be applied in the event that such guarantees are not met.
- 3.3 The TS shall specify all essential technical and performance characteristics and requirements, including guaranteed or acceptable maximum or minimum values, as appropriate. Whenever necessary, the Procuring Entity shall include an additional ad-hoc Tendering form (to be an Attachment to the Letter of Tender), where the tenderer shall provide detailed information on such

- technical performance characteristics in respect to the corresponding acceptable or guaranteedvalues.
- 3.4 When the Procuring Entity requests that the tenderer provides in its Tender a part or all of the Technical Specifications, technical schedules, or other technical information, the Procuring Entity shall specify in detail the nature and extent of the required information and the manner in whichit has to be presented by the tenderer in its Tender.
- 3.5 If a summary of the Technical Specifications(TS) has to be provided, the Procuring Entity shall insert information in the table below. The tenderer shall prepare a similar table to justify compliance with the requirements.

Summary of Technical Specifications: The Goods and Related Services shall comply with following Technical Specifications and Standards:

OPERATING THEATRE TABLE

1. OPERATING THEATRE TABLE WITH ACCESSORIES

ITEM DESCRIPTION

1.General Description Operating table suitable for use in theatre for major operations. It should be capable of performing lateral tilt, up-down movement, Trendelenburg and reverse Trendelenburg position, and back section refraction. The movement should be electrohydraulic with manual option control system). 2. Composition 2.1 Main Unit 2.2 Accessories 3. Performance specifications 3.1 **Main Unit** 3.1.1 Table Top - Approx. Length 2000 X width 600 mm - X-ray Permeable - Detachable 3.1.2 Head Rest 3.1.3 Leg Rests - Detachable / Separable - Made of scratch resistant, hard wearing and easy to clean 3.2 Material of main unit material 3.3 Height of table top - Adjustable, mechanical operated, min 600mm to 1100mm 3.4 Table top movements 3.4.1. Trendelenburg - Forward: 25°, Reverse: 25° 3.4.2 Lateral – tilt $- \sim 20^{\circ}$ both to the left and right 3.4.3 Back- section refraction - 90° 4.4 Table top turn - 180° 3.5 Main Unit Movement - Mobile with antistatic castors with braking mechanism 3.6 Maximum Load Weight - 250Kg 4. Accessories - High density type easy to clean, 3" thickness with 4 4.1 Mattress sections, breathable, waterproof that doesn't stick to the table. 4.2 Arm board with mattress 4.3 Shoulder support with pad 4.4 Foot board 4.5 Knee crutches 4.6 Screen frame 4.7 Body support with pads

4.8	I.V. pole, height adjustable	
4.9	Orthopedic attachment	
5.	Operating Environment	
5.1	Power requirements	- 240V A/C 50Hz, 13A, 3 pin plug, single phase, 5m long
	-	cable with PE
5.2	Ambient temperature	- 10 to 40°C
5.3	Relative Humidity	- 20% to 90%, No dew Condensation
6.	Quality Standards	
6.1	Manufacturing Standards	- ISO certified.
6.2	Conformity to Standards	- CE, FDA approved and marked.
7.	Local back up service	
7.1	Local agent	- Should be locally available and reliable.
7.2	Capacity to service equipment	- Agent shall have adequate facilities, spare parts, qualified
		and Skilled technical staff.
8.	Delivery point	
8.1	Inspection and Testing	- At County Hospital
9.	Pre installation requirements	- Nil
10.	Installation and testing	
10.1	1	- At the site advised, as per manufacturer's instructions.
	set-up of the equipment	
11.	Training	
	User training	- On Operation, User Maintenance and daily upkeep.
11.2	Maintenance training	- Technical training on preventive maintenance and fault-
10		finding.
12.	Technical documentations	
12.1		- English language, Printed and Soft, 1 set
12.2		- English language, Printed and Soft, 1 set
13.	Commissioning	
13.1	Testing and commissioning	- To the satisfaction of the client
14.	Warranty	(NB: Provide Warranty card)
14.1	Equipment	- Minimum of Two (2) year after commissioning on all parts

2. OPERATING THEATRE LIGHT

ITEM DESCRIPTION	OPERATING THEATRE LIGHT
	type. The surgical light should consist of two lamp head, ructed from light weight material preferable aluminum, ackup power supply to last for at least 2 hours.
2. Composition2.1 Main and Auxillary lamp head2.2 Antisurge protector - As	per Manufacturer's specification, 100V-250Vac

3.1.1 Dimeter	3. P	Performance specifications	
3.1.2 Rotation 3.1.3 Maximum Light Intensity 3.1.4 Focus 3.1.5 Field 3.1.6 Field 3.1.7 Light Colour Temperature 3.1.8 Lighting Control 3.1.9 Lighting Bulb 3.1.10 Lighting Bulb 3.1.11 Mounting Ceiling Height 3.1.11 Mounting Ceiling Height 3.1.12 Physical Characteristics 4.1 Physical Characteristics 4.1 Main and Auxilliary Units 4. Physical Characteristics 4.1 Main and Auxilliary Units 5. Operating Environment 5.1 Power requirements 5.2 Ambient temperature 5.3 Relative Humidity 5. Quality Standards 6.1 Manufacturing Standards 6.2 Product Conformity Standards 6.3 Delivery point 8.1 Inspection and Testing 6.1 Inspection and Testing 6.2 Pre installation requirements 7.1 Pre installation and testing 6.2 Installation and testing 6.3 Installation and testing 6.4 Installation and testing 6.5 Conjour ment erral axis 6.7 Adjustable 6.8 Conjour tender at least 500mm 6. Adjustable 6.9 Control mounted at least 500mm 6. Above 150,000 lux at 1 meter each 6.1 Adjustable 6.2 Constant to a depth of at least 500mm 6. Adjustable 6.2 Constant to a depth of at least 500mm 6. Adjustable 6.2 Constant to a depth of at least 500mm 6. Adjustable 6.2 Control mounted at ca convenient place preferable on the head lamp 6.2 Light field diameter of 300mm at 1M 6.3 Minimum 2.5M above floor 6. Automatic changeover and charger unit 6. Physical Characteristics 6. Product Conformity Standards 6. Product Conformity Standar	3.1	Main and Auxilliary lamp head.	
3.1.3 Maximum Light Intensity 3.1.4 Focus 3.1.5 Field 3.1.6 Field 3.1.7 Light Colour Temperature 3.1.8 Lighting Control 3.1.9 3.1.10 Lighting Bulb 3.1.11 Mounting Ceiling Height 3.2 Accessories 3.3 Emergency Backup Power 3.3.1 Maximum Light Intensity 3.3.1 Field 3.3.1 Maximum Light Intensity 3.3.1 Field 3.4 Characteristics 4.1 Main and Auxilliary Units 5. Operating Environment 5.1 Power requirements 5.2 Ambient temperature 5.3 Relative Humidity 5. Quality Standards 6.1 Manufacturing Standards 6.2 Product Conformity Standards 6.3 Product Conformity Standards 6.4 Capacity to service equipment 6.5 Delivery point 8.1 Inspection and Testing 6.1 Installation and testing 6.2 Installation and testing 6.3 Installation and testing 6.4 Installation and testing 6.5 Operating Environment 7.6 Complete installation and destring 6.7 Pre installation and testing 6.8 Delivery point 6.9 Pre installation and testing 6.1 Installation and testing 6.2 Complete installation and destring 6.3 Installation and testing 6.4 Complete installation and set-up of the light at permanufacturer's instructions in accordance with the Acceptance Testing Procedures.	3.1.1	Diameter	- main and auxiliary unit
3.1.4 Focus 3.1.5 Field 3.1.6 Field 3.1.7 Light Colour Temperature 3.1.8 Lighting Control 3.1.9 3.1.10 Lighting Bulb 3.1.11 Mounting Ceiling Height 3.2 Accessories 3.3 Emergency Backup Power 3.3.1 Emergency Backup Power 3.3.1 Physical Characteristics 4.1 Main and Auxilliary Units 5. Operating Environment 5.1 Power requirements 5.2 Ambient temperature 5.3 Relative Humidity 5.4 Product Conformity Standards 6.5 Product Conformity Standards 6.7 Local back up service 7.1 Local agent 7.2 Capacity to service equipment 8.1 Inspection and Testing 7.2 Pre installation requirements 8.1 Delivery point 8.1 Installation and testing 10.1 Complete installation and secupe of the light at per manufacturer's instructions in accordance with the Acceptance Testing Procedures. 11. Training	3.1.2	Rotation	- 360° along the central axis
3.1.5 Field 3.1.6 Field 3.1.7 Light Colour Temperature 3.1.8 Lighting Control 3.1.9 Shadowless 3.300 to 4800 K 3.1.9 - Deeming range 30-100% 3.1.10 Lighting Bulb 3.1.11 Mounting Ceiling Height 3.2 Accessories 3.3 Emergency Backup Power 3.3.1 For at least 2 hours 3.3.1 Physical Characteristics 4.1 Main and Auxilliary Units 5. Operating Environment 5.1 Power requirements 5.2 Ambient temperature 5.3 Relative Humidity 5.4 Product Conformity Standards 6.6 Quality Standards 6.7 Local back up service 7.1 Local agent 7.2 Capacity to service equipment 8.1 Inspection and Testing 9. Pre installation requirements 9.1 Pre installation and testing 10.1 Installation and testing 10.1 Training A. Adjustable - Constant to a depth of at least 500mm - Shadowless - Soulo to a 4800 K - Colour rendering index >95% - Deeming range 30-100% - Electronic system with touch button light intensity - Control mounted at a convenient place preferable on the head lamp - Low Voltage LEDs, service life > 40,000 hours - Light field diameter of 300mm at 1M - Minimum 2.5M above floor - All mounting ceiling anchor plates, bolts, nuts - For at least 2 hours - Automatic changeover and charger unit - Minimum 2.5M above floor - All mounting ceiling anchor plates, bolts, nuts - For at least 2 hours - Automatic changeover and charger unit - Minimum 2.5M above floor - All mounting ceiling anchor plates, bolts, nuts - For at least 2 hours - Automatic changeover and charger unit - Minimum 2.5M above floor - All mounting ceiling anchor plates, bolts, nuts - For at least 2 hours - Automatic changeover and charger unit - Minimum 2.5M above floor - All mounting ceiling achor plates, bolts, nuts - For at l	3.1.3	Maximum Light Intensity	- Above 150,000 lux at 1 meter each
3.1.6 Field 3.1.6 Field 3.1.7 Light Colour Temperature Shadowless 3.1.7 Light Colour Temperature 3.1.8 Lighting Control 3.1.9 Shadowless 3.1.10 Lighting Bulb 3.1.11 Mounting Ceiling Height 3.2 Accessories 3.3 Emergency Backup Power 3.3.1 Main and Auxilliary Units 4. Physical Characteristics 4.1 Main and Auxilliary Units 5. Operating Environment 5.1 Power requirements 5.2 Ambient temperature 5.3 Relative Humidity 5. Quality Standards 6.1 Product Conformity Standards 6.2 Product Conformity Standards 6.3 Product Conformity Standards 6.4 Physical Characteristics 9.1 Pre installation and testing 10.1 Installation and testing 10.1 Installation and testing 10.1 Training - Complete installation and set-up of the light at per manufacturer's instructions in accordance with the Acceptance Testing Procedures.	3.1.4	•	- Adjustable
3.1.6 Field 3.1.7 Light Colour Temperature 3.1.8 Lighting Control 3.1.9 - Electronic system with touch button light intensity - Control mounted at a convenient place preferable on the head lamp - Low Voltage LEDs, service life > 40,000 hours - Light field diameter of 300mm at 1M - Minimum 2.5M above floor - All mounting celing anchor plates, bolts, nuts - For at least 2 hours - With sealed batteries - Automatic changeover and charger unit 4. Physical Characteristics - Automatic changeover and charger unit 4. Physical Characteristics - Automatic changeover and charger unit 5. Operating Environment - Ceiling Mount 5. Operating Environment - 10 to 40°C - 20%-90% - 20%-90% - Should be locally available and reliable Agent shall have adequate facilities, spare parts, qualified and Skilled technical staff. 8. Delivery point - Complete installation and testing - Complete installation and set-up of the light at per manufacturer's instructions in accordance with the Acceptance Testing Procedures.	3.1.5	Field	· ·
- Colour rendering index > 95% - Deeming range 30-100% 3.1.10 3.1.10 3.1.11 Mounting Ceiling Height 3.2 Accessories 3.3 Emergency Backup Power 3.3.3.1 - All mounting ceiling anchor plates, bolts, nuts - Light field diameter of 300mm at 1M - Minimum 2.5M above floor - All mounting ceiling anchor plates, bolts, nuts - For at least 2 hours - With sealed batteries - Automatic changeover and charger unit 4. Physical Characteristics 4.1 Main and Auxilliary Units - Ceiling Mount 5. Operating Environment 5.1 Power requirements - 220 - 240V AC, 50Hz, single phase with PE - 10 to 40°C - 20%-90% 6. Quality Standards 6.1 Manufacturing Standards 6.2 Product Conformity Standards 6.3 Product Conformity Standards 6.4 Local back up service 7.1 Local agent 7.2 Capacity to service equipment 8. Delivery point R.1 Inspection and Testing 9. Pre installation requirements - To be conducted 10. Installation and testing 10.1 Complete installation and - Complete installation and eventual the Acceptance Testing Procedures.	3.1.6	Field	
- Colour rendering index > 95% - Deeming range 30-1009 3.1.19 3.1.10 Lighting Bulb - Low Voltage LEDs, service life > 40,000 hours - Light field diameter of 30mm at 1M - Minimum 2.5M above floor - All mounting ceiling Height 3.2 Accessories 3.3 Emergency Backup Power 3.3.1 Physical Characteristics 4.1 Physical Characteristics 4.1 Main and Auxilliary Units - Ceiling Mount 5. Operating Environment 5.1 Power requirements - Ambient temperature - 10 to 40°C - 20%-90% 6. Quality Standards 6.1 Manufacturing Standards 6.2 Product Conformity Standards 6.3 Product Conformity Standards 6.4 Local agent 7.2 Capacity to service equipment 8. Delivery point R.1 Inspection and Testing 9. Pre installation requirements - To be conducted - Complete installation and testing 10.1 Installation and testing 10.1 Training	3.1.7	Light Colour Temperature	- 3600 to 4800 K
- Deeming range 30-100% 3.1.9 - Electronic system with touch button light intensity - Control mounted at a convenient place preferable on the head lamp - Low Voltage LEDs, service life > 40,000 hours - Light field diameter of 300mm at 1M - Light field diameter of 300mm at 1M - Light field diameter of 300mm at 1M - Minimum 2.5M above floor - All mounting ceiling anchor plates, bolts, nuts - For at least 2 hours - With sealed batteries - Automatic changeover and charger unit 4. Physical Characteristics - Automatic changeover and charger unit 5. Operating Environment - Ceiling Mount - Ceiling Mount - For at least 2 hours - All mounting ceiling anchor plates, bolts, nuts - For at least 2 hours - All mounting ceiling anchor plates, bolts, nuts - For at least 2 hours - All mounting ceiling anchor plates, bolts, nuts - For at least 2 hours - For at least 2 hours - All mounting ceiling anchor plates, b			- Colour rendering index >95%
3.1.9 Electronic system with touch button light intensity 3.1.10 Lighting Bulb Low Voltage LEDs, service life > 40,000 hours 3.1.11 Mounting Ceiling Height Accessories Light field diameter of 300mm at 1M 3.1.11 Mounting Ceiling Height Accessories All mounting ceiling anchor plates, bolts, nuts 3.3.1 For at least 2 hours With sealed batteries Automatic changeover and charger unit 4. Physical Characteristics Physical Characteristics Automatic changeover and charger unit 4. Physical Characteristics Automatic changeover and charger unit 5. Operating Environment Power requirements -220 - 240V AC, 50Hz, single phase with PE 5.2 Ambient temperature -10 to 40°C 5.3 Relative Humidity -20%-90% 6. Quality Standards -1SO certified 6.1 Product Conformity Standards -1SO certified 6.2 Product Conformity Standards -1SO certified 6.3 Product Conformity Standards -1SO certified 6.4 EU, IEC -1DA and CE approved 7. Local back up service -3 Should be locally available and reliable -3 Agent shall have adequate facilities, spare parts, qualified and Skilled technical staff. 8. Delivery point -3 County Hospital 9. Pre installation requirements -10 be conducted 10. Installation and testing -2 Complete installation and set-up of the light at permanufacturer's instructions in accordance with the Acceptance Testing Procedures.			- Deeming range 30-100%
3.1.10 Lighting Bulb 3.1.11 Mounting Ceiling Height 3.2 Accessories 3.3 Emergency Backup Power 3.3.11 Main and Auxiliary Units 5. Operating Environment 5.1 Power requirements 5.2 Ambient temperature 5.3 Relative Humidity 5.4 Product Conformity Standards 6.1 Product Conformity Standards 6.2 Product Conformity Standards 6.3 Delivery point 7.2 Capacity to service equipment 7.4 Local back up service 7.5 Delivery point 8.1 Inspection and Testing 7.0 Pre installation and testing 7.1 Complete installation and testing 7.1 Training 7. Control mounted at a convenient place preferable on the head lamp 1. Low Voltage LEDs, service life > 40,000 hours 1. Low Voltage LEDs, service life > 40,000 hours 1. Low Voltage LEDs, service life > 40,000 hours 1. Low John and 1. Minum 2.5M above floor 2. All mounting ceiling anchor plates, bolts, nuts 1. For at least 2 hours 2. All mounting ceiling anchor plates, bolts, nuts 2. Eight field diameter of 300mm at 1M 2. Minimu 2.5M above floor 3. All mounting ceiling anchor plates, bolts, nuts 3. For at least 2 hours 4. Min munu 2.5M above floor 4. All mounting ceiling anchor plates, bolts, nuts 4. Physical Characteristics 4. Automatic changeover and charger unit 4. Ceiling Mount 4. Ceiling Mount 4. Ceiling Mount 5. Ceiling Mount 5. Ceiling Mount 6. Quality Standards 6.1 Io to 40°C 2. 20%-90% 6. Quality Standards 6.1 Io to 40°C 2. EU, EC. 4. FDA and CE approved. 7. Local back up service 7. Local agent 7. Capacity to service equipment 6. Should be locally available and reliable. 4. Agent shall have adequate facilities, spare parts, qualified and Skilled technical staff. 8. Delivery point 8. De	3.1.8	Lighting Control	
Sample S	3.1.9		
3.1.10 Lighting Bulb 3.1.11 Mounting Ceiling Height 3.2 Accessories 3.3 Emergency Backup Power 3.3.1			
- Light field diameter of 300mm at 1M - Minimum 2.5M above floor - All mounting ceiling anchor plates, bolts, nuts - For at least 2 hours - With sealed batteries - Automatic changeover and charger unit - Ceiling Mount - Ceili	3.1.10) Lighting Bulb	•
3.1.11 Mounting Ceiling Height 3.2 Accessories Emergency Backup Power 3.3.1 4. Physical Characteristics 4.1 Main and Auxilliary Units 5. Operating Environment 5.1 Power requirements 5.2 Ambient temperature 5.3 Relative Humidity 6. Quality Standards 6.1 Manufacturing Standards 6.2 Product Conformity Standards 7.1 Local back up service 7.1 Local agent 7.2 Capacity to service equipment 7.2 Capacity to service equipment 8.1 Inspection and Testing 9. Pre installation requirements 9.1 Pre installation and testing 10.1 Complete installation and Capacity of the light at per manufacturer's instructions in accordance with the Acceptance Testing Procedures. 11. Training			
3.2 Accessories 3.3 Emergency Backup Power 3.3.1	3.1.1	1 Mounting Ceiling Height	
3.3.1 Emergency Backup Power 3.3.1			
- With sealed batteries - Automatic changeover and charger unit 4. Physical Characteristics 4.1 Main and Auxilliary Units - Ceiling Mount 5. Operating Environment 5.1 Power requirements 5.2 Ambient temperature 5.3 Relative Humidity - 200 - 240V AC, 50Hz, single phase with PE - 10 to 40°C - 20%-90% 6. Quality Standards 6.1 Manufacturing Standards 6.2 Product Conformity Standards - ISO certified EU, IEC FDA and CE approved. 7. Local back up service 7.1 Local agent - Should be locally available and reliable Agent shall have adequate facilities, spare parts, qualified and Skilled technical staff. 8. Delivery point - County Hospital 9. Pre installation requirements 9.1 Pre installation and testing - Complete installation and set-up of the light at per manufacturer's instructions in accordance with the Acceptance Testing Procedures.			
- Automatic changeover and charger unit 4. Physical Characteristics 4.1 Main and Auxilliary Units - Ceiling Mount 5. Operating Environment 5.1 Power requirements 5.2 Ambient temperature 5.3 Relative Humidity - 20%-90% 6. Quality Standards 6.1 Manufacturing Standards 6.2 Product Conformity Standards 6.2 Product Conformity Standards 6.2 Product Conformity Standards 6.3 Product Conformity Standards 6.4 Local back up service 7.1 Local back up service 7.2 Capacity to service equipment Capacity to service equipment 8. Delivery point 8.1 Inspection and Testing - County Hospital 9. Pre installation requirements 9.1 Pre installation and testing 10.1 Complete installation and - Complete installation and set-up of the light at per manufacturer's instructions in accordance with the Acceptance Testing Procedures.		=morgone, =wemap 1 0 mor	
4. Physical Characteristics 4.1 Main and Auxilliary Units - Ceiling Mount 5. Operating Environment 5.1 Power requirements - 220 - 240V AC, 50Hz, single phase with PE - 10 to 40°C - 20%-90% 6. Quality Standards 6.1 Manufacturing Standards 6.2 Product Conformity Standards - EU, IEC FDA and CE approved. 7. Local back up service 7.1 Local agent - Should be locally available and reliable Agent shall have adequate facilities, spare parts, qualified and Skilled technical staff. 8. Delivery point Inspection and Testing - County Hospital 9. Pre installation requirements 9.1 Pre installation assessment - To be conducted 10. Installation and testing 10.1 Complete installation and - Complete installation and set-up of the light at per manufacturer's instructions in accordance with the Acceptance Testing Procedures.			
4.1 Main and Auxilliary Units - Ceiling Mount 5. Operating Environment 5.1 Power requirements 5.2 Ambient temperature 5.3 Relative Humidity - 20%-90% 6. Quality Standards 6.1 Manufacturing Standards 6.2 Product Conformity Standards 6.2 Product Conformity Standards 6.3 Agent Shall have adequate facilities, spare parts, qualified and Skilled technical staff. 7. Local back up service 7.1 Local agent 7.2 Capacity to service equipment 8. Delivery point 8.1 Inspection and Testing 9. Pre installation requirements 9.1 Pre installation and testing 10.1 Complete installation and - Complete installation and set-up of the light at per manufacturer's instructions in accordance with the Acceptance Testing Procedures. 11. Training			
4.1 Main and Auxilliary Units - Ceiling Mount 5. Operating Environment 5.1 Power requirements 5.2 Ambient temperature 5.3 Relative Humidity - 20% - 90% 6. Quality Standards 6.1 Manufacturing Standards 6.2 Product Conformity Standards 6.2 Product Conformity Standards 6.3 EU, IEC FDA and CE approved. 7. Local back up service 7.1 Local agent 7.2 Capacity to service equipment Capacity to service equipment 8. Delivery point 8.1 Inspection and Testing - County Hospital 9. Pre installation requirements 9.1 Pre installation and testing 10.1 Complete installation and - Complete installation and set-up of the light at per manufacturer's instructions in accordance with the Acceptance Testing Procedures. 11. Training	4.	Physical Characteristics	
5. Operating Environment 5.1 Power requirements 5.2 Ambient temperature 5.3 Relative Humidity 5.3 Relative Humidity 6. Quality Standards 6.1 Manufacturing Standards 6.2 Product Conformity Standards 6.2 Product Conformity Standards 6.3 Agent Agent 6.4 Local back up service 7.5 Local agent 7.6 Capacity to service equipment 7.7 Capacity to service equipment 8.1 Inspection and Testing 8. Delivery point 8.1 Inspection and Testing 9. Pre installation requirements 9.1 Pre installation assessment 9.1 To be conducted 10. Installation and testing 10.1 Complete installation and 10. Training 11. Training	4.1	•	- Ceiling Mount
5.1 Power requirements 5.2 Ambient temperature 5.3 Relative Humidity 6. Quality Standards 6.1 Manufacturing Standards 6.2 Product Conformity Standards 6.2 Product Conformity Standards 6.3 Local back up service 7.1 Local agent 7.2 Capacity to service equipment 8.1 Inspection and Testing 9. Pre installation requirements 9.1 Pre installation and testing 10.1 Installation and testing 10.1 Training - 220 - 240V AC, 50Hz, single phase with PE - 10 to 40°C - 20%-90% - Ito 40°C - 20%-90% - IsO certified EU, IEC FDA and CE approved. - Should be locally available and reliable Agent shall have adequate facilities, spare parts, qualified and Skilled technical staff. - County Hospital - County Hospital - To be conducted - Complete installation and set-up of the light at per manufacturer's instructions in accordance with the Acceptance Testing Procedures.		•	
5.2 Ambient temperature 5.3 Relative Humidity - 20%-90% 6. Quality Standards 6.1 Manufacturing Standards 6.2 Product Conformity Standards 6.3 EU, IEC FDA and CE approved. 7. Local back up service 7.1 Local agent 7.2 Capacity to service equipment Capacity to service equipment Capacity to service equipment 8. Delivery point Inspection and Testing - County Hospital 9. Pre installation requirements 9.1 Pre installation assessment - To be conducted 10. Installation and testing Complete installation and - Complete installation and set-up of the light at per manufacturer's instructions in accordance with the Acceptance Testing Procedures. 11. Training	5.	Operating Environment	
 6. Quality Standards 6.1 Manufacturing Standards 6.2 Product Conformity Standards 6.3 Product Conformity Standards 6.4 EU, IEC. - FDA and CE approved. 7. Local back up service 7.1 Local agent - Should be locally available and reliable. - Agent shall have adequate facilities, spare parts, qualified and Skilled technical staff. 8. Delivery point Inspection and Testing - County Hospital - To be conducted - To be conducted - Complete installation and set-up of the light at per manufacturer's instructions in accordance with the Acceptance Testing Procedures. 11. Training 	5.1		- 220 - 240V AC, 50Hz, single phase with PE
 6. Quality Standards 6.1 Manufacturing Standards 6.2 Product Conformity Standards 6.3 Product Conformity Standards 6.4 EU, IEC. - FDA and CE approved. 7. Local back up service 7.1 Local agent - Should be locally available and reliable. - Agent shall have adequate facilities, spare parts, qualified and Skilled technical staff. 8. Delivery point Inspection and Testing - County Hospital - To be conducted - To be conducted - Complete installation and set-up of the light at per manufacturer's instructions in accordance with the Acceptance Testing Procedures. 11. Training 	5.2		- 10 to 40°C
6.1 Manufacturing Standards 6.2 Product Conformity Standards 6.2 Product Conformity Standards 6.2 Product Conformity Standards 6.3 Product Conformity Standards 6.4 Product Conformity Standards 6.5 Product Conformity Standards 6.6 Product Conformity Standards 6.7 Product Conformity Standards 6.8 Product Conformity Standards 6.9 Pre installation and Testing 6.0 Pre installation requirements 6.1 Inspection and Testing 6.2 Pre installation requirements 6.3 Pre installation and testing 6.4 Product Conformity Standards 6.5 Pre installation and Testing 6.6 Pre installation requirements 6.7 Pre installation and testing 6.8 Pre installation and testing 6.9 Pre installation and testing 6.0 Installation and testing 6.1 Product Conformity Standards 6.2 Product Conformity Standards 6.3 Product Conformity Standards 6.4 Product Conformity Standards 6.5 Product Conformity Standards 6.6 Product Conformity Standards 6.6 Product Conformity Standards 6.7 Product Conformity Standards 6.8 Product Conformity Standards 6.9 Product Conformity Standards 6.9 Product Conformity Standards 6.1 Product Conformity Standards 6.2 Product Conformity Standards 6.2 Product Conformity Standards 6.3 Product Conformity Standards 6.4 Product Conformity Standards 6.5 Product Conformity Standards 6.6 Product Conformity Standards 6.7 Product Conformity Standards 6.7 Product Conformity Standards 6.8 Product Conformity Standards 6.8 Product Conformity Standards 6.9 Product Conformity Standards 6.0 Product	5.3	Relative Humidity	- 20%-90%
6.1 Manufacturing Standards 6.2 Product Conformity Standards 6.2 Product Conformity Standards 6.2 Product Conformity Standards 6.3 Product Conformity Standards 7. Local back up service 7.1 Local agent 7.2 Capacity to service equipment 7.3 Capacity to service equipment 8.4 Delivery point 8.5 Delivery point 8.6 Delivery point 8.7 Pre installation and Testing 8. Pre installation requirements 9.1 Pre installation assessment 9.1 Installation and testing 10.1 Complete installation and set-up of the light at per manufacturer's instructions in accordance with the Acceptance Testing Procedures. 11. Training		O124 C411-	
 Froduct Conformity Standards EU, IEC. FDA and CE approved. Local back up service Local agent Capacity to service equipment Should be locally available and reliable. Agent shall have adequate facilities, spare parts, qualified and Skilled technical staff. Delivery point Inspection and Testing County Hospital Pre installation requirements To be conducted Installation and testing Complete installation and set-up of the light at per manufacturer's instructions in accordance with the Acceptance Testing Procedures. Training 			ICO contified
- FDA and CE approved. 7. Local back up service 7.1 Local agent 7.2 Capacity to service equipment 8. Delivery point 8.1 Inspection and Testing 9. Pre installation requirements 9.1 Pre installation assessment 9.1 Installation and testing 10.1 Complete installation and 10. Installation and testing 10.1 Training 10.1 Training			
7. Local back up service 7.1 Local agent 7.2 Capacity to service equipment 8. Delivery point 8.1 Inspection and Testing 9. Pre installation requirements 9.1 Pre installation assessment 10.1 Installation and testing 10.1 Complete installation and 10. Installation and testing 10.1 Training 10.1 Training	0.2	Product Conformity Standards	
7.1 Local agent 7.2 Capacity to service equipment - Should be locally available and reliable Agent shall have adequate facilities, spare parts, qualified and Skilled technical staff. 8. Delivery point 8.1 Inspection and Testing - County Hospital 9. Pre installation requirements 9.1 Pre installation assessment - To be conducted 10. Installation and testing 10.1 Complete installation and - Complete installation and set-up of the light at per manufacturer's instructions in accordance with the Acceptance Testing Procedures.			- FDA and CE approved.
7.1 Local agent 7.2 Capacity to service equipment - Should be locally available and reliable Agent shall have adequate facilities, spare parts, qualified and Skilled technical staff. 8. Delivery point 8.1 Inspection and Testing - County Hospital 9. Pre installation requirements 9.1 Pre installation assessment - To be conducted 10. Installation and testing 10.1 Complete installation and - Complete installation and set-up of the light at per manufacturer's instructions in accordance with the Acceptance Testing Procedures.	7	Local back un service	
7.2 Capacity to service equipment - Agent shall have adequate facilities, spare parts, qualified and Skilled technical staff. 8. Delivery point 8.1 Inspection and Testing - County Hospital 9. Pre installation requirements 9.1 Pre installation assessment - To be conducted 10. Installation and testing 10.1 Complete installation and set-up of the light at per manufacturer's instructions in accordance with the Acceptance Testing Procedures. 11. Training			- Should be locally available and reliable
and Skilled technical staff. 8. Delivery point 8.1 Inspection and Testing - County Hospital 9. Pre installation requirements 9.1 Pre installation assessment - To be conducted 10. Installation and testing 10.1 Complete installation and set-up of the light at per manufacturer's instructions in accordance with the Acceptance Testing Procedures. 11. Training		•	
8. Delivery point 8.1 Inspection and Testing - County Hospital 9. Pre installation requirements 9.1 Pre installation assessment - To be conducted 10. Installation and testing 10.1 Complete installation and - Complete installation and set-up of the light at per manufacturer's instructions in accordance with the Acceptance Testing Procedures.	'.2	capacity to service equipment	
 8.1 Inspection and Testing - County Hospital 9. Pre installation requirements 9.1 Pre installation assessment - To be conducted 10. Installation and testing 10.1 Complete installation and manufacturer's instructions in accordance with the Acceptance Testing Procedures. 11. Training 			and amino common smil.
 8.1 Inspection and Testing - County Hospital 9. Pre installation requirements 9.1 Pre installation assessment - To be conducted 10. Installation and testing 10.1 Complete installation and manufacturer's instructions in accordance with the Acceptance Testing Procedures. 11. Training 	8.	Delivery point	
 9. Pre installation requirements 9.1 Pre installation assessment - To be conducted 10. Installation and testing 10.1 Complete installation and manufacturer's instructions in accordance with the Acceptance Testing Procedures. 11. Training 		v <u>-</u>	- County Hospital
9.1 Pre installation assessment - To be conducted 10. Installation and testing 10.1 Complete installation and - Complete installation and set-up of the light at per manufacturer's instructions in accordance with the Acceptance Testing Procedures. 11. Training			* *
9.1 Pre installation assessment - To be conducted 10. Installation and testing 10.1 Complete installation and - Complete installation and set-up of the light at per manufacturer's instructions in accordance with the Acceptance Testing Procedures. 11. Training	9.	Pre installation requirements	
10.1 Complete installation and - Complete installation and set-up of the light at per manufacturer's instructions in accordance with the Acceptance Testing Procedures. 11. Training	9.1		- To be conducted
10.1 Complete installation and - Complete installation and set-up of the light at per manufacturer's instructions in accordance with the Acceptance Testing Procedures. 11. Training			
manufacturer's instructions in accordance with the Acceptance Testing Procedures. 11. Training	10.	Installation and testing	
Acceptance Testing Procedures. 11. Training	10.1	Complete installation and	
11. Training			manufacturer's instructions in accordance with the
g .			Acceptance Testing Procedures.
9			
11.1 User training - Onsite on Operation, User Maintenance and daily upkeep.		<u> </u>	
	11.1	User training	- Onsite on Operation, User Maintenance and daily upkeep.

11.2	Maintenance training	- Onsite technical training on preventive maintenance and fault finding.
12.	Technical documentations	
12.1	User Manuals	- English language, 1 set
12.2	Service Manuals	- English language, 1 set
13.	Commissioning	
13.1	Testing and commissioning	- To the satisfaction of the user in accordance with the Acceptance Testing Procedures
14.	Warranty	
14.1	Equipment	- Minimum of One (1) year after commissioning on all parts
		(provide Warrantee card)
14.2	Equipment system	- Minimum of one (1) year

3. ANAESTHETIC MACHINE WITH 2 VENTILATORS

ITEM DESCRIPTION	ANAESTHETIC MACHINE WITH
	VENTILATOR
1.General Description	
Inhalation anaesthetic machine with electron	ic ventilator complete with all accessories for low and high
flow anaesthesia, adult, paediatric and infant	application. It should include a patient monitor unit.
2. Composition	
2.1 Main Unit	- 1 Unit.
2.2 Electronic Ventilator	- 1 Unit
2.3 Patient Monitor	- 1 Unit
2.4 Accessories complete start-up kit	- 1 Set
3. Performance specifications	
3.1 Main Unit	
3.1.1 Anaesthetic trolley with minimum 2 d	lrawers and a table top, with yokes for Oxygen (O ₂)
and Nitrous Oxide (N2O) portable cyl	inder and support for circle systems including hoses and
absorbers and support for central pipe	line gas system. Model on current production
3.1.2 Gas delivery system	- 3 gas delivery system (O ₂ , N ₂ O and air) with both inlets for central gas pipeline system, and separate O ₂ and N ₂ O portable
	cylinders.
3.1.3 Yokes	- To support portable Oxygen (O ₂) and Nitrous Oxide (N ₂ O)
	cylinders
3.1.5 Pressure regulators and gauges	- Intergraded in the trolley
for O ₂ and N ₂ O	
3.1.6 Central gas pipeline system	- Standard BS connections and colour codes for O ₂ , N ₂ O,
	and Air
	- Integrated, user-controlled, heated system
3.1.8 All patient connecting hoses	- Corrugated, Transparent, autoclavable (134°C), φ 22 mm,
	with ISO connectors
3.1.9 CO ₂ absorber	- Integrated, complete with Soda lime and switch for
	Magill's circuit.
3.1.10 Software	- With anesthetic agent calculation
3.1.11 Plug and Play Multi-Gas Modules	- Comprehensive breath-by-breath analysis of gasses
3.1.12 Internal Battery,	- 2 to 4 hours of supply (Lithium ion battery)

0.1.1	2.4	
3.1.1	3 Accessories:	- Complete sets Adult, paeds (Breathing circuits for
		ventilator, Face masks (transparent type), Breathing bags
		- Magill's circuit complete with adult mask
		- Aynes Paed circuit
		- CO ₂ absorber gas out let
3.2	Vaporizer	- Minimum Halothane and Isoflurane
3.2.1	Compensation	- Circuit gas leakage and automatic compliance compensated
	Keyed filler according to ISO stand	
3.3	Safety controls	
3.3.1	O ₂ supply failure	- Audible alarm with reset
	Hypoxyguard	- Available
	O ₂ Flush	- Available
1	Ventilator	
	Type	- Microprocessor controlled and electrical/gas driven
	Application	- Suitable for adult, paediatric and infant application
	11	without changing parts between patient types
3.4.3		- Ventilation with ambient air possible
	Modes	- IPPV, PCV, PCV-VG, SIMV,
		SIMV-VG, SIMV-PC, SIMV-VC, VCV, CPAP/PS, PSV
(+Ap	nea	
		back-up)
2.5	Dieploy	Colour dienlay minimum 10.1" touch comes
1	Display Display parameters	- Colour display minimum 12.1" touch-screen Minute Volume, Tidel Volume, Peter Pressure Peek
3.3.1	Display parameters	- Minute Volume, Tidal Volume, Rate, Pressure Peak
fucal		Response, PEEP, FiO2, Graphic Trends + numerical data and
fresh	I.	gas flow O noncontago
26	Patient monitor	gas flow O ₂ percentage - To be mounted on the anesthetic machine
	Parameters Parameters	
3.0.1	r arameters	- Pulse rate, SpO ₂ , Temperature, Blood pressure (NIPB),
260	Diaploy	ECG 5 leads, ETCO ₂ , IBP
	Display	- Colour Display minimum 12" 5 Parameter display
	Accessories: as startup kits.	- SpO ₂ , Reusable (Adult, Paeds, Infant Sensors) - Temperature
probe	28	DD outfor rougable (for I area adult adult amall adult Dand
		- BP cuffs, reusable (for Large adult, adult, small adult, Paed
		and Thigh) ECG. 5 Loads. ETCO: IRP accessories
		- ECG 5 Leads, ETCO ₂ , IBP accessories
27	Pook up Down Cum-1-	- 3X5 Ltr Sodalime.
3.7	Back-up Power Supply	- UPS rated 1,000VA
1	Dhygical Characteristics	
4. 4.1	Physical Characteristics Main Unit	- Mobile, with rolling stand (trolley mounting), castor-
7.1	maili Ollit	wheel 125mm with one press -central brake system, stable
base.		wheel 125mm with one press -central brake system, stable
base.	Outer dimensions	- Compact design
5.	Operating Environment	,
5.1	Power requirements	- 240V AC, 50Hz, 13A, 3 pin plug, 3m long cord with PE
5.2	Auxiliary electrical outlets	- 4 in number
5.3	Ambient temperature	- 10° C to 40° C
5.4	Relative Humidity	- 20% to 90% No dew Condensation
6.	Quality Standards	to your tit the works with
6.1	Manufacturing Standards	- ISO certified.
6.2	Conformity to Standards	- CE, FDA approved and marked.
7.	Local back up service	, TT
	Local back up bet vice	

7.1 Local agent	- Should be locally available and reliable.
7.2 Capacity to service equipment	- Agent shall have adequate facilities, spare parts, qualified
	and Skilled technical staff.
8. Delivery point	
8.1 Inspection and Testing	- At County Hospital
9. Installation and testing	
9.1 Complete installation and	- At the site advised, as per manufacturer's instructions.
set-up of the equipment	
10. Training	
10.1 User training	- On Operation, User Maintenance and daily upkeep.
10.2 Maintenance training	- Technical training on preventive maintenance and fault-
finding.	
11. Technical documentations	
11.1 User Manuals	- Printed and electronic Copies, English language, 1 set
11.2 Service Manuals	- English language, 1 set
12. Commissioning	
12.1 Testing and commissioning	- To the satisfaction of the client
13. Warranty	
13.1 Equipment	- Minimum of Two (2) years after commissioning on all parts
	(provide Warrantee card)

4. CRASH CART AND DEFIBRILLATOR

A) ITEM DESCRIPTION	DEFIBRILLATOR MACHINE	
1.General Description		
A biphasic Defibrillator, portable, durable, reliable, with manual defibrillation, AED, pacer and monitorin		
modes for emergency use in a hos		
2. Composition	<u> </u>	
2.1 Main Unit	- Biphasic defibrillator	
2.2 Antisurge protector	- As per Manufacturer's specification, 100V-250Vac	
3. Performance specifications		
3.1 Main Unit	- with visual and voice prompts	
3.1.1 Waveform	- Biphasic	
3.1.2 Energy display	- Indicates both selected and delivered energy	
3.1.3 Display	- High resolution coloured LCD ≥7"; Information – Heart rate,	
	alarm selection and limits, Delivered energy, lead/pads, Alarm	
	On/Off, SPO2, AES functions and prompts	
3.1.4 Battery Pack Up	- Rechargeable, Above 50 Shocks at maximum energy, battery status indication	
3.1.5 Paddles	- Standard Adult and paediatric	
3.1.6 Inbuilt recorder	- Thermal, high resolution	
3.1.7 Data storage	-Events, waveform, voice recording and data export	
3.2. Monitoring	- Vital signs	
3.2.1 ECG	$- \ge 5$ lead cable, paddles; lead selection, user-selectable heart	
rate		
	alarm, smart alarms, adult and paed	
3.2.2 SPO2	- measurement, user-selectable speed and alarm, adult and paed	
3.2.3 EtCO2	- Continuous waveform	
3.3 Pacer		
3.3.1 Multi-Function Electrode		
3.3.2 Output protection	- Fully defibrillator protected and isolated.	
3.4 Accessories	- Minimum 5 paper	
3.4 Antisurge protector	- As per manufacturers' specification, 100V-250Vac	

4.	Physical Characteristics	
4.1	Main Unit	- Portable
5.	Operating Environment	
5.1	Power requirements	- 240V AC, 13A, 3 pin 50Hz, single phase
5.2	Ambient temperature	- 10 to 40 degrees centigrade
5.3	Relative Humidity	- 20%-90%, No dew Condensation
6.	Quality Standards	
6.1	Manufacturing Standards	- ISO certified.
6.2	Conformity to Standards	- CE, FDA approved and marked.
7.	Local back up service	
7.1	Local agent	- Should be locally available and reliable.
7.2	Capacity to service equipment	- Agent shall have adequate facilities, spare parts, qualified
		and Skilled technical staff.
8.	Delivery point	
8.1	Inspection and Testing	- At County Hospital
9.	Pre installation requirements	s - Nil
10.	Installation and testing	
10.1	Complete installation and	- At the site advised, as per manufacturer's instructions.
	set-up of the equipment	
11.	Training	
11.1	User training	- On Operation, User Maintenance and daily upkeep.
11.2	Maintenance training	- Technical training on preventive maintenance and fault-
		finding.
12.	Technical documentations	T 11.1
12.1	User Manuals	- English language, 1 set
12.2		- English language, 1 set
13.	Commissioning	
13.1	Testing and commissioning	- To the satisfaction of the client
14.	Warranty	(NB: Provide Warranty)
14.1	Equipment	- Minimum of One (1) year after commissioning on all parts
B)	Crash Cart	

- B) Crash Cart
 - Lightweight polymer construction with smooth rounded corners to assist cleaning
 - Single seal mechanism to secure the top compartment, drawers and tilt out side bins (security seal compatible)
 - Top storage compartment with clear removable cover provides instant access to first line equipment
 - Strapless defibrillator platform on swingarm adjusts to accommodate various modes
 - extension lightweight polymer drawers, depths 76, 152, 229 & 305mm, (max load per drawer 6/11/20kg)
 - Supplied with a front/rear mounting backboard (model dependent)
 - Ergonomic handle with fifth wheel steering provides maximum control when a tending an emergency situation
 - Incorporates Microban offering antimicrobial protection
 - 1 5mm Antistatic castors, 2 off braking

5. ELECTROSURGICAL DIATHERMY MACHINE

ITEM DESCRIPTION

ELECTROSURGICAL UNIT

1. General Description

High frequency electro surgical (diathermy) machine suitable for general surgery. The unit should be microprocessor based and capable of performing cutting, coagulation and blend functions at varying power output and complete with foot switch, electrodes and a cart (trolley).

2. Composition

2.1 Main Unit

3. I	Performance specifications	
3.1	Main Unit	
3.1.1	Output Power Nominal	 high frequency output of about 400W adjustable up and down with touch button keys or convenient controls. With automatic output regulation against excess impedance (TUR)
3.1.2	Cutting	 Monopolar, bipolar and blend functions Activation by finger-switch and/or foot switch.
3.1.3	Coagulation	- Monopolar, bipolar, low forced and spray, Activation by finger switch and / or foot switch
3.1.4	Bipolar	- Very low voltage
3.1.5		- Modulated pulse or Hemostatic or equivalent
3.1.6	Display	- Digital Read out
3.1.7		 Active patient electrode with standard electrode handle, with finger switch and connecting cable, reusable and autoclavable at 134 °C.
3.1.8	Patient Plate	- Patient (in different) plate, reusable, rubber, with connecting cable, autoclavable at 134 °C.
3.1.9	Foot Switch	- Two pedal foot switch for cut and coagulation, water proof, cable length not less than 4 metres.
3.1.1	0 Safety Alarms	 Dosage rate control, Leakage current, Patient plate spilt and Short circuit (Audible and visual alarms).
3.2	Accessories	- Full set as startup kit
4.	Physical Characteristics	
4.1	Main Unit	- Mounted on mobile cart, 2 wheel with braking system
5.	Operating Environment	
5.1	Power requirements	- 240V A/C 50Hz, 13A, 3 pin plug, single phase, 3m long
		cable with PE
5.2	Ambient temperature	- 10 to 40°C
5.3	Relative Humidity	- 20% to 90%, No dew Condensation
6.	Quality Standards	ICO conticad
6.1	Manufacturing Standards Conformity to Standards	- ISO certified.- CE, FDA approved and marked.
7.	Local back up service	- CE, FDA approved and marked.
7.1	Local agent	- Should be locally available and reliable.
7.1	Capacity to service equipment	- Agent shall have adequate facilities, spare parts, qualified
		and Skilled technical staff.
8.	Delivery point	
8.1	Inspection and Testing	- At County Hospital
9.	Pre installation requirements	- Nil
10.	Installation and testing	
10.1	Complete installation and	- At the site advised, as per manufacturer's instructions.

	set-up of the equipment	
11.	Training	
11.1	User training	- On Operation, User Maintenance and daily upkeep.
11.2	Maintenance training	- Technical training on preventive maintenance and fault-
		finding.
12.	Technical documentations	
12.1	User Manuals	- English language, Printed and Soft, 1 set
12.2	Service Manuals	- English language, Printed and Soft, 1 set
13.	Commissioning	
13.1	Testing and commissioning	- To the satisfaction of the client
14.	Warranty	(NB: Provide Warranty card)
14.1	Equipment	- Minimum of Two (2) year after commissioning on all parts

<u>6. INFANT INCUBATOR</u>

ITEM	DESCRIPTION	INFANT INCUBATOR WITH PHOTOTHERAPY
	1 10,	bile, reliable, with carefully monitored environment for
2.1 Ma	iposition in Unit isurge protector (UPS)	Infant Incubator with phototherapy.As per Manufacturer's specification, 100V-250Vac
3. Pe 3.1	rformance specifications Main Unit	- Infant baby incubator, double wall canopy, IV pole, mattress tilt / rotation and digital display with
standar	d	panels; Capable for Up-Gas scoring. Fitted with oxygen port, tubings and cylinder; and suction.
3.1.1 temper	Specific performance	with trend monitoring capabilities. Alarm system and timer setting, canopy light system; X-ray slot - Measurement and control of skin and air
3.1.2 3.2	Accessories Antisurge protector (UPS)	and servo humidifier.Phototherapy light meter and Eye shield.As per manufacturers' specification, 100V-250Vac
4. 4.1 swivel	Physical Characteristics Main Unit	- Mobile, with rolling stand (trolley mounting), 4" polyamide / antistatic castor-wheel with brake, with storage cabinets / drawers.
5. 5.1	Operating Environment Power requirements	- 220 - 240V AC, 13A, 3 pin 50Hz, single phase

5.2	Ambient temperature	- 15 to 37 degrees centigrade
5.3	Relative Humidity	- 5%-95%, No dew Condensation
	•	
6.	Quality Standards	
6.1	Manufacturing Standards	- ISO certified.
6.2	Conformity to Standards	- CE marked.
6.3	Authority for Standards	- FDA approved.
0.5	rumonty for Standards	1 D11 approved.
7.	Local back up service	
7.1	Local agent	- Should be locally available and reliable.
7.2	Capacity to service equipment	- Agent shall have adequate facilities, spare parts,
	1 1	- Agent shan have adequate facilities, spare parts,
quali	neu	and Skilled technical staff.
		and Skined technical staff.
8.	Delivery point	
8.1	Inspection and Testing	- At County Hospital
0.1	mspection and resultg	11 County Hospital
9.	Pre installation requirements	
9.1	Nil	
7.1		
10.	Installation and testing	
	Complete installation and	- At the site advised, as per manufacturer's
	actions.	The tire site advised, as per mandraters s
IIIStr	set-up of the equipment	
	set-up of the equipment	
11.	Training	
11.1	User training	- On Operation, User Maintenance and daily upkeep.
	Maintenance training	- Technical training on preventive maintenance and
fault-	<u> </u>	- Technical training on preventive maintenance and
rauit-	•	finding.
		mung.
12.	Technical documentations	
12.1	User Manuals	- English language, 1 set
12.1	Service Manuals	- English language, 1 set - English language, 1 set
12.2	Set vice ivialitials	- English language, 1 set
13.	Commissioning	
13.1	Testing and commissioning	- To the satisfaction of the client
13.1	1 come and commissioning	2 5 MA SMASIMONION OF MIC CHOIC
14.	Warranty	
14.1	Equipment	- Minimum of One (1) year after commissioning on all
parts	1 · F	(-,)
		(provide Warrantee card)
14.2	Equipment system	- Minimum of one (1) year
1 7.4	Equipment bystem	Transministration one (1) your

7. INFANT RESCUSCITAIRE

ITEM DESCRIPTION	INFANT WARMER WITH RESCUSCITAIRE.

1.	General Description		
	Standard Infant Warmer, Mobile, reliable, with carefully monitored environment for neonates.		
2.	Composition		
2.2	Main Unit		
3	Performance Specifications		
3.1	Main Unit	- with instrument tray, IV pole, Monitor shelf, X-ray cassette tray, Mattress	
3.1.1	Illumination	- Examination light	
3.1.2	Alarm system	- Controller heating alarms and resuscitaire alarms, Probe Short or open circuit / No probe, System fail, Power fail AC.	
		- Interruption Alarm silence/reset intervals.	
	Alerts	- Manual mode, regular Apgar alerts and visual timer	
	Set temperature / heater output	– with display	
	Gas delivery system Portable Oxygen (O ₂)	 O₂ gas delivery system with both inlet with portable cylinders. Bull nose type connection 	
		- Fitted with standard oxygen port	
3.1.7	Pressure regulator and gauge for O ₂	- Intergraded in the trolley	
3.1.8		- Adjustable with container	
3.2 250V	Antisurge protector (UPS)	- As per manufacturers' specification, 100V-	
,	Back-up Power Supply	- Internal Battery , 2-4 hours of supply	
4.	Physical Characteristics	mount 2 mount of supply	
4.1	Main Unit	- Mobile, with rolling stand (trolley mounting), 4" swivel	
		polyamide / antistatic castor-wheel with brake, stable base, with	
		storage cabinets / drawers.	
5.	Operating Environment	storage eachies / drawers.	
5.1	Power requirements	- 240V AC, 50Hz, 13A, 3 pin plug, 3m long cord with PE	
5.2	Ambient temperature	- 10° C to 40° C	
5.3	Relative Humidity	- 20% to 90% No dew Condensation	
6.	Quality Standards	2070 to 7070 140 dew Condensation	
6. 1	Manufacturing Standards	- ISO certified.	
6.2	Conformity to Standards	- ISO certified CE, FDA approved and marked.	
7.	Local back up service	- CE, I'DA approved and marked.	
7.1	-	- Should be locally available and reliable.	
	Local agent	· · · · · · · · · · · · · · · · · · ·	
7.2	Capacity to service equipment	 Agent shall have adequate facilities, spare parts, qualified and Skilled technical staff. 	
8.	Delivery point		
8.1	Inspection and Testing	- At County Hospital	
	Installation and testing		
9.1	Complete installation and set-up of the equipment	- At the site advised, as per manufacturer's instructions.	
10.	Training		
10.1	User training	- Onsite On Operation, User Maintenance and daily upkeep.	
10.2	Maintenance training	- Onsite technical training on preventive maintenance and fault-finding.	
11.	Technical documentations		
11.1	User Manuals	- Printed and electronic Copies, English language	
11.2	Service Manuals	- Printed and electronic Copies, English language	
12.	Commissioning	1 / 0	
12.1	Testing and commissioning	- To the satisfaction of the client	
	Warranty		
13.1	Equipment	- Minimum of One (1) years after commissioning on all parts	
	1 ·- F	(-, jenie mier tommissioning on an parts	

8. PHOTOTHERAPY UNIT

ITEM	ITEM DESCRIPTION INFANT PHOTOTHERAPY UNIT			
1.Gene	ral Description			
An Infa	An Infant Phototherapy Unit, mobile, reliable, with carefully monitored light therapy			
2. Com	position			
2.1 Mai	in Unit	- Infant Phototherapy Unit.		
2.2 Ant	surge protector (UPS)	As per Manufacturer's specification, 100V-250Vac		
3. Per	rformance specifications			
3.1	Main Unit	- Infant baby phototherapy Unit, Blue LED of 490nm		
	•	ent area, Height adjustable with treatment distance 25		
		2, measured and reported, Lamp source continuous		
	covering $\pm 90^{\circ}$, Built-in non resettable	timer, transparent infant bed with up/down tiltable		
facility				
3.1.2		- Phototherapy light meter and Eye shield.		
3.2	Antisurge protector (UPS)	- As per manufacturers' specification, 100V-250Vac		
4.	Physical Characteristics			
4.1	Main Unit	- Mobile, with rolling stand (trolley mounting), 4"		
swivel				
		polyamide / antistatic castor-wheel with brake, with		
		storage cabinets / drawers.		
5.	Operating Environment			
5.1		- 220 - 240V AC, 13A, 3 pin 50Hz, single phase		
5.2		- 15 to 37 degrees centigrade		
5.3		- 5%-95%, No dew Condensation		
	0 14 94 1 1			
6.	Quality Standards	ICO contico d		
6.1	\mathcal{E}	- ISO certified. - CE marked.		
6.3	3			
6.4		- FDA approved. - Valid certificates		
0.4	Certified	- vand certificates		
7.	Local back up service			
	<u> </u>	Should be locally available and reliable.		
	- ·	Agent shall have adequate facilities, spare parts,		
qualifie				
		and Skilled technical staff.		
8.	Delivery point			
8.1 I	nspection and Testing -	At County Hospital		
9. P	Pre installation requirements			
	-			

9.1	Nil	
10.	Installation and testing	
10.1	Complete installation and	- At the site advised, as per manufacturer's
instru	actions.	
	set-up of the equipment	
11.	Training	
11.1	User training	- On Operation, User Maintenance and daily upkeep.
	Maintenance training	- Technical training on preventive maintenance and
fault-		
		finding.
12.	Technical documentations	
12.1	User Manuals	- English language, 1 set
	Service Manuals	- English language, 1 set
12.2	Service iviandans	English language, 1 sec
13.	Commissioning	
13.1	Testing and commissioning	- To the satisfaction of the client
14.	Warranty	
14.1	Equipment	- Minimum of One (1) year after commissioning on all
parts		
		(provide Warrantee card)
14.2	Equipment system	- Minimum of one (1) year

9. SPACE HEATER

Description: Space / room heater, with oil filled radiator and fan, ≥7 fins, adjustable	
heating power: Min, Med and Maximum power of approximately 2,000 Watts. ON/OFF	
indicator light with room thermostat and cord storage	
Type: Mobile stand, with universal castors	
Power requirements : 240V A/C, 50Hz, 13A, 3 pin single phase, Cable with PE	
Quality Standards	
- Manufacturing Standards - ISO certified.	
- Conformity to Standards - CE approved and marked	
Warranty: A minimum of 1 year after delivery (provide Warranty card)	
* * *	

10. OXYGEN SPLITTER, 5 WAY

1	Description: Oxygen Flow splitter for distribution of medical oxygen from a source for	
	oxygen delivery.	
	- 5 way / outlets and 1 inlet with independent flow meters of range up to 2 lpm	
	- Complete with humidifier bottles	
2	Type: Portable stand, table top	
3	Technical documentations: Provided in English language	
4	Quality Standards	
	- Manufacturing Standards - ISO certified.	
	- Conformity to Standards - CE approved and marked	

	- Authority for Standards - FDA approved
5	Warranty: A minimum of 1 year after delivery (provide Warranty card)

11. PATIENT MONITOR

ITEM DESCRIPTION	PATIENT (CARDIAC) MONITOR			
1.General Description				
	A Complete cardiac monitor, multi-parameter, mobile, high image quality, reliable and environment			
friendly for point of care monitoring.				
2. Composition				
2.1 Main Unit	- Cardiac monitor.			
2.2 Antisurge protector	- As per Manufacturer's specification, 100V-250Vac			
	1			
3. Performance specifications				
3.1 Main Unit	- Cardiac Monitor, at least 15" screen, high resolution			
	Coloured display.			
	- With Internal battery (Maintenance free) with at least 2			
	hours minimum operating time for the monitoring.			
3.1.1 Accessories				
3.1.1.1 NIBP	- Reusable large adult, Adult, paediatric and neonate cuffs.			
3.1.2.2 SPO2 and Pulse Rate	- Reusable finger probes. Monitoring adults, paediatric and			
	neonates probes.			
3.1.2.3 Temperature	- Reusable skin surface probe.			
3.1.2.4 ECG- Respiratory and Heart rate	es - 5-lead, monitoring adults and neonates.			
3.2 Antisurge protector	- As per manufacturers' specification, 100V-250Vac			
4. Physical Characteristics				
4.1 Main Unit	- Mobile, with rolling stand (trolley mounting), 4" castor-			
	wheel with brake, stable base and cable trays.			
5. Operating Environment				
5.1 Power requirements	- 240V AC, 13A, 3 pin 50Hz, single phase			
5.2 Ambient temperature	- 10 to 40 degrees centigrade			
5.3 Relative Humidity	- 20%-90%, No dew Condensation			
6. Quality Standards				
6.1 Manufacturing Standards	- ISO certified.			
6.2 Conformity to Standards	- CE, FDA approved and marked.			
7. Local back up service	• •			
7.1 Local agent	- Should be locally available and reliable.			
7.2 Capacity to service equipment	- Agent shall have adequate facilities, spare parts, qualified			
	and Skilled technical staff.			
8. Delivery point				
8.1 Inspection and Testing	- At County Hospital			
9. Pre installation requirements	- Nil			
10. Installation and testing				
10.1 Complete installation and	- At the site advised, as per manufacturer's instructions.			
set-up of the equipment				
11. Training				
11.1 User training	- On Operation, User Maintenance and daily upkeep.			
11.2 Maintenance training	- Technical training on preventive maintenance and fault-			
	finding.			
12. Technical documentations				
12.1 User Manuals	- English language, 1 set			
12.2 Service Manuals	- English language, 1 set			
13. Commissioning				
13.1 Testing and commissioning	- To the satisfaction of the client			
	B: Provide Warranty)			
(2)	V./			

ITEM	DESCRIPTION	FETAL MONITOR		
1.Gene	1.General Description			
A fetal	A fetal monitor to meet the need of whole stages of labour for monitoring fetus and mother.			
2. Con	nposition			
2.1 Ma	in Unit	- Fetal Monitor, with low ultrasound power		
2.2 An	tisurge protector (UPS)	- As per Manufacturer's specification, 100V-250Vac		
3. Pe	rformance specifications			
3.1	Main Unit	- Fetal monitor, folding up to 90°, Interface of		
parame	eters of			
		FHR, TOCO and fetal movement.		
3.2	Specific performance			
3.2.1	Display	- Atleast 12" TFT color screen		
3.2.1.1	Data storage	- Atleast 40 hours with playback and printed		
3.2.1.2	Alarm system	- For various parameters		
3.2.2	Transducer	- 3 in 1; Fetal Heart Rate (FHR), Uterine activity		
(TOCC	0),			
		Fetal movement (FM); Pulsed Doppler, high		
sensitiv	vity;			
		Frequency approximately 1MHz, wired		
3.2.3	Built-in printer	- Thermal		
3.2.4	Internal Battery	- Li-Ion, minimum of 4 hours continuous monitoring		
3.3	Accessories (10 printer papers)			
3.4	Antisurge protector (UPS)	- As per manufacturers' specification, 100V-250Vac		
3.5	Trolley			
4.	Physical Characteristics			
4.1	Main Unit	- Portable, Desktop, on a trolley		
5.	Operating Environment			
5.1	Power requirements	- 220 - 240V AC, 13A, 3 pin 50Hz, single phase		
5.2	Ambient temperature	- 15 to 37 degrees centigrade		
5.3	Relative Humidity	- 5%-95%, No dew Condensation		
6.	Quality Standards			
6.1	Manufacturing Standards	- ISO certified.		
6.2	Conformity to Standards	- CE marked.		
6.3	Authority for Standards	- FDA approved.		
6.4	Certified	- Valid certificates		
7.	Local back up service			
7.1	Local agent	- Should be locally available and reliable.		
7.2	Capacity to service equipment	- Agent shall have adequate facilities, spare parts,		
		qualified and Skilled technical staff.		
	Delivery point			
	Inspection and Testing	- At County Hospital		
	Pre installation requirements			
-	Nil			
	Installation and testing			
	Complete installation and	- At the site advised, as per manufacturer's		
instruc				
S	set-up of the equipment			

11	m · ·	
11.	Training	
11.1	User training	- On Operation, User Maintenance and daily upkeep.
11.2	Maintenance training	- Technical training on preventive maintenance and
fault-	-	
		finding.
12.	Technical documentations	
12.1	User Manuals	- English language, 1 set
12.2	Service Manuals	- English language, 1 set
13.	Commissioning	
13.1	Testing and commissioning	- To the satisfaction of the client
14.	Warranty	
14.1	Equipment	- Minimum of One (1) year after commissioning on all
parts		
		(provide Warrantee card)
14.2	Equipment system	- Minimum of one (1) year

13. FETAL DOPPLER, HANDHELD

1	Description: - Handheld ultrasound device, heart rate fetal monitor, digital display with color	
	LCD, and battery level indicator, Sound, Probe working frequency ≥ 2.0MHz, type B equipment,	
	complete with user manual, Earphone, Carry bag and Gel	
2	Power requirements : 240V A/C, 50Hz, 13A, 3 pin single phase, Cable with PE with	
	rechargeable batteries	

14. OXYGEN THERAPY (REGULATOR) APPARATUS; GAUGE

1 **Description:**

- A complete set of regulator, flowmeter and humidifier bottle: Bullnose fitted with O-ring, Hand tightened regulators, single flowmeter up to a maximum of 15litres per minute with maneuverable glandnut to fit either horizontal or vertical outlets to cylinder; sintered filter fitted in the humidifier bottle.
- Conformity to Standards
- CE marked and approved.
- Warranty 1 year

15. WALL OXYGEN FLOWMETER (DUAL)

1 **Description:**

- A complete set of Standard Oxygen wall Plug-in Flow meters (Dual), Complete with Humidifier Bottles, single flowmeter up to a maximum of 15litres per minute, sintered filter fitted in the humidifier bottle. Twin control and Patient tubes with Mask both Adults and Paeds.
- Conformity to Standards
- CE marked and approved.
- Warranty 1 year

16. OXYGEN CONCENTRATOR

1 **Description:** Dual Oxygen Concentrator, high flow, high pressure complete with humidifier bottles to continuously produce medical oxygen using PSA principle.

	- Maximum flow: 10L/min	
	- Oxygen purity: ≥90% (V/V)	
	- Running Noise: ≤50 dB	
	- Working time: 10,000 hours, continuous	
	- Display: Digital	
	- Alarms: High and low voltage, power off.	
	- Nebulizing Kit	
	- Oxygen nebulization and oxygen analyzing monitor	
2	Type: Mobile, with 4 low friction lockable universal castors	
3	Power requirements : 240V A/C, 50Hz, 13A, 3 pin single phase, Cable with PE	
4	Quality Standards	
	- Manufacturing Standards - ISO certified.	
	- Conformity to Standards - CE approved and marked	
	- Authority for Standards - FDA approved	
5	Warranty: A minimum of 1 year after delivery (provide Warranty card)	

17. ADULT WEIGHING SCALE

1	Description: Heavy duty, precise scales with measuring rod -0-150kgs with 100 grams
	graduation (analogue),
2	Model: SECA or its equivalent

18. INFANT WEIGHING SCALE WITH MEASURING ROD

1	Description: Heavy duty, digital scale, with head and foot positioners, measuring rod
2	Model: SECA or its equivalent

19. SUCTION MACHINE - ELECTRICAL

1	Description: High vacuum, high flow electric Suction machine / pump; 60L/Min, with		
	autoclavable collection double bottles / jars, 2 X 4 litres, tubing: all sizes, silicon, fitted with		
	antibacterial filter, float valve system for the suction bottles and an internal hydrophobic filter.		
	Provide foot switch. Oil less and maintenance free piston pump, Non-stop operation duty cycle,		
	Maximum vacuum of -0.90 bar.		
2	Type: Mobile, with low friction lockable universal castors		
3	Power requirements : 240V A/C, 50Hz, 13A, 3 pin single phase, Cable with PE		
4	Quality Standards		
	- Manufacturing Standards - ISO certified.		
	- Conformity to Standards - CE approved and marked		
5	Warranty: A minimum of 1 year after delivery (provide Warranty card)		

20. SUCTION MACHINE, MANUAL

1 **Description:** Manual Suction Machine for clinical suction with high negative pressure and high flow rate. With stepping rod, pedal and piston and non-return valve.

Maximum pressure; ≥-0.08MPa
 Pedal force: ≤245 N
 Reservoir Volume: ≥ 1 Ltr
 Maximum flowrate: ≥ 20L/min
 Complete tubing: all sizes, silicon, fitted with antibacterial filter, float valve system for the suction bottles.
 Type: portable

21. EXAMINATION LAMP.

1	Description: Medical examination lamp with one 3W LED white light, High		
	illumination of 70,000 lux at 200mm working distance, lifetime up to 50,000 hours,		
	Goose neck adjustable extension, with Adjustable brightness and adjustable Spot		
	diameter		
2	Type: Mobile stand, with 5 low friction lockable universal castors		
2	Power requirements : 240V A/C, 50Hz, 13A, 3 pin single phase, Cable with PE		
3	Power requirements : 240 V A/C, 50Hz, 13A, 3 pin single phase, Cable with PE		
4	Quality Standards		
	Quality Standards - Manufacturing Standards - ISO certified.		
	Quality Standards		
	Quality Standards - Manufacturing Standards - ISO certified.		

22. DIGITAL BLOOD PRESSURE MACHINE

1	Description: Automatic Blood Pressure with hypertension indicator and enhance		
	intelligence. Cuff with circumference 22 to 42cms. Model: Omron HEM – 7322 or its		
	equivalent.		
2	Type: Portable, desktop.		
3	Power requirements : 240V A/C, 50Hz, 13A, 3-pin single phase, Cable with PE to		
	6VDV Adaptor and rechargeable battery.		
4	Quality Standards		
	- Manufacturing Standards - ISO certified.		
	- Conformity to Standards - CE approved and marked		
5	Warranty: A minimum of 1 year after delivery (provide Warranty card)		

23. NEBULIZER

1	Description: Ultrasonic type, light weight, Silent operation, compact, with nebulization control, inbuilt timer, frequency 1.7MHz or more, nebulizing rate of up to 5 l/min,		
	minimum drug nebulization flow ≤ 0.2 ml/min, water-based, complete with nebulization kit		
	of adult and child masks and small and big medication cups and, should have a dust filter,		
	nebulization accessing kit- flowrate ≥ 74m, air pressure≥ 35 PS		
2	Type: Portable		
3	Power requirements : 240V A/C, 50Hz, 13A, 3 pin single phase, Cable with PE		
4	Quality Standards		
	- Manufacturing Standards - ISO certified.		
	- Conformity to Standards - CE approved and marked		
	- UL approved		
5	Warranty: A minimum of 1 year after delivery (provide Warranty card)		

24. SYRINGE INFUSION PUMP

- 1. Should be easy to use and nurse friendly.
- 2. Should have automatic syringe size and model detection
- 3. Should have large format LCD/TFT display.
- 4. Should have a minimum flow rate range from 0.1–1200 ml/hr for 50ml syringe, 0.1–100 ml/hr for 20ml syringe and 0.1–60 ml/hr for 10ml syringe.
- 5. Syringe range from 20-50/60 ml.
- 6. Should have a flow rate accuracy of $\pm 2\%$
- 7. Should have a bolus rate up to 1000ml/hr for 50 ml syringe.
- 8. Should have automatic and manual bolus.
- 9. Should have at least 3 levels of programmable occlusion pressure.
- 10. Should have automatic bolus reduction system to avoid accidental bolus delivery after occlusion incident.
- 11. Should have a rechargeable battery with back up time of minimum 3 hours.
- 12. Pump must triger following alarms with visual indication:
 - i. Occlusion Pressure Alarm
 - ii. KVO or 3 min pre- alarm
 - iii. Syringe empty and volume infused alarm
 - iv. Internal malfunction and Battery Charge Low Alarm
 - v. Syringe disengaged and incorrectly placed alarm
 - vi. Alarm loudnesscontrol.
 - vii. No mains
 - viii. Line disconnected (rapid pressure drops).
- 13. Should work with input 200 to 240Vac 50 Hz supply.
- 14. Should have safety certificate from a competent authority CE / FDA (US) / STQC CB certificate / STQC S certificate or valid detailed electrical and functional safety test report
 - from ERT. Copy of the certificate/test report shall be produced along with the technical bid

25. PULSE OXIMETER

1 **Description:**

- Handheld pulse oximeter for monitoring peripheral oxygen saturation (SPO/Band pulse rate (PR) for adult, pediatric or neonatal patients.
- High resolution
- Alarm indication
- Rechargeable battery (Lithium-Ion) and or Alkaline AA with 2-hour charging time, 36hours.
- With rechargeable stand
- Power requirements: 240V A/C, 50Hz, 13A, 3 pin single phase, Cable with PE
- **Ouality Standards**
 - Manufacturing Standards
- ISO certified.
- Conformity to Standards
- CE approved and marked

26. AMBU BAG – ADULT

Technical Specification- Ambu Bag		
III	Ambu Bag – Adult	
1	Should have autoclave-at-134 deg. C silicon rubber bellow to withstand heat	
2	Should provide Oxygen connecting tube with autoclavable face mask	
3	Should be supplied with a carry pouch.	
4	It should have a bag volume of 1700 ml. and a variation of ± 100ml. will be accepted.	
5	Should have an expiratory resistance of 2.2cms of water	
6	Should have an inspiratory resistance of 3.3cms of water.	
7	It should have controlled flow rates and ventilation, and with reduced airway pressure.	
8	Should have a port in the bag to connect oxygen with reservoir bag	

27. AMBU BAG – PAEDIATRIC

Ambu Bag –Paeds		
1	Should have autoclave-at-134 deg. C silicon rubber bellow to withstand heat	
2	Should provide Oxygen connecting tube with autoclavable face mask	
3	Should be supplied with a carry pouch.	
4	It should have a bag volume of 500 ml. and a variation of \pm 100ml. will be accepted.	
5	Should have an expiratory resistance of 2.2cms of water	
6	Should have an inspiratory resistance of 3.3cms of water.	
7	It should have controlled flow rates and ventilation, and with reduced airway pressure.	
8	Should have a port in the bag to connect oxygen with reservoir bag	

28. DELIVERY SET

	Description	
1	Tray -18"x12" x2, 2 kidney dish 10", 2bowls 4", 2 kocher artery forceps straight, 16	
	7&18cm, 1 kocher artery forceps curved 16cm, 1 cord scissors 12.5cm, 1 Episiotomy	
	scissors 14cm, 1 mayor needle holder 18cm, 1 dissecting scissor sharp straight 15cm, 1	
	operating scissor sharp curved 15cm, 1 dissecting forceps toothed 14cm. Non toothed	
	dissecting forceps 15cm. 1 Rampleysponge holding forceps 25.5cm. 3 cusco vaginal	
	speculum small, medium, large. 2artery forceps curved10cm	
2	All stainless steel	

29. CAESAREAN SECTION SET KIT.

Contains stainless steel instruments as follows:

•	OPERATING/DISSECTING SCISSORS S/B STR 14.5CM	1
•	FOERSTER SPONGE SERRATED STR 25CM	4
•	ALLIS TISSUE FORCEPS 5X6 19CM	1
•	MAYO SCISSORS 17CM STR TC	1
•	DRESSING FORCEPS 20CM	1
•	TISSUE FORCEPS 1:2 20CM	6
•	HEANEY NEEDLE HOLDER 21CM TC	1
•	ROCHESTER PEAN FORCEPS CVD 16CM	1
•	SCALPEL HANDLE #3	1
•	GREEN-ARMYTAGE FORCEPS 21CM	1
•	JACKSON RETRACTOR 90X40MM	1
•	DELEE RETRACTOR UNIVERSAL 70X50MM	1
•	INSTRUMENT TRAY WITH COVER	1

- LARGE BOWLS 4"
- CORD CLAMPS 14.5 CM

2

30. PATIENT ADJUSTABLE BED

- Heavy Duty, standard Height Patient Bed, single crank.
- Perforated Heavy gauge Steel bed platform, Preferably Epoxy coated, 2-segment
- Backrest with Crank Mechanism
- Stainless Steel Head-side IV rod
- Headboard and Footboard, Preferably Epoxy coated Steel
- Mobile with 5", Heavy Duty, Silent, Swivel Polyamide Castors, 2 with braking system
- Complete with 5" Heavy duty Mattress covered with Mackintosh

31. BEDSIDE LOCKER

- o Heavy gauge Steel Bedside Cabinet (Locker), preferably Epoxy coated
- o One Cabinet door and drawer
- o Fibre finish top / ABS top
- o Mobile with 2 ½" heavy Duty Polyamide Castors, two with brake system.

32. DELIVERY BED, HYDRAULIC

1	Heavy Duty, Height adjustable - Hydraulic with foot pedal		
2	High stainless steel base,		
	Pair of heavy duty, Height adjustable / Collapsible Side Rails Preferably Stainless Steel		
	Stainless Steel Head-side Telescopic IV rod,		
	Antistatic leg,		
	Manual crutches with antistatic pads,		
	Liquid collection pan stainless steel,		
	Steel frame which is epoxy coated,		
	Leg support stands,		
	At least 5" thick mattress, with heavy duty mackintosh.		
	1 Detachable Stainless Steel Tray		
	Headboard, Detachable, Preferably Stainless Steel framed		
3	Auxiliary (Leg end) section can slide under the main section		
4	Provided with a pair of upholstery lithotomy positions		
5	Four Functions: Back rest-gas Spring, Height Adjustment –Hydraulic with minimum height of at least		
	600mm, Trendelenburg and re-Trendelenburg		
	Trendelenburg /Re-Trendelenburg positions adjustment through Crank Mechanism		
6	Mobile with 5", Heavy Duty, Silent, Swivel Polyamide – antistatic Castors, with central locking		
	mechanism.		
7	Quality Standards		
	- Manufacturing Standards - ISO certified.		
	- Conformity to Standards - CE approved and marked		

33. BABY COT

	Description
1	Heavy Duty, 2-section, Stainless steel Pediatric Cot Bed
2	Manually adjustable backrest

3	Pair of heavy duty, Height adjustable / Collapsible Side Rails		
4.	Size: standard		
5	Mobile with 5", Heavy Duty, Siler	Mobile with 5", Heavy Duty, Silent, Swivel Polyamide Castors, with braking system	
6	Heavy duty 5" mattress with heavy duty mackintosh		
7	Quality Standards		
	- Manufacturing Standards	- ISO certified.	
	- Conformity to Standards	- CE approved and marked	

34. EXAMINATION COUCH METAL

1	Description:				
	- Heavy duty, robust constructed manual examination couch with standard height.				
	- Fixed height hollow-section framework, firmly welded, All Stainless Steel				
	- Two sections, adjustable headrest				
	- Leather upholstery and rubber feet plugs				
2	Quality Standards				
	- Manufacturing Standards - ISO certified.				
	- Conformity to Standards - CE approved and marked				

35. TROLLEY INSTRUMENT

1	ITEM DESCRIPTION: Portable instrument stand, with removable instrument tray, height adjusted stand, All stainless steel, tubular frame, robust constructed, 100 mm Antistatic swivel castors,		
2	Quality Standards		
	- Manufacturing Standards	- ISO certified.	
	- Conformity to Standards	- CE approved and marked	

36. TROLLEY, DRESSING

1	ITEM DESCRIPTION: Instrument trolley for transporting instruments in the hospital. All		
	stainless steel, tubular frame, robust constructed, with two stainless steel shelves and guard rail on		
	three sides of the shelf top, on four (4) 100 mm Antistatic swivel castors, 2 lockable brakes.		
2	Size: Approximately L700 x W 460 x H 860 mm		
3	Quality Standards		
	- Manufacturing Standards - ISO certified.		
	- Conformity to Standards - CE approved and marked		

37. TROLLEY, GAS CYLINDER

1	ITEM DESCRIPTION: Medical gasses trolley for transporting medical gasses in the hospital.			
	- All 304-stainless steel, tubular frame, robust construct, Hind stand is foldable.			
	- On four (4) Castors (2 front fixed solid padded ≥8" wheel and 2 hind swivel with lockable			
	brake castors			
2	Size: Approximately 850x550x1,030 mm			
3	Quality Standards			
	- Manufacturing Standards - ISO certified.			
	- Conformity to Standards - CE approved and marked			

38. TROLLEY, GENERAL PURPOSE

1 ITEM DESCRIPTION: General purpose trolley for transporting items in the hospital.
--

	- All 304-stainless steel, tubular frame, robust construct			
	- Can carry up to 200 kgs load			
	- With three shelves			
	- 4 heavy duty castors – 6"			
2	Size: Approximately			
3	Quality Standards			
	- Manufacturing Standards - ISO certified.			
	- Conformity to Standards - CE approved and marked			

39. TROLLEY EMERGENCY

Should be supplied with:

- 1. Aluminum alloy column with ABS engineering plastic construction
- 2. Base with dia–125mm castors, two of which with brakes
- 3. Five drawers (2 small, 2 middle & 1 large) small and middle drawers inner with medicine divider can be organized easily and freely
- 4. With related accessories for emergency as bellow:
- IV pole; waste bin; file cassette; needle disposal container; defibrillator shelf; lateral side with sliding working shelf; back with ABS plastic cardiac boards; oxygen tank holder
- 5.Central locking system
 - Smooth rounded corners to assist cleaning
 - Strapless defibrillator platform on swingarm adjusts to accommodate various modes
 - Size of approximately 850 x 520 x 1040mm

Quality Standards

- Manufacturing Standards - ISO certified.
- Conformity to Standards - CE approved and marked

40. TROLLEY, MEDICINE / DRUG

1	ITEM DESCRIPTION: Transportation drug trolley, Frame work of Stainless Steel tube			
	mounted on 10 cms Antistatic swivel castors, Stainless Steel shelves with three side railing			
	on top shelf, Two drawer under the shelf			
2	Size: Approximately L650 x W 400 x H 900 mm			
3	Quality Standards			
	- Manufacturing Standards - ISO certified.			
	- Conformity to Standards - CE approved and marked			

41. TROLLEY PATIENT STRETCHER						
1	Gene	General Description:				
	Resuscitation patient trolley with IV pole, Oxygen Cylinders and with adjustable sides, constructed					
	from	preferably chrome plated	mild steel and mobile on castors.			
2	Physical Specifications					
	2.1	Movements	- Back rest, trendelenburg/reverse trendelenburg, up and down			
	2.2	Operation	- By hydraulic mechanical system			
	2.3	Side guard rails	- Foldaway store away under the mattress footprint			
	2.4	Mattress	- High density, water proof and fire resistance			
	2.5	Height adjustment	- Dual sided foot pedal			
	2.6	Mobile	- On four antistatic smooth castors diameter 150mm with brakes			
			and central locking system with Retractable 5th wheel steering system			
	2.7	IV pole	- Provided			
	2.8	Storage compartment	- For file storage and other accessories			
	2.9	Oxygen cylinder	- Trough provided, Medium size 11kg (1.36m²) mild, bullnose			
	2.10	Resuscitation bags	- Provided, adult and pediatric			

	2.11	Weight to handle	- Approximately 200- 300kgs	
3	Approximate Size: Dimensions (Overall) 2050(L) X 780(W) X 620 - 900(H) mm			
4	Qualit	ty Standards		
	- Man	ufacturing Standards	- ISO certified.	
	- Conf	Formity to Standards	- CE approved and marked	

42. WHEEL CHAIR ADULT

1	Wheel chairs- Folding type, steel type, solid rubber wheels padded footrest (fixed)		
2	Size: Approximately 940 x 230 x 890mm		
3	Capacity: 200kg		
4	Model: Contrex or its equivalent		
5	Quality Standards		
	- Manufacturing Standards - ISO certified.		
	- Conformity to Standards - CE approved and marked		

Detailed Technical Specifications and Standards [insert whenever necessary]. [Insert detailed description of TS]

4.	Drawings	N/A

This Tendering document includes	[Insert "th	le
following" or "no"]drawings. [If documents shall	be included, insert the followi	ng
List of Drawings l.		

List of Drawings			
Drawing No.	Drawing Name	Purpose	

5. Inspections and Tests

The following inspections and tests shall be performed	.[Insert list of
inspections and	
tests]	

PART 3 - CONDITIONS OF CONTRACT AND CONTRACT FORMS

SECTION VI - GENERAL CONDITIONS OF CONTRACT

1. Definitions

In the Conditions of Contract ("these Conditions"), which include Special Conditions, Parts A and B, and these General Conditions, the following words and expressions shall have the meanings stated. Words indicating persons or parties include corporations and other legal entities, except where the context requires otherwise.

- a) "Contract" means the Contract Agreement entered into between the Procuring Entity and the Supplier, together with the Contract Documents referred to therein, including all attachments, appendices, and all documents incorporated by reference therein.
- b) "Contract Documents" means the documents listed in the Contract Agreement, including any amendments thereto.
- c) "Contract Price" means the price payable to the Supplier as specified in the Contract Agreement, subject to such additions and adjustments thereto or deductions therefrom, as may be made pursuant to the Contract.
- d) "Day" means calendar day.
- e) "Completion" means the fulfilment of the Related Services by the Supplier in accordance with the terms and conditions set forth in the Contract.
- f) "GCC" means the General Conditions of Contract.
- g) "Goods" means all of the commodities, raw material, machinery and equipment, and/or other materials that the Supplier is required to supply to the Procuring Entity under the Contract.
- h) "Procuring Entity" means the Procuring Entity purchasing the Goods and Related Services, as specified in the SCC.
- i) "Related Services" means the services incidental to the supply of the goods, such as insurance, delivery, installation, commissioning, training and initial maintenance and other such obligations of the Supplier under the Contract.
- j) "SCC" means the Special Conditions of Contract.
- k) "Subcontractor" means any person, private or government entity, or a combination of the above, to whom any part of the Goods to be supplied or execution of any part of the Related Services is subcontracted by the Supplier.
- 1) "Supplier" means the person, private or government entity, or a combination of the above, whose Tender to perform the Contract has been accepted by the Procuring Entity and is named as such in the Contract Agreement.
- m) "Base Date" means a date 30 day prior to the submission of tenders.
- n) "Laws" means all national legislation, statutes, ordinances, and regulations and by-laws of any legally constituted public authority.
- o) "Letter of Acceptance" means the letter of formal acceptance, signed by the contractor. Procuring Entity, including any annexed memoranda comprising agreements between and signed by both Parties.
- p) "Procuring Entity" means the Entity named in the Special Conditions of Contract.

2. Interpretation

21. If the context so requires it, singular means plural and vice versa.

22. Incoterms

- a) Unless inconsistent with any provision of the Contract, the meaning of any trade term and the rights and obligations of parties thereunder shall be as prescribed by Incoterms **specified in the SCC**.
- b) The terms EXW and CIP and other similar terms, when used, shall be governed by the rules prescribed in the current edition of Incoterms specified in the SCC and published by the International Chamber of Commerce in Paris, France.

3. Contract Documents

Subject to the order of precedence set forth in the Contract Agreement, all documents forming the Contract (and all parts thereof) are intended to be correlative, complementary, and mutually explanatory. The Contract Agreement shall be read as a whole. The documents forming the Contract shall be interpreted in the following order of priority:

- a) the Contract Agreement,
- b) the Letter of Acceptance,
- c) the General Conditions of Contract
- d) Special Conditions of Contract
- e) the Form of Tender,
- f) the Specifications and Schedules of the Drawings (if any), and
- g) the Schedules of Requirements, Price Schedule and any other documents forming part of the Contract.

4. Fraud and Corruption

- 3.1 The supplier shall comply with anti-corruption laws and guidelines and the prevailing sanctions, policies and procedures as set forth in the Laws of Kenya.
- The Supplier shall disclose any commissions, gratuity or fees that may have been paid or are tobe paid to agents or any other person with respect to the Tendering process or execution of the Contract. The information disclosed must include at least the name and address of the agent or other party, the amount and currency, and the purpose of the commission, gratuity or fee.

4.1 Entire Agreement

4.3.1 The Contract constitutes the entire agreement between the Procuring Entity and the Supplier and supersedes all communications, negotiations and agreements (whether written or oral) of the parties with respect thereto made prior to the date of Contract.

4.2 Amendment

No amendment or other variation of the Contract shall be valid unless it is in writing, is dated, expressly refers to the Contract, and is signed by a duly authorized representative of each partythereto.

4.3 Non-waiver

- a) Subject to GCC Sub-Clause 4.5(b) below, no relaxation, forbearance, delay, or indulgence byeither party in enforcing any of the terms and conditions of the Contract or the granting of time by either party to the other shall prejudice, affect, or restrict the rights of that party under the Contract, neither shall any waiver by either party of any breach of Contractoperate as waiver of any subsequent or continuing breach of Contract.
- b) Any waiver of a party's rights, powers, or remedies under the Contract must be in writing, dated, and signed by an authorized representative of the party granting such waiver, andmust specify the right and the extent to which it is being waived.

4.4 Severability

If any provision or condition of the Contract is prohibited or rendered invalid or unenforceable,

such prohibition, invalidity or unenforceability shall not affect the validity or enforceability of any other provisions and conditions of the Contract.

5. Language

- 5.1 The Contract as well as all correspondence and documents relating to the Contract exchanged by the Supplier and the Procuring Entity, shall be written in the **English Language**. Supporting documents and printed literature that are part of the Contract may be in another language provided they are accompanied by an accurate and certified translation of the relevant passages in the **English Language**, in which case, for purposes of interpretation of the Contract, the English language is translation shall govern.
- 52 The Supplier shall bear all costs of translation to the governing language and all risks of theaccuracy of such translation, for documents provided by the Supplier.

6. Joint Venture, Consortium or Association

6.1 If the Supplier is a joint venture, consortium, or association, all of the parties shall be jointly andseverally liable to the Procuring Entity for the fulfilment of the provisions of the Contract and shall designate one member of the joint venture, consortium, or association to act as a leaderwith authority to bind the joint venture, consortium, or association. The composition or the constitution of the joint venture, consortium, or association shall not be altered without the prior written consent of the Procuring Entity.

7. Eligibility

- 7.1 The Supplier and its Subcontractors shall have the nationality of an eligible country. A Supplier or Sub-contractor shall be deemed to have the nationality of a country if it is a citizen or constituted, incorporated, or registered, and operates in conformity with the provisions of the laws of that country.
- All Goods and Related Services to be supplied under the Contract shall have their origin in Eligible Countries. For the purpose of this Clause, origin means the country where the goodshave been grown, mined, cultivated, produced, manufactured, or processed; or through manufacture, processing, or assembly, another commercially recognized article results that differs substantially in its basic characteristics from its components.
- 73 The Tenderer, if a Kenyan firm, must submit with its tender a valid tax compliance certificate from the Kenya Revenue Authority.

8. Notices

- 8.1 Any notice given by one party to the other pursuant to the Contract shall be in writing to the address specified in the SCC. The term "in writing" means communicated in written form withproof of receipt.
- 82 A notice shall be effective when delivered or on the notice's effective date, whichever is later.

9. Governing Law

- 9.1 The Contract shall be governed by and interpreted in accordance with the laws of Kenya.
- Throughout the execution of the Contract, the Supplier shall comply with the import of goods andservices prohibitions in Kenya:
 - a) where, as a matter of law, compliance or official regulations, Kenya prohibits commercial relations with that country or any import of goods from that country or any payments toany country, person, or entity in that country; or
 - b) by an act of compliance with a decision of the United Nations Security Council taken underChapter VII of the Charter of the United Nations, Kenya prohibits any import of goodsfrom that country or any payments to any country, person, or entity.

10. Settlement of Disputes

10.1 The Procuring Entity and the Supplier shall make every effort to resolve amicably by direct negotiation any disagreement or dispute arising between them under or in connection with the Contract.

10.2 If, after thirty (30) days, the parties have failed to resolve their dispute or difference by such mutual consultation, then either the Procuring Entity or the Supplier may give notice to the other party of its intention to commence arbitration, as hereinafter provided, as to the matter in dispute, and no arbitration in respect of this matter may be commenced unless such notice is given. Any dispute or difference in respect of which a notice of intention to commence arbitration has been given in accordance with this Clause shall be finally settled by arbitration. Arbitration may be commenced prior to or after delivery of the Goods under the Contract.

102 Arbitration proceedings shall be conducted as follows:

- 1021 Any claim or dispute between the Parties arising out of or in connection with the Contract not settled amicably in accordance with Sub-Clause 10.1 shall be finally settled by arbitration.
- 1022 No arbitration proceedings shall be commenced on any claim or dispute where notice of a claimor dispute has not been given by the applying party within thirty days of the occurrence or discovery of the matter or issue giving rise to the dispute.
- 1023 Notwithstanding the issue of a notice as stated above, the arbitration of such a claim or dispute shall not commence unless an attempt has in the first instance been made by the parties to settle such claim or dispute amicably with or without the assistance of third parties. Proof of such attempt shall be required.
- 1024 The Arbitrator shall, without prejudice to the generality of his powers, have powers to direct such measurements, computations, or valuations as may in his opinion be desirable in order to determine the rights of the parties and assess and award any sums which ought to have been the subject of or included in any due payments.
- 1025 Neither Party shall be limited in the proceedings before the arbitrators to the evidence, or to the reasons for the dispute given in its notice of a claim or dispute.
- 1026 Arbitration may be commenced prior to or after delivery of the goods. The obligations of the Parties shall not be altered by reason of any arbitration being conducted during the progress of the delivery of goods.
- 1027 The terms of the remuneration of each or all the members of Arbitration shall be mutually agreedupon by the Parties when agreeing the terms of appointment. Each Party shall be responsible for paying one-half of this remuneration.

10.3 Arbitration Proceedings

- Arbitration proceedings with national suppliers will be conducted in accordance with the Arbitration Laws of Kenya. In case of any claim or dispute, such claim or dispute shall be notified in writing by either party to the other with a request to submit it to arbitration and to concur in the appointment of an Arbitrator within thirty days of the notice. The dispute shall be referred to the arbitration and final decision of a person or persons to be agreed between the parties. Failing agreement to concur in the appointment of an Arbitrator, the Arbitrator shall be appointed, on the request of the applying party, by the Chairman or Vice Chairman of any of thefollowing professional institutions;
 - i) Kenya National Chamber of Commerce
 - ii) Chartered Institute of Arbitrators (Kenya Branch)
 - iii) The Law Society of Kenya
- 1032 The institution written to first by the aggrieved party shall take precedence over all other institutions.

10.4 Arbitration with Foreign Suppliers

104.1 Arbitration with foreign suppliers shall be conducted in accordance with the arbitration rules of the United Nations Commission on International Trade Law (UNCITRAL); or with proceedings administered by the International Chamber of Commerce (ICC) and conducted under the ICC Rules of Arbitration; by one or more arbitrators appointed in accordance with said arbitration

1042 The place of arbitration shall be a location specified in the SCC; and the arbitration shall be conducted in the language for communications defined in Sub-Clause 1.4 [Law and Language].

10.5 Alternative Arbitration Proceedings

Alternatively, the Parties may refer the matter to the Nairobi Centre for International Arbitration (NCIA) which offers a neutral venue for the conduct of national and international arbitration with commitment to providing institutional support to the arbitral process.

10.6 Failure to Comply with Arbitrator's Decision

- 1061 The award of such Arbitrator shall be final and binding upon the parties.
- 10.6.1 In the event that a Party fails to comply with a final and binding Arbitrator's decision, then the other Party may, without prejudice to any other rights it may have, refer the matter to a competent court of law.

10.7 Contract operations continue

Notwithstanding any reference to arbitration herein,

- a) the parties shall continue to perform their respective obligations under the Contract unless they otherwise agree; and
- b) the Procuring Entity shall pay the Supplier any monies due the Supplier.

11. Inspections and Audit by the Procuring Entity

- 11.1 The Supplier shall keep, and shall cause its Subcontractors to keep, accurate and systematic accounts and records in respect of the Goods in such form and details as will clearly identify relevant time, changes and costs.
- Pursuant to paragraph 2.2 of Instruction to Tenderers, the Supplier shall permit and shall cause its subcontractors to permit, the Procuring Entity and/or persons appointed by the Procuring Entityor by other statutory bodies of the Government to inspect the Site and/or the accounts andrecords relating to the procurement process, selection and/or contract execution, and to have such accounts and records audited by auditors appointed by the Procuring Entity. The Supplier's andits Subcontractors' attention is drawn to Sub- Clause 3.1 which provides, inter alia, that acts intended to materially impede the exercise of the Procuring Entity's inspection and audit rights constitute a prohibited practice subject to contract termination, as well as to a determination of ineligibility.

12. Scope of Supply

12.1 The Goods and Related Services to be supplied shall be as specified in the Schedule of Requirements.

13. Delivery and Documents

13.1 Subject to GCC Sub-Clause 33.1, the delivery of the Goods and completion of the RelatedServices shall be in accordance with the List of Goods and Delivery Schedule specified in the Supply Requirements. The details of shipping and other documents to be furnished by the Supplier are specified in the SCC.

14. Supplier's Responsibilities

14.1 The Supplier shall supply all the Goods and Related Services included in the Scope of Supply in accordance with GCC Clause 12, and the Delivery and Completion Schedule, as per GCC Clause 13.

15. Contract Price

15.1 Prices charged by the Supplier for the Goods supplied and the Related Services performed under

the Contract shall not vary from the prices quoted by the Supplier in its Tender, with the exception of any price adjustments authorized in the SCC.

Where the contract price is different from the corrected tender price, in order to ensure the supplier is not paid less or more relative to the contract price (*which would be the tender price*), any partial payment valuation based on rates in the schedule of prices in the Tender, will be adjusted by a <u>plus or minus</u> percentage. The percentage already worked out during tender evaluation is worked out as follows: (*corrected tender price – tender price*)/tender price X 100.

16. Terms of Payment

- 16.1 The Supplier shall request for payment by submitting invoice(s), delivery note(s) and any other relevant documents as specified in the **SCC** to the Procuring Entity.
- Payments shall be made promptly by the Procuring Entity, but not later than thirty (30) days after submission of an invoice by the Supplier, and after the Procuring Entity has accepted it.
- Where a Procuring Entity rejects Goods and Related Services, in part or wholly, the procuring Entity shall promptly inform the Supplier to collect, replace or rectify as appropriate and give reasons for rejection. The Supplier shall submit a fresh invoice, delivery note and any other relevant documents as specified in the SCC.
- 164 The currencies in which payments shall be made to the Supplier under this Contract shall bethose in which the Tender price is expressed.
- In the event that the Procuring Entity fails to pay the Supplier any payment by its due date or within the period set forth in the SCC, the Procuring Entity may pay to the Supplier interest on the amount of such delayed payment at the rate shown in the SCC, for the period of delay untilpayment has been made in full, whether before or after judgment or arbitrage award.

17. Taxes and Duties

- 17.1 The Supplier shall be entirely responsible for all taxes, duties, license fees, and other such levies incurred to deliver the Goods and Related Services to the Procuring Entity at the final delivery point.
- 17.3 If any tax exemptions, reductions, allowances or privileges may be available to the Supplier in Kenya, the Supplier shall inform the Procuring Entity and the Procuring Entity shall use its best efforts to enable the Supplier to benefit from any such tax savings to the maximum allowable extent.

18. Performance Security

- 181 If required as specified in the SCC, the Supplier shall, within twenty-eight (28) days of the notification of contract award, provide a performance security for the performance of the Contract in the amount specified in the SCC.
- The proceeds of the Performance Security shall be payable to the Procuring Entity as compensation for any loss resulting from the Supplier's failure to complete its obligations under the Contract.
- As specified in **the SCC**, the Performance Security, if required, shall be denominated in the currency(ies) of the Contract, or in a freely convertible currency acceptable to the Procuring Entity; and shall be in one of the formats stipulated by the Procuring Entity in **the SCC**, or in another format acceptable to the Procuring Entity.
- 184 The Performance Security shall be discharged by the Procuring Entity and returned to the Supplier not later than thirty (30) days following the date of Completion of the Supplier'sperformance obligations under the Contract, including any warranty obligations, unless specified otherwise in the SCC.

19. Copyright

19.1 The copyright in all drawings, documents, and other materials containing data and information furnished to the Procuring Entity by the Supplier herein shall remain vested in the Supplier, or,

if they are furnished to the Procuring Entity directly or through the Supplier by any third party, including suppliers of materials, the copyright in such materials shall remain vested in such third party.

20. Confidential Information

- 20.1 The Procuring Entity and the Supplier shall keep confidential and shall not, without the written consent of the other party hereto, divulge to any third party any documents, data, or other information furnished directly or indirectly by the other party hereto in connection with the Contract, whether such information has been furnished prior to, during or following completion ortermination of the Contract. Notwithstanding the above, the Supplier may furnish to its Sub-Supplier such documents, data, and other information it receives from the Procuring Entity to the extent required for the Sub Supplier to perform its work under the Contract, in which event the Supplier shall obtain from such Sub Supplier undertaking of confidentiality similar to that imposed on the Supplier under GCC Clause 20.
- 20.2 The Procuring Entity shall not use such documents, data, and other information received from the Supplier for any purposes unrelated to the contract. Similarly, the Supplier shall not use such documents, data, and other information received from the Procuring Entity for any purpose other than the performance of the Contract.
- 20.3 The obligation of a party under GCC Sub-Clauses 20.1 and 20.2 above, however, shall not applyto information that:
 - a) the Procuring Entity or Supplier need to share with other arms of Government or other bodies participating in the financing of the Contract; such parties shall de disclosed in **the SCC**;
 - b) now or hereafter enters the public domain through no fault of that party;
 - c) can be proven to have been possessed by that party at the time of disclosure and whichwas not previously obtained, directly or indirectly, from the other party; or
 - **d**) otherwise lawfully becomes available to that party from a third party that has no obligation of confidentiality.
- 20.4 The above provisions of GCC Clause 20 shall not in any way modify any undertaking of confidentiality given by either of the parties hereto prior to the date of the Contract in respect of the Supply or any part thereof.
- 20.5 The provisions of GCC Clause 20 shall survive completion or termination, for whatever reason, of the Contract.

21. Subcontracting

- 21.1 The Supplier shall notify the Procuring Entity in writing of all subcontracts awarded under the Contract if not already specified in the Tender. Such notification, in the original Tender or later shall not relieve the Supplier from any of its obligations, duties, responsibilities, or liability underthe Contract.
- 212 Subcontracts shall comply with the provisions of GCC Clauses 3 and 7.

22. Specifications and Standards

- 22.1 Technical Specifications and Drawings
 - a) The Goods and Related Services supplied under this Contract shall conform to the technical specifications and standards mentioned in Section VI, Schedule of Requirements and, whenno applicable standard is mentioned, the standard shall be equivalent or superior to the official standards whose application is appropriate to the Goods' country of origin.
 - b) The Supplier shall be entitled to disclaim responsibility for any design, data, drawing, specification or other document, or any modification thereof provided or designed by or on behalf of the Procuring Entity, by giving a notice of such disclaimer to the ProcuringEntity.
 - c) Wherever references are made in the Contract to codes and standards in accordance with

which it shall be executed, the edition or the revised version of such codes and standards shall be those specified in the Schedule of Requirements. During Contract execution, any changes in any such codes and standards shall be applied only after approval by the Procuring Entity and shall be treated in accordance with GCC Clause 33.

23. Packing and Documents

- 23.1 The Supplier shall provide such packing of the Goods as is required to prevent their damage or deterioration during transit to their final destination, as indicated in the Contract. During transit, the packing shall be sufficient to withstand, without limitation, rough handling and exposure to extreme temperatures, salt and precipitation, and open storage. Packing case size and weights shall take into consideration, where appropriate, the remoteness of the goods' final destination and the absence of heavy handling facilities at all points in transit.
- The packing, marking, and documentation within and outside the packages shall comply strictly with such special requirements as shall be expressly provided for in the Contract, including additional requirements, if any, specified **in the SCC**, and in any other instructions ordered bythe Procuring Entity.

24. Insurance

24.1 Unless otherwise specified in the **SCC**, the Goods supplied under the Contract shall be fully insured—in a freely convertible currency from an eligible country—against loss or damage incidental to manufacture or acquisition, transportation, storage, and delivery, in accordance with the applicable Incoterms or in the manner specified in the **SCC**.

25. Transportation and Incidental Services

- 25.1 Unless otherwise specified in the SCC, responsibility for arranging transportation of the Goods shall be in accordance with the specified Incoterms.
- The Supplier may be required to provide any or all of the following services, including additional services, if any, specified in SCC:
 - a) performance or supervision of on-site assembly and/or start-up of the supplied Goods;
 - b) furnishing of tools required for assembly and/or maintenance of the supplied Goods;
 - c) furnishing of a detailed operations and maintenance manual for each appropriate unit of the supplied Goods;
 - d) performance or supervision or maintenance and/or repair of the supplied Goods, for a period of time agreed by the parties, provided that this service shall not relieve the Supplier of any warranty obligations under this Contract; and
 - e) training of the Procuring Entity's personnel, at the Supplier's plant and/or on-site, in assembly, start-up, operation, maintenance, and/or repair of the supplied Goods.
- Prices charged by the Supplier for incidental services, if not included in the Contract Price for the Goods, shall be agreed upon in advance by the parties and shall not exceed the prevailing rates charged to other parties by the Supplier for similar services

26. Inspections and Tests

- The Supplier shall at its own expense and at no cost to the Procuring Entity carry out all such tests and/or inspections of the Goods and Related Services as are specified in the SCC.
- The inspections and tests may be conducted on the premises of the Supplier or its Subcontractor, at point of delivery, and/or at the Goods' final destination, or in another place in Kenya as specified in the SCC. Subject to GCC Sub-Clause 26.3, if conducted on the premises of the Supplier or its Subcontractor, all reasonable facilities and assistance, including access to drawings and production data, shall be furnished to the inspectors at no charge to the Procuring Entity.
- 263 The Procuring Entity or its designated representative shall be entitled to attend the tests and/or inspections referred to in GCC Sub-Clause 26.2, provided that the Procuring Entity bear all of itsown costs and expenses incurred in connection with such attendance including, but not limited to,

all travelling and board and lodging expenses.

- Whenever the Supplier is ready to carry out any such test and inspection, it shall give areasonable advance notice, including the place and time, to the Procuring Entity. The Supplier shall obtain from any relevant third party or manufacturer any necessary permission or consent to enable the Procuring Entity or its designated representative to attend the test and/or inspection.
- The Procuring Entity may require the Supplier to carry out any test and/or inspection not required by the Contract but deemed necessary to verify that the characteristics and performance of the Goods comply with the technical specifications codes and standards under the Contract, provided that the Supplier's reasonable costs and expenses incurred in the carrying out of such test and/or inspection shall be added to the Contract Price. Further, if such test and/or inspection impedes the progress of manufacturing and/or the Supplier's performance of its other obligations under the Contract, due allowance will be made in respect of the Delivery Dates and Completion Dates and the other obligations so affected.
- The Supplier shall provide the Procuring Entity with a report of the results of any such test and/or inspection.
- The Procuring Entity may reject any Goods or any part thereof that fail to pass any test and/or inspection or do not conform to the specifications. The Supplier shall either rectify or replacesuch rejected Goods or parts thereof or make alterations necessary to meet the specifications atno cost to the Procuring Entity, and shall repeat the test and/or inspection, at no cost to the Procuring Entity, upon giving a notice pursuant to GCC Sub- Clause 26.4.
- The Supplier agrees that neither the execution of a test and/or inspection of the Goods or any part thereof, nor the attendance by the Procuring Entity or its representative, nor the issue of anyreport pursuant to GCC Sub-Clause 26.6, shall release the Supplier from any warranties or other obligations under the Contract.

27. Liquidated Damages

27.1 Except as provided under GCC Clause 32, if the Supplier fails to deliver any or all of the Goods by the Date(s) of delivery or perform the Related Services within the period specified in the Contract, the Procuring Entity may without prejudice to all its other remedies under the Contract, deduct from the Contract Price, as liquidated damages, a sum equivalent to the percentage specified in the SCC of the delivered price of the delayed Goods or unperformed Services for each week or part thereof of delay until actual delivery or performance, up to a maximum deduction of the percentage specified in those SCC. Once the maximum is reached, the ProcuringEntity may terminate the Contract pursuant to GCC Clause 35.

28. Warranty

- 28.1 The Supplier warrants that all the Goods are new, unused, and of the most recent or current models, and that they incorporate all recent improvements in design and materials, unless provided otherwise in the Contract.
- Subject to GCC Sub-Clause 22.1(b), the Supplier further warrants that the Goods shall be free from defects arising from any act or omission of the Supplier or arising from design, materials, and workmanship, under normal use in the conditions prevailing in the country of final destination.
- Unless otherwise specified in the **SCC**, the warranty shall remain valid for twelve (12) months after the Goods, or any portion thereof as the case may be, have been delivered to and accepted at the final destination indicated in the **SCC**, or for eighteen (18) months after the date of shipment from the port or place of loading in the country of origin, whichever period concludes earlier.
- 28.4 The Procuring Entity shall give notice to the Supplier stating the nature of any such defects together with all available evidence thereof, promptly following the discovery thereof. The Procuring Entity shall afford all reasonable opportunity for the Supplier to inspect such defects.
- 285 Upon receipt of such notice, the Supplier shall, within the period specified in the SCC,

expeditiously repair or replace the defective Goods or parts thereof, at no cost to the Procuring Entity.

If having been notified, the Supplier fails to remedy the defect within the period specified in the SCC, the Procuring Entity may proceed to take within a reasonable period such remedial actionas may be necessary, at the Supplier's risk and expense and without prejudice to any other rights which the Procuring Entity may have against the Supplier under the Contract.

29. Patent Indemnity

- 29.1 The Supplier shall, subject to the Procuring Entity's compliance with GCC Sub-Clause 29.2, indemnify and hold harmless the Procuring Entity and its employees and officers from and against any and all suits, actions or administrative proceedings, claims, demands, losses, damages, costs, and expenses of any nature, including attorney's fees and expenses, which the Procuring Entity may suffer as a result of any infringement or alleged infringement of any patent, utility model, registered design, trademark, copyright, or other intellectual property right registered or otherwise existing at the date of the Contract by reason of:
 - a) the installation of the Goods by the Supplier or the use of the Goods in the country where the Site is located; and
 - b) the sale in any country of the products produced by the Goods.

Such indemnity shall not cover any use of the Goods or any part thereof other than for the purpose indicated by or to be reasonably inferred from the Contract, neither any infringementresulting from the use of the Goods or any part thereof, or any products produced therebyin association or combination with any other equipment, plant, or materials not supplied by the Supplier, pursuant to the Contract.

- If any proceedings are brought or any claim is made against the Procuring Entity arising out of the matters referred to in GCC Sub-Clause 29.1, the Procuring Entity shall promptly give the Supplier a notice thereof, and the Supplier may at its own expense and in the Procuring Entity's name conduct such proceedings or claim and any negotiations for the settlement of any such proceedings or claim.
- If the Supplier fails to notify the Procuring Entity within twenty-eight (28) days after receipt of such notice that it intends to conduct any such proceedings or claim, then the Procuring Entity shall be free to conduct the same on its own behalf.
- The Procuring Entity shall, at the Supplier's request, afford all available assistance to the Supplierin conducting such proceedings or claim, and shall be reimbursed by the Supplier for all reasonable expenses incurred in so doing.
- The Procuring Entity shall indemnify and hold harmless the Supplier and its employees, officers, and Subcontractors from and against any and all suits, actions or administrative proceedings, claims, demands, losses, damages, costs, and expenses of any nature, including attorney's fees and expenses, which the Supplier may suffer as a result of any infringement or alleged infringement of any patent, utility model, registered design, trademark, copyright, or other intellectual property right registered or otherwise existing at the date of the Contract arising out of or in connection with any design, data, drawing, specification, or other documents or materials provided or designed by or on behalf of the Procuring Entity.

30. Limitation of Liability

- 30.1 Except in cases of criminal negligence or willful misconduct,
 - a) the Supplier shall not be liable to the Procuring Entity, whether in contract, tort, or otherwise, for any indirect or consequential loss or damage, loss of use, loss of production, or loss of profits or interest costs, provided that this exclusion shall not apply to any obligation of the Supplier to pay liquidated damages to the Procuring Entity, and
 - b) the aggregate liability of the Supplier to the Procuring Entity, whether under the Contract, intort or otherwise, shall not exceed the total Contract Price, provided that this limitation shall not apply to the cost of repairing or replacing defective equipment, or to any obligation of the

supplier to indemnify the Procuring Entity with respect to patent infringement.

31. Change in Laws and Regulations

31.1 Unless otherwise specified in the Contract, if after the date of 30 days prior to date of Tender submission, any law, regulation, ordinance, order or bylaw having the force of law is enacted, promulgated, abrogated, or changed in Kenya (which shall be deemed to include any change in interpretation or application by the competent authorities) that subsequently affects the Delivery Date and/or the Contract Price, then such Delivery Date and/or Contract Price shall be correspondingly increased or decreased, to the extent that the Supplier has thereby been affected in the performance of any of its obligations under the Contract. Notwithstanding the foregoing, such additional or reduced cost shall not be separately paid or credited if the same has already been accounted for in the price adjustment provisions where applicable, in accordance with GCC Clause 15.

32. Force Majeure

- The Supplier shall not be liable for forfeiture of its Performance Security, liquidated damages, or termination for default if and to the extent that its delay in performance or other failure to perform its obligations under the Contract is the result of an event of Force Majeure.
- For purposes of this Clause, "Force Majeure" means an event or situation beyond the control of the Supplier that is not foreseeable, is unavoidable, and its origin is not due to negligence orlack of care on the part of the Supplier. Such events may include, but not be limited to, acts of the Procuring Entity in its sovereign capacity, wars or revolutions, fires, floods, epidemics, quarantine restrictions, and freight embargoes.
- 32.3 If a Force Majeure situation arises, the Supplier shall promptly notify the Procuring Entity in writing of such condition and the cause thereof. Unless otherwise directed by the Procuring Entity in writing, the Supplier shall continue to perform its obligations under the Contract as far as is reasonably practical, and shall seek all reasonable alternative means for performance not prevented by the Force Majeure event.

33. Change Orders and Contract Amendments

- 33.1 The Procuring Entity may at any time order the Supplier through notice in accordance GCC Clause 8, to make changes within the general scope of the Contract in any one or more of the following:
 - a) drawings, designs, or specifications, where Goods to be furnished under the Contract are to be specifically manufactured for the Procuring Entity;
 - b) the method of shipment or packing;
 - c) the place of delivery; and
 - d) the Related Services to be provided by the Supplier.
- If any such change causes an increase or decrease in the cost of, or the time required for, the Supplier's performance of any provisions under the Contract, an equitable adjustment shall be made in the Contract Price or in the Delivery/Completion Schedule, or both, and the Contract shall accordingly be amended. Any claims by the Supplier for adjustment under this Clause must be asserted within twenty-eight (28) days from the date of the Supplier's receipt of the Procuring Entity's change order.
- Prices to be charged by the Supplier for any Related Services that might be needed but which were not included in the Contract shall be agreed upon in advance by the parties and shall not exceed the prevailing rates charged to other parties by the Supplier for similar services.
- Value Engineering: The Supplier may prepare, at its own cost, a value engineering proposal at any time during the performance of the contract. The value engineering proposal shall, at a minimum, include the following;
 - a) the proposed change(s), and a description of the difference to the existing contract requirements;
 - b) a full cost/benefit analysis of the proposed change(s) including a description and estimate of

- costs (including life cycle costs) the Procuring Entity may incur in implementing the value engineering proposal; and
- c) a description of any effect(s) of the change on performance/functionality.
- 33.5 The Procuring Entity may accept the value engineering proposal if the proposal demonstrates benefits that:
 - a) accelerates the delivery period; or
 - b) reduces the Contract Price or the life cycle costs to the Procuring Entity; or
 - c) improves the quality, efficiency or sustainability of the Goods; or
 - d) yields any other benefits to the Procuring Entity, without compromising the necessary functions of the Facilities.
- 33.6 If the value engineering proposal is approved by the Procuring Entity and results in:
 - a) a reduction of the Contract Price; the amount to be paid to the Supplier shall be the percentage specified **in the SCC** of the reduction in the Contract Price; or
 - b) an increase in the Contract Price; but results in a reduction in life cycle costs due to any benefit described in
 - (a) to (d) above, the amount to be paid to the Supplier shall be the full increase in the Contract Price.
- 33.7 Subject to the above, no variation in or modification of the terms of the Contract shall be made except by written amendment signed by the parties.

34. Extensions of Time

- 34.1 If at any time during performance of the Contract, the Supplier or its subcontractors should encounter conditions impeding timely delivery of the Goods or completion of Related Services pursuant to GCC Clause 13, the Supplier shall promptly notify the Procuring Entity in writing of the delay, its likely duration, and its cause. As soon as practicable after receipt of the Supplier's notice, the Procuring Entity shall evaluate the situation and may at its discretion extend the Supplier's time for performance, in which case the extension shall be ratified by the parties by amendment of the Contract.
- Except in case of Force Majeure, as provided under GCC Clause 32, a delay by the Supplier in the performance of its Delivery and Completion obligations shall render the Supplier liable to the imposition of liquidated damages pursuant to GCC Clause 26, unless an extension of time is agreed upon, pursuant to GCC Sub-Clause 34.1.

35. Termination

- 35.1 Termination for Default
 - a) The Procuring Entity, without prejudice to any other remedy for breach of Contract, by written notice of default sent to the Supplier, may terminate the Contract in whole or in part:
 - i) if the Supplier fails to deliver any or all of the Goods within the period specified in the Contract, or within any extension thereof granted by the Procuring Entity pursuant to GCC Clause 34;
 - ii) if the Supplier fails to perform any other obligation under the Contract; or
 - iii) if the Supplier, in the judgment of the Procuring Entity has engaged in Fraud and Corruption, as defined in paragraph 2.2 a of the Appendix to the GCC, in competing for or in executing the Contract.
 - b) In the event the Procuring Entity terminates the Contract in whole or in part, pursuant to GCC Clause 35.1(a), the Procuring Entity may procure, upon such terms and in suchmanner as it deems appropriate, Goods or Related Services similar to those undelivered or not performed, and the Supplier shall be liable to the Procuring Entity for any additional costs for such similar Goods or Related Services. However, the Supplier shall continue performance of the Contract to the extent not terminated.
- 352 Termination for Insolvency.

The Procuring Entity may at any time terminate the Contract by giving notice to the Supplier if

the Supplier becomes bankrupt or otherwise insolvent. In such event, termination will be without compensation to the Supplier, provided that such termination will not prejudice or affect any rightof action or remedy that has accrued or will accrue thereafter to the Procuring Entity

352 Termination for Convenience.

- a) The Procuring Entity, by notice sent to the Supplier, may terminate the Contract, in wholeor in part, at any time for its convenience. The notice of termination shall specify that termination is for the Procuring Entity's convenience, the extent to which performance of the Supplier under the Contract is terminated, and the date upon which such termination becomes effective.
- b) The Goods that are complete and ready for shipment within twenty-eight (28) days after the Supplier's receipt of notice of termination shall be accepted by the Procuring Entity at the Contract terms and prices. For the remaining Goods, the Procuring Entity may elect:
 - i) to have any portion completed and delivered at the Contract terms and prices; and/or
 - to cancel the remainder and pay to the Supplier an agreed amount for partially completed Goods and Related Services and for materials and parts previously procured by the Supplier.

36. Assignment

36.1 Neither the Procuring Entity nor the Supplier shall assign, in whole or in part, their obligations under this Contract, except with prior written consent of the other party.

37. Export Restriction

37.1 Notwithstanding any obligation under the Contract to complete all export formalities, any export restrictions attributable to the Procuring Entity, to Kenya, or to the use of the products/goods, systems or services to be supplied, which arise from trade regulations from a country supplying those products/goods, systems or services, and which substantially impede the Supplier frommeeting its obligations under the Contract, shall release the Supplier from the obligation to provide deliveries or services, always provided, however, that the Supplier can demonstrate to the satisfaction of the Procuring Entity that it has completed all formalities in a timely manner, including applying for permits, authorizations and licenses necessary for the export of the products/goods, systems or services under the terms of the Contract. Termination of the Contract on this basis shall be for the Procuring Entity's convenience pursuant to Sub-Clause 35.3.

SECTION VII - SPECIAL CONDITIONS OF CONTRACT

The following Special Conditions of Contract (SCC) shall supplement and/or amend the General Conditions of Contract (GCC). Whenever there is a conflict, the provisions herein shall prevail over those in the GCC.

[The Procuring Entity shall select insert the appropriate wording using the samples below or other acceptable wording, and delete the text in italics].

SECTION VII - SPECIAL CONDITIONS OF CONTRACT

The following Special Conditions of Contract (SCC) shall supplement and / or amend the General Conditions of Contract (GCC). Whenever there is a conflict, the provisions herein shall prevail over those in the GCC.

[The Procuring Entity shall select insert the appropriate wording using the samples below or other acceptable wording, and delete the text in italics]

Number of GC	Amendments of, and Supplements to, Clauses in the General Conditions of Contract			
Clause				
GCC 1.1(h)	The Procuring Entity is: COUNTY OF GOVERNMENT OF KISUMU			
GCC 4.2 (a)	The meaning of the trade terms shall be as prescribed by Incoterms. If the meaning of any trade term and the rights and obligations of the parties thereunder shall not be as prescribed by Incoterms, they shall be as prescribed by: [exceptional; refer to other internationally accepted trade terms]N/A			
GCC 4.2 (b)	The version edition of Incoterms shall be <i>INCOTERMS 2015</i>			
GCC 8.1	For <u>notices</u> , the Procuring Entity's address shall be: PROSPERITY BUILDING Attention: [insert full name of person, if applicable] Postal address (full postal address) Physical Address (full Location Address- insert city, street name, Building named floor number, room number) Telephone: [include telephone number, including country and city codes] Electronic mail address: [insert e-mail address, if applicable]			
GCC 10.4.2				
GCC 10.4.2	The place of arbitration shall be: KISUMU, KENYA Details of Shipping and other Documents to be furnished by the Supplier are [insert trequired documents, such as a negotiable bill of lading, a non-negotiable sea way bill, a airway bill, a railway consignment note, a road consignment note, insurance certificate Manufacturer's or Supplier's warranty certificate, inspection certificate issued nominated inspection agency, Supplier's factory shipping details etc.].			
	The above documents shall be received by the Procuring Entity before arrival of the Goods and, if not received, the Supplier will be responsible for any consequent expenses.			
GCC 15.1	The prices charged for the Goods supplied and the related Services performed [insert "shall" or "shall not," as appropriate] be adjustable. If prices are adjustable, the following method shall be used to calculate the price			
	adjustment [see attachment to these SCC for a sample Price Adjustment Formula]			
GCC 16.1	Sample provision GCC 16.1—The method and conditions of payment to be made to the Supplier under this Contract shall be as follows:			
	A. Payment for Goods supplied from abroad: Payment of foreign currency portion shall be made in [insert currency of the Contract Price] in the following manner:			
	(i) Advance Payment: Ten (10) percent of the Contract Price shall be paid within thirty (30) days of signing of the Contract, and upon submission of claim and a bank guarantee for equivalent amount valid until the Goods are delivered and, in the form, provided in the Tendering document or another form acceptable to the Procuring Entity.			

i e	
	(ii) On Shipment: Eighty (80) percent of the Contract Price of the Goods shipped shall be paid through irrevocable confirmed letter of credit opened in favour of the Supplier in a bank in its country, upon submission of documents specified in GCC Clause 12.
	(iii) On Acceptance: Ten (10) percent of the Contract Price of Goods received shall be paid within thirty (30) days of receipt of the Goods upon submission of claim supported by the acceptance certificate issued by the Procuring Entity.
	B. Payment of local currency portion of a foreign Supplier shall be made in Kenya shillings within thirty (30) days of presentation of claim supported by a certificate from the Procuring Entity declaring that the Goods have been delivered and that all other contracted Services have been performed. C. Payment for Goods and Services supplied from within Kenya:
	Payment for Goods and Services supplied from within Kenya shall be made in[currency], as follows:
	(i) Advance Payment: Ten (10) percent of the Contract Price shall be paid within thirty (30) days of signing of the Contract against an invoice and a bank guarantee for the equivalent amount and in the form provided in the Tendering document or another form acceptable to the Procuring Entity.
	(ii) On Delivery: Eighty (80) percent of the Contract Price shall be paid on receipt of the Goods and upon submission of the documents specified in GCC Clause 13. The bank guarantee shall then be released.
	(iii) On Acceptance: The remaining ten (10) percent of the Contract Price shall be paid to the Supplier within thirty (30) days after the date of the acceptance certificate for the respective delivery issued by the Procuring Entity.
GCC 16.5	The payment-delay period after which the Procuring Entity shall pay interest to the supplier shall be [insert number] days.
	The interest rate that shall be applied is [insert number] %
GCC 18.1	A Performance Security shall be required – N/A
GCC 18.3	If required, the Performance Security shall be in the form of: a Performance Bond
	If required, the Performance security shall be denominated in: KENYA SHILLINGS
GCC 18.4	Discharge of the Performance Security shall take place: [insert date if different from the one indicated in sub clause GCC 18.4]
GCC 23.2	The packing, marking and documentation within and outside the packages shall be:[insert in detail the type of packing required, the markings in the packing and all documentation required]
GCC 24.1	The insurance coverage shall be as specified in the Incoterms. If not in accordance with Incoterms, insurance shall be as follows: [insert specific insurance provisions agreed upon, including coverage, currency and amount]
GCC 25.1	Responsibility for transportation of the Goods shall be as specified in the Incoterms.
	-

1	
	If not in accordance with Incoterms, responsibility for transportations shall be as follows: [insert "The Supplier is required under the Contract to transport the Goods to a specified place of final destination within Kenya, defined as the Project Site, transport to such place of destination in Kenya, including insurance and storage, as shall be specified in the Contract, shall be arranged by the Supplier, and related costs shall be included in the Contract Price"; or any other agreed upon trade terms (specify the respective responsibilities of the Procuring Entity and the Supplier)] – N/A
GCC 25.2	Incidental services to be provided are:
	[Selected services covered under GCC Clause 25.2 and/or other should be specified with the desired features. The price quoted in the Tender price or agreed with the selected Supplier shall be included in the Contract Price.] N/A
GCC 26.1	The inspections and tests shall be: [insert nature, frequency, procedures for carrying out
	the inspections and tests]
GCC 26.2	The Inspections and tests shall be conducted at: $[insert\ name(s)\ of\ location(s)]$
GCC 27.1	The liquidated damage shall be: [insert number] % per week
GCC 27.1	The maximum amount of liquidated damages shall be: [insert number] %
GCC 28.3	The period of validity of the Warranty shall be: [insert number] days For purposes of the Warranty, the place(s) of final destination(s) shall be: [insert name(s) of location(s)]
	Sample provision
	GCC 28.3—In partial modification of the provisions, the warranty period shall be hours of operation or months from date of acceptance of the Goods or () months from the date of shipment, whichever occurs earlier. The Supplier shall, in addition, comply with the performance and/or consumption guarantees specified under the Contract. If, for reasons attributable to the Supplier, these guarantees are not attained in whole or in part, the Supplier shall, at its discretion, either:
	(a) make such changes, modifications, and/or additions to the Goods or any partthereof as may be necessary in order to attain the contractual guarantees specified in the Contract at its own cost and expense and to carry out further performance tests inaccordance with GCC 26.7,
	or
	(b) pay liquidated damages to the Procuring Entity with respect to the failure to meet the contractual guarantees. The rate of these liquidated damages shall be ().
	[The rate should be higher than the adjustment rate used in the Tender evaluation under $TDS\ 34.6(f)$]
GCC 28.6 28.5,	The period for repair or replacement shall be: [insert number(s)] days. N/A
GCC 33.6	If the value engineering proposal is approved by the Procuring Entity the amount to be paid to the Supplier shall be% (insert appropriate percentage.
	The percentage is normally up to 50%) of the reduction in the Contract Price.

SECTION VIII - CONTRACT FORMS

This Section contains forms which, once completed, will form part of the Contract. The forms for Performance Security and Advance Payment Security, when required, shall only be completed by the successful tenderer after contract award.

FORM No. 1: NOTIFICATION OF INTENTION TO AWARD

This Notification of Intention to Award shall be sent to each Tenderer that submitted a Tender. Send

this Notification to the Tenderer's Authorized Representative named in the Tender Information Form on the format below. **FORMAT** 1. For the attention of Tenderer's Authorized Representative Name:______[insert Authorized Representative's name] I)ii) Address: [insert Authorized Representative's Address] Telephone: [insert Authorized Representative's telephone/fax numbers] iii) Email Address: [insert Authorized Representative's email address] iv) [IMPORTANT: insert the date that this Notification is transmitted to Tenderers. The Notification must be sent to all Tenderers simultaneously. This means on the same date and as close to the same time as possible.] 2 Date of transmission: [email] on [date] (local time) This Notification is sent by (Name and designation) 3. Notification of Intention to Award Employer:_____[insert the name of the Employer] I)Project: [insert name of project] Contract title: _____[insert the name of the contract] iii) Country: _____[insert country where ITT is issued] iv) ITT No: _______[insert ITT reference number from Procurement Plan] This Notification of Intention to Award (Notification) notifies you of our decision to award the above contract. The transmission of this Notification begins the Standstill Period. During the Standstill Period, you may: Request a debriefing in relation to the evaluation of your tender 4. Submit a Procurement-related Complaint in relation to the decision to award the contract. a) The successful tenderer Name of successful Tender i) Address of the successful Tender ii) Contract price of the successful Tender Kenya Shillings (in words b) Other Tenderers Names of all Tenderers that submitted a Tender. If the Tender's price was evaluated include the evaluated price as well as the Tender price as read out. For Tenders not evaluated, give onemain reason the Tender was unsuccessful.

S/No.	Name of Tender	Tender Price	Tender's evaluated	One Reason Why Not Evaluated		
		as read out price (N		as read out price (Note a)	price (Note a)	
1						
2						
3						
4						
5						

(Note a) State NE if not evaluated

- 5. How to request a debriefing
 - a) DEADLINE: The deadline to request a debriefing expires at midnight on [insert date] (local time).
 - b) You may request a debriefing in relation to the results of the evaluation of your Tender. If you decide to request a debriefing your written request must be made within three (5) Business Days of receipt of this Notification of Intention to Award.
 - c) Provide the contract name, reference number, name of the Tenderer, contact details; and address the request for debriefing as follows:

1)	Attention:	[insert full name of person, if applicable]
ii)	Title/position:	[insert title/position]
ii)	Agency:	[insert name of Employer]
iii)	Email address:	[insert email address]

- d) If your request for a debriefing is received within the 3 Days deadline, we will provide the debriefing within five (3) Business Days of receipt of your request. If we are unable to provide the debriefing within this period, the Standstill Period shall be extended by five (3) Days after the date that the debriefing is provided. If this happens, we will notify you and confirm the date that the extended Standstill Period will end.
- e) The debriefing may be in writing, by phone, video conference call or in person. We shall promptly advise you in writing how the debriefing will take place and confirm the date and time.
- f) If the deadline to request a debriefing has expired, you may still request a debriefing. Inthis case, we will provide the debriefing as soon as practicable, and normally no later than fifteen (15) Days from the date of publication of the Contract Award Notice.
- 6. How to make a complaint
 - a) Period: Procurement-related Complaint challenging the decision to award shall be submitted by midnight, [insert date] (local time).
 - b) Provide the contract name, reference number, name of the Tenderer, contact details; and address the Procurement-related Complaint as follows:

I)	Attention:	_[insert full name of person, if applicable]
ii)	Title/position:	[insert title/position]
iii)	Agency:	[insert name of Employer]
iv)	Email address:	[insert email address]

- c) At this point in the procurement process, you may submit a Procurement-related Complaint challenging the decision to award the contract. You do not need to have requested, or received, a debriefing before making this complaint. Your complaint must be submitted within the Standstill Period and received by us before the Standstill Period ends.
- d) Further information: For more information refer to the Public Procurement and Disposals Act2015 and its Regulations available from the Website www.ppra.go.ke or email complaints@ppra.go.ke.

You should read these documents before preparing and submitting your complaint.

- e) There are four essential requirements:
- i) You must be an 'interested party'. In this case, that means a Tenderer who submitted a Tender in this tendering process, and is the recipient of a Notification of Intention to Award.
- ii) The complaint can only challenge the decision to award the contract.
- iii) You must submit the complaint within the period stated above.
- iv) You must include, in your complaint, all of the information required to support your complaint.

7. Standstill Period

- i) DEADLINE: The Standstill Period is due to end at midnight on [insert date] (local time).
- ii) The Standstill Period lasts FOURTEEN (14) Days after the date of transmission of this Notification of Intention to Award.
- iii) The Standstill Period may be extended as stated in paragraph Section 5 (d) above.

If you have any questions regarding this Notification please do not hesitate to contact us.

On behalf of the Employer:	
Signature:	
Name:	Title/

FORM NO. 2 - REQUEST FOR REVIEW

FORM FOR REVIEW(r.203(1))

PUBLIC PROCUREMENT ADMINISTRATIVE REVIEW BOARD **APPLICATION NO......OF......20......** BETWEENAPPLICANT ANDRESPONDENT (Procuring Entity) Request for review of the decision of the (Name of the Procuring Entity ofdated the ...day of REQUEST FOR REVIEW the whole/part of the above mentioned decision on the following grounds, namely: 1. By this memorandum, the Applicant requests the Board for an order/orders that: 1. 2. FOR OFFICIAL USE ONLY Lodged with the Secretary Public Procurement Administrative Review Board onday of20...... **SIGNED Board Secretary**

FORM NO. 3 LETTER OF AWARD

[Use letter head paper of the Procuring Entity]	
[Date]	
To:[name and address of the Supplier]	
Subject:Notification of Award Contract No	
This is to notify you that your Tender dated	
You are requested to furnish the Performance Security within 30 days in accordance with the Conditionsof Contract, using for that purpose the of the Performance Security Form included in Section X, Contract Forms, of the Tendering document.	
Authorized Signature:	Name

Attachment: Contract Agreement

FORM NO. 4 - CONTRACT AGREEMENT

[The s	succes	sful tenderer shall fill in this form in accordance with the instructions indicated]				
year]. princi Entity of [in	BET pal pl '"), o usert:	WEEN (1)				
1.	WHEREAS the Procuring Entity invited Tenders for certain Goods and ancillary services, viz.,					
	i)	In this Agreement words and expressions shall have the same meanings as are respectively assigned to them in the Contract documents referred to.				
	The following documents shall be deemed to form and be read and construed as part of this Agreement. This Agreement shall prevail over all other contract documents.					
		a) the Letter of Acceptance				
		b) the Letter of Tender				
		c) the Addenda Nos(if any)				
		d) Special Conditions of Contract				
		e) General Conditions of Contract				
		f) the Specification (including Schedule of Requirements and Technical Specifications)				
		g) the completed Schedules (including Price Schedules)				
		h) any other document listed in GCC as forming part of the Contract				
	iii)	In consideration of the payments to be made by the Procuring Entity to the Supplier as specified in this Agreement, the Supplier hereby covenants with the Procuring Entity to provide the Goods and Services and to remedy defects therein in conformity in all respects with the provisions of the Contract.				
2.	and S	Procuring Entity hereby covenants to pay the Supplier in consideration of the provision of the Goods ervices and the remedying of defects therein, the Contract Price or such other sum asmay become ble under the provisions of the Contract at the times and in the manner prescribed by the fact.				
3.	IN WITNESS whereof the parties hereto have caused this Agreement to be executed in accordancewith the laws of Kenya on the day, month and year indicated above.					
<u>For</u> a	nd on	behalf of the Procuring Entity				
Signe	d:	[insert signature]				
in the	capa	city of[insert title or other appropriate designation] In the presence of				
		[insert identification of official witness] For and on behalf of the Supplier				
Signe	d:	[insert signature of authorized representative(s) of the Supplier] in the capacity of_				
		[insert title or other appropriate designation] in the presence of				
		[insert identification of official witness]				

FORM NO. 5 - PERFORMANCE SECURITY [Option 1 - Unconditional Demand Bank Guarantee]

Ben	eficiary:	[insert name and Address of	
	loyer]		
-	•	[Insert date of issue]	
Gua		[Insert name and address of place of issue, unless indicated in the	
1.	We have b	een informed that(herein	naft
	er called "th	e Contractor") has entered into Contract No	
		with (name of Employer) dated (the	
	Employer as Contract").	with (name of Employer) (the the Beneficiary), for the execution of (hereinafter called "the	
2.		we understand that, according to the conditions of the Contract, a guarantee is required.	
3.	At the request of the Contractor, we as Guarantor, hereby irrevocably undertake to paythe Beneficiary any sum or sums not exceeding in total an amount of(in words),¹ such sum being payable in the types and proportions of currencies in which the Contract Price is payable, upon receipt by us of the Beneficiary's complying demand supported by the Beneficiary's statement, whether in the demand itself or in a separate signed document accompanying or identifying the demand, stating that the Applicant is in breach of its obligation(s) under the Contract, without the Beneficiary needing to prove or to show grounds for your demand or the sum specified therein.		
4.		e shall expire, no later than the Day of	
5.	months] [one	agrees to a one-time extension of this guarantee for a period not to exceed, year], in response to the Beneficiary's written request for such extension, so presented to the Guarantor before the expiry of the guarantee."	
 [Nar	ne of Authorize	d Official, signature(s) and seals/stamps]	

Note: All italicized text (including footnotes) is for use in preparing this form and shall be deleted from the final product.

FORM No. 6 - PERFORMANCE SECURITY [Option 2– Performance Bond]

[Note: Procuring Entities are advised to use Performance Security – Unconditional Demand Bank Guarantee instead of Performance Bond due to difficulties involved in calling Bond holder to action]

[Guarantor letterhead or SWIFT identifier code]

Beneficiary:________ [insert name and Address of Employer] Date:______ [Insert date of issue]

PERFORMANCE BOND No.:______

Guarantor:[Insert name and address of place of issue, unless indicated in the

letterhead]

1.	By this Bond	as	Principal	(hereinafter	called	"the
	Contractor") and] as	Surety (her	einafter called	"the Surety	/"),are
	held and firmly bound unto] as	Obligee	(hereinafte	r called "the E	mployer")	in the
	amount of for the payment of which sum well	and truly	to be made	de in the types	and propo	ortions
	of currencies in which the Contract Price is paya	able, the	Contractor	and the Surety	bind thems	selves,
	their heirs, executors, administrators, successors	and assi	gns, jointly	and severally	firmly by	these
	presents.					

- 3. NOW, THEREFORE, the Condition of this Obligation is such that, if the Contractor shall promptly and faithfully perform the said Contract (including any amendments thereto), then this obligation shall be null and void; otherwise, it shall remain in full force and effect. Whenever the Contractor shall be, and declared by the Employer to be, in default under the Contract, the Employer having performed the Employer's obligations thereunder, the Surety may promptly remedy the default, or shall promptly:
 - 1) complete the Contract in accordance with its terms and conditions; or
 - obtain a tender or tenders from qualified tenderers for submission to the Employer for completing the Contract in accordance with its terms and conditions, and upon determination by the Employer and the Surety of the lowest responsive Tenderers, arrange for a Contract between such Tenderer, and Employer and make available as work progresses (even though there should be a default or a succession of defaults under the Contract or Contracts of completion arranged under this paragraph) sufficient funds to pay the cost of completion less the Balance of the Contract Price; but not exceeding, including other costs and damages for which the Surety may be liable hereunder, the amount set forth in the first paragraph hereof. The term "Balance of the Contract Price," as used in this paragraph, shall mean the total amount payable by Employer to Contractor under the Contract, less the amount properlypaid by Employer to Contractor; or
 - 3) pay the Employer the amount required by Employer to complete the Contract in accordance with its terms and conditions up to a total not exceeding the amount of this Bond.
- 4. The Surety shall not be liable for a greater sum than the specified penalty of this Bond.
- 5. Any suit under this Bond must be instituted before the expiration of one year from the date of the issuing of the Taking-Over Certificate. No right of action shall accrue on this Bond to or for the use of any person or corporation other than the Employer named herein or the heirs, executors, administrators, successors, and assigns of the Employer.
- 6. In testimony whereof, the Contractor has hereunto set his hand and affixed his seal, and the Surety has

		aled with his corporate seal duly attested by the signature of his leof	gal
SIGNEI	O ON	on behalf of	
Ву		in the capacity of	
In the p	resence of		
SIGNEI	O ON	on behalf of	
Ву		in the capacity of	
In the p	resence of		

<u>FORM NO. 7 - ADVANCE PAYMENT SECURITY [Demand Bank [Guarantor letterhead]</u>	Guarantee]
Beneficiary: name and Address of Employer] Date:[Insert date of issue]	[Insert
ADVANCE PAYMENT GUARANTEE No.: number]	[Insert guarantee reference
Guarantor: [Insert name and address of place of issue, unless indica	ted in the letterhead]
1. We have been informed that (hereinafter Contract No dated with the contract No hereinafter (hereinafter hereinafter	called "the Contractor") has entered into ne Beneficiary, for the execution of
(hereinafter called "the Contract").	
2. Furthermore, we understand that, according to the conditions sum_(<i>in words</i>) is to be made against an advance payment guarant	
3. At the request of the Contractor, we as Guarantor, hereby irrevenum or sums not exceeding in total an amount of upon receipt by us of the Beneficiary's complying demand supported the demand itself or in a separate signed document accompanying or in Applicant:	(in words) ¹ I by the Beneficiary's statement, whether in
(a) has used the advance payment for purposes other than the costs(b) has failed to repay the advance payment in accordance with the which the Applicant has failed to repay.	
4. A demand under this guarantee may be presented as from the prometric from the Beneficiary's bank stating that the advance payment referred on its account numberat	to above has been credited to the Contractor
5. The maximum amount of this guarantee shall be progressive payment repaid by the Contractor as specified in copies of interim state be presented to us. This guarantee shall expire, at the latest, upon o certificate indicating that ninety (90) percent of the Accepted Contracertified for payment, or on the day of, 2, demand for payment under this guarantee must be received by us at the	atements or payment certificates which shall ur receipt of a copy of the interim payment act Amount, less provisional sums, has been whichever is earlier. Consequently, any
6. The Guarantor agrees to a one-time extension of this guarantee <i>year</i>], in response to the Beneficiary's written request for such ext Guarantor before the expiry of the guarantee.	
[Name of Authorized Official, signature(s) and seals/stamps]	

Note: All italicized text (including footnotes) is for use in preparing this form and shall be deleted from the final product.

¹The Guarantor shall insert an amount representing the amount of the advance payment and denominated either in the currency of the advance payment as specified in the Contract.

² Insert the expected expiration date of the Time for Completion. The Employer should note that in the event of an extension of the time for completion of the Contract, the Employer would need to request an extension of this guarantee from the Guarantor.

Such request must be in writing and must be made prior to the expiration date established in the guarantee.

FORM NO. 8 BENEFICIAL OWNERSHIP DISCLOSURE FORM

INSTRUCTIONS TO TENDERERS: DELETE THIS BOX ONCE YOU HAVE COMPLETED THE FORM

This Beneficial Ownership Disclosure Form ("Form") is to be completed by the successful tenderer. In case of joint venture, the tenderer must submit a separate Form for each member. The beneficial ownership information to be submitted in this Form shall be current as of the date of its submission.

For the purposes of this Form, a Beneficial Owner of a Tenderer is any natural person who ultimately owns or controls the Tenderer by meeting one or more of the following conditions:

- *Directly or indirectly holding 25% or more of the shares.*
- Directly or in directly holding 25% or more of the voting rights.
- Directly or indirectly having the right to appoint a majority of the board of directors or equivalent governing body of the Tenderer.

Tender Reference No.:	[insert identification	
no] Name of the Assignment:	[insert name of the assignment]	
to:insert comple	ete name of Procuring Entity]	
In response to your notification of award datedadditional information on beneficial ownership:options that are not applicable]	[insert date of notification of award] to furnish [select one option as applicable and delete the	
I) We here by provide the following beneficial ow	vnership information.	

Details of beneficial ownership

Identity of Beneficial Owner	Directly or indirectly holding 25% or more of the shares (Yes / No)	Directly or indirectly holding 25 % or more of the Voting Rights (Yes / No)	Directly or indirectly having the right to appoint a majority of the board of the directors or an equivalent governing body of the Tenderer (Yes / No)
[include full name (last, middle, first), nationality, country of residence]			

OR

ii) We declare that there is no Beneficial Owner meeting one or more of the following conditions: directly or indirectly holding 25% or more of the shares. Directly or indirectly holding 25% or more of the voting rights. Directly or indirectly having the right to appoint a majority of the board of directors or equivalent governing body of the Tenderer.

OR

We declare that we are unable to identify any Beneficial Owner meeting one or more of the following conditions. [If this option is selected, the Tenderer shall provide explanation on why it is unable to identify any Beneficial Owner]

Directly or indirectly holding 25% or more of the shares. Directly or indirectly holding 25% or more of the voting rights.

Name of the Tenderer*[insert complete name of the Tenderer]
Name of the person duly authorized to sign the Tender on behalf of the Tenderer: ** [insert complete name of person duly authorized to sign the Tender]
Title of the person signing the Tender[insert complete title of the person signing the Tender]
Signature of the person named above
Date signed [insert date of signing] day of [Insert month], [insert year]

Directly or indirectly having the right to appoint a majority of the board of directors or equivalent governing

body of the Tenderer]"